New England ACCUTEST Laboratories

IT'S ALL IN THE CHEMISTRY

05/21/07

Technical Report for

GEI Consultants, Inc.

Indoor & Outdoor Air Samples Tufts St., Somerville MA

045162

Accutest Job Number: M64467

Sampling Date: 04/27/07

Report to:

GEI Consultants, Inc.
400 Unicorn Park Drive
Woburn, MA 01801-1970

ATTN: Raymond Siegener, Ph.D.

Total number of pages in report: **140**

ACCUTEST LABORATORIES 50 YEARS 1956-2006

ACCREDITED IN ACCORDANCE WITH nelac

Test results contained within this data package meet the requirements of the National Environmental Laboratory Accreditation Conference and/or state specific certification programs as applicable.

Reza Tand
Lab Director

Certifications: MA (M-MA136) CT (PH-0109) NH (250204) RI (00071) ME (MA136) FL (E87579) NY (23346) NJ (MA926) NAVY USACE

This report shall not be reproduced, except in its entirety, without the written approval of Accutest Laboratories.

Table of Contents

-1-

Section 1: Sample Summary **3**
Section 2: Case Narrative/Conformance Summary **4**
Section 3: Sample Results **5**
3.1: M64467-1: 045162-97R FRANK-SS1 6
Section 4: Misc. Forms **7**
4.1: Chain of Custody 8
4.2: Summa Canister and Flow Controller Log 13
4.3: Sample Tracking Chronicle 14
Section 5: GC/MS Volatiles - QC Data Summaries **15**
5.1: Method Blank Summary 16
5.2: Blank Spike Summary 17
5.3: Duplicate Summary 18
5.4: Instrument Performance Checks (BFB) 19
5.5: Internal Standard Area Summaries 21
5.6: Initial Calibration RT/ISTD Area Summaries 22
5.7: Surrogate Recovery Summaries 36
5.8: Initial and Continuing Calibration Summaries 37
Section 6: GC/MS Volatiles - Raw Data **46**
6.1: Samples 47
6.2: Method Blanks 55
6.3: Blank Spikes 57
6.4: Duplicates 60
6.5: Instrument Performance Checks (BFB) 69
6.6: Initial and Continuing Calibarations 87
6.7: Instrument Run Logs 139

Accutest Laboratories

Sample Summary

GEI Consultants, Inc.

Job No: M64467

Indoor & Outdoor Air Samples Tufts St., Somerville MA
Project No: 045162

Sample Number	Collected Date	Time	By	Received	Matrix Code	Type	Client Sample ID
M64467-1	04/27/07	16:20	TAD	04/30/07	AIR	Air	045162-97R FRANK-SS1

SAMPLE DELIVERY GROUP CASE NARRATIVE

Client:	GEI Consultants, Inc.	**Job No**	M64467
Site:	Indoor & Outdoor Air Samples Tufts St., Somerville MA	**Report Date**	5/19/2007 3:11:49 PM

1 Sample was collected on 04/27/2007 and were received at Accutest on 04/30/2007. These Samples received an Accutest job number of M64467. A listing of the Laboratory Sample ID, Client Sample ID and dates of collection are presented in the Results Summary Section of this report.

Except as noted below, all method specified calibrations and quality control performance criteria were met for this job. For more information, please refer to QC summary pages.

Volatiles by GCMS By Method TO-15

Matrix AIR **Batch ID:** MSJ533

- All samples were analyzed within the recommended method holding time.
- All method blanks for this batch meet method specific criteria.
- Sample(s) M64467-1DUP were used as the QC samples indicated.

The Accutest Laboratories of New England certifies that all analysis were performed within method specification. It is further recommended that this report to be used in its entirety.The Accutest Laboratories of NE, Laboratory Director or assignee as verified by the signature on the cover page has authorized the release of this report(M64467).

Sample Results

Report of Analysis

Accutest Laboratories

Report of Analysis

Client Sample ID: 045162-97R FRANK-SS1
Lab Sample ID: M64467-1 **Date Sampled:** 04/27/07
Matrix: AIR - Air Summa ID: M126 **Date Received:** 04/30/07
Method: TO-15 **Percent Solids:** n/a
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	J10289.D	1	05/10/07	DFT	n/a	n/a	MSJ533
Run #2							

	Initial Volume
Run #1	400 ml
Run #2	

CAS No.	MW	Compound	Result	RL	Units	Q	Result	RL	Units
75-00-3	64.52	Chloroethane	ND	0.20	ppbv		ND	0.53	ug/m3
56-23-5	153.8	Carbon tetrachloride	ND	0.20	ppbv		ND	1.3	ug/m3
75-34-3	98.96	1,1-Dichloroethane	ND	0.20	ppbv		ND	0.81	ug/m3
75-35-4	96.94	1,1-Dichloroethylene	ND	0.20	ppbv		ND	0.79	ug/m3
107-06-2	98.96	1,2-Dichloroethane	ND	0.20	ppbv		ND	0.81	ug/m3
156-60-5	96.94	trans-1,2-Dichloroethylene	ND	0.20	ppbv		ND	0.79	ug/m3
156-59-2	96.94	cis-1,2-Dichloroethylene	ND	0.20	ppbv		ND	0.79	ug/m3
71-55-6	133.4	1,1,1-Trichloroethane	ND	0.20	ppbv		ND	1.1	ug/m3
79-34-5	167.9	1,1,2,2-Tetrachloroethane	ND	0.20	ppbv		ND	1.4	ug/m3
79-00-5	133.4	1,1,2-Trichloroethane	ND	0.20	ppbv		ND	1.1	ug/m3
127-18-4	165.8	Tetrachloroethylene	8.9	0.20	ppbv		60	1.4	ug/m3
79-01-6	131.4	Trichloroethylene	ND	0.20	ppbv		ND	1.1	ug/m3
75-01-4	62.5	Vinyl chloride	ND	0.20	ppbv		ND	0.51	ug/m3

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
460-00-4	4-Bromofluorobenzene	95%		57-139%

ND = Not detected
RL = Reporting Limit
E = Indicates value exceeds calibration range
J = Indicates an estimated value
B = Indicates analyte found in associated method blank
N = Indicates presumptive evidence of a compound

Misc. Forms

Custody Documents and Other Forms

ACCUTEST Laboratories

CHAIN OF CUSTODY

Air Sampling Field Data Sheet

FED-EX Tracking #: | Bottle Order Control #: | Lab Quote #: | Lab Job #: M64467

Client / Reporting Information

Company Name: GEI
Address: 400 Unicorn Park Dr
City: Woburn State: MA Zip: 01801
Project Contact: Mark Ensign E-mail: mensign@geiconsultants.com
Phone #: 781-721-4000 Fax #: 781-721-4073
Sampler(s) Name(s):

Project Name: 50 Tufts St
Street: 50 Tufts St
City: Somerville State: MA
Project #: 045162
Client Purchase Order #:

Weather Parameters

Temperature (Fahrenheit) Start: 52.1 Maximum: Stop: 51.5 Minimum:
Atmospheric Pressure (inches of Hg) Start: 29.93 Maximum: Stop: 29.90 Minimum:
Other weather comment:

Requested Analysis

Standard TO-15 Reporting List | MA DEP APH

Lab Sample #	Field ID / Point of Collection	Air Type Indoor(I) Soil Vap(SV) Ambient(A)	Sampling Equipment Info: Canister Serial #	Canister Size 6L or 1L	Flow Controller Serial #	Start Sampling Information: Date	Time (24hr clock)	Canister Pressure ("Hg)	Interior Temp (F)	Sampler Init.	Stop Sampling Information: Date	Time (24hr clock)	Canister Pressure ("Hg)	Interior Temp (F)	Sampler Init.
-1	045162-97RFRANK-SS1	SV	M126	6	MFC002	4-27-07	1515	29	–	TAD	4-27-07	1620	4	–	TAD

Turnaround Time (Business days)

Standard - 15 Days ✓
10 Day
5 Day
3 Day
2 Day
1 Day
Other

Approved By:
Date:

Data Deliverable Information

Comm A
Comm B
Full T1
Other:

Comments / Remarks

See attached list

Sample Custody must be documented below each time samples change possession, including courier delivery.

Relinquished by Laboratory:	Date Time:	Received By:	Relinquished By:	Date Time:	Received By:
1 GEI lab	4-27-07 1030	1 [signature]	2 [signature]	4-27-07 1700	2 GEI lab
Relinquished by: 3 GEI Lab	Date Time: 4/30/07 1500	Received By: 3 [signature]	Relinquished By: 4 [signature]	Date Time: 4/30/07 1500	Received By: 4 [signature]
Relinquished By: 5 [signature]	Date Time: 4/30/07 17:8	Received By: 5 [signature]	Custody Seal #		

SM013-08 (2/14/06)

4.1 4

M64467: Chain of Custody
Page 1 of 5

Subslab

~~Indoor~~ Air Sampling Analyte List

- Chloroethane
- Carbon Tetrachloride
- 1,1-Dichloroethane
- 1,1-Dichloroethylene
- 1,2-Dichloroethane
- trans-1,2-Dichloroethylene
- cis-1,2-Dichloroethylene
- 1,1,1-Trichloroethane
- 1,1,2,2-Tetrachloroethane
- 1,1,2-Trichloroethane
- Tetrachloroethylene
- Trichloroethylene
- Vinyl Chloride

Summa Certification Blanks

Sample ID	Canister ID
M64467-1	M126

Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\
Data File : J9864.D
Acq On : 30 Mar 2007 8:10 pm
Operator : PhilipB
Sample : M63703-14 (M126)
Misc : MS13284, MSJ510,,,,,1
ALS Vial : 9 Sample Multiplier: 1

Quant Time: Apr 02 09:19:54 2007
Quant Method : C:\MSDCHEM\1\METHODS\J032207T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Fri Mar 23 07:52:47 2007
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) BROMOCHLOROMETHANE	8.63	128	312328	10.00	PPBV	0.00
35) 1,4-DIFLUOROBENZENE	10.44	114	549807	10.00	PPBV	0.00
49) CHLOROBENZENE-D5	14.67	117	470170	10.00	PPBV	0.00

System Monitoring Compounds

61) 4-BROMOFLUOROBENZENE	16.29	95	124434	4.19	PPBV	0.00

Spiked Amount 5.000 Range 51 - 142 Recovery = 83.80%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

J032207T.M Mon Apr 02 09:31:25 2007 Page: 1

4.1 4

M64467: Chain of Custody
Page 4 of 5

Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\
Data File : J9864.D
Acq On : 30 Mar 2007 8:10 pm
Operator : PhilipB
Sample : M63703-14 (M126)
Misc : MS13284, MSJ510,,,,,1
ALS Vial : 9 Sample Multiplier: 1

Quant Time: Apr 02 09:19:54 2007
Quant Method : C:\MSDCHEM\1\METHODS\J032207T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Fri Mar 23 07:52:47 2007
Response via : Initial Calibration



J032207T.M Mon Apr 02 09:31:26 2007 Page: 2

4.1 4

M64467: Chain of Custody
Page 5 of 5

Summa Canister and Flow Controller Log

Job Number: M64467
Account: GEI GEI Consultants, Inc.
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA
Received: 04/30/07

SUMMA CANISTERS

Shipping							Receiving						
Summa ID	L	Vac " Hg	Date Out	By	SCC Batch	SCC FileID	Sample Number	Date In	By	Vac " Hg	Pres psig	Final psig	Dil Fact
M126	6	29.4	04/02/07	PB	CP391		M64467-1	05/10/07	DFT	1.5			1

FLOW CONTROLLERS

Shipping					Receiving		
Flow Crtl ID	Date Out	By	cc/ min	Time hrs.	Date In	By	cc/ min
MC002	04/02/07	PB	80	1	05/15/07	PB	84.3

Accutest Bottle Order(s):
PB/4-2-07/GEI/TUFTS ST.

Prep Date	Room Temp(F)	Bar Pres "Hg
04/02/07	70	29.92

Internal Sample Tracking Chronicle

GEI Consultants, Inc.

Job No: M64467

Indoor & Outdoor Air Samples Tufts St., Somerville MA
Project No: 045162

Sample Number	Method	Analyzed	By	Prepped	By	Test Codes
M64467-1	Collected: 27-APR-07 16:20 By: TAD		Received: 30-APR-07 By: RS			
045162-97R	FRANK-SS1					
M64467-1	TO-15	10-MAY-07 13:03	DFT			VTO15STD

GC/MS Volatiles

QC Data Summaries

Includes the following where applicable:

- Method Blank Summaries
- Blank Spike Summaries
- Matrix Spike and Duplicate Summaries
- Instrument Performance Checks (BFB)
- Internal Standard Area Summaries
- Initial Calibration RT/ISTD Area Summaries
- Surrogate Recovery Summaries
- Initial and Continuing Calibration Summaries

Method Blank Summary

Job Number: M64467
Account: GEI GEI Consultants, Inc.
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
MSJ533-MB	J10288.D	1	05/10/07	DFT	n/a	n/a	MSJ533

The QC reported here applies to the following samples: **Method:** TO-15

M64467-1

CAS No.	Compound	Result	RL	Units	Q	Result	RL	Units
75-00-3	Chloroethane	ND	0.20	ppbv		ND	0.53	ug/m3
56-23-5	Carbon tetrachloride	ND	0.20	ppbv		ND	1.3	ug/m3
75-34-3	1,1-Dichloroethane	ND	0.20	ppbv		ND	0.81	ug/m3
75-35-4	1,1-Dichloroethylene	ND	0.20	ppbv		ND	0.79	ug/m3
107-06-2	1,2-Dichloroethane	ND	0.20	ppbv		ND	0.81	ug/m3
156-60-5	trans-1,2-Dichloroethylene	ND	0.20	ppbv		ND	0.79	ug/m3
156-59-2	cis-1,2-Dichloroethylene	ND	0.20	ppbv		ND	0.79	ug/m3
71-55-6	1,1,1-Trichloroethane	ND	0.20	ppbv		ND	1.1	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.20	ppbv		ND	1.4	ug/m3
79-00-5	1,1,2-Trichloroethane	ND	0.20	ppbv		ND	1.1	ug/m3
127-18-4	Tetrachloroethylene	ND	0.20	ppbv		ND	1.4	ug/m3
79-01-6	Trichloroethylene	ND	0.20	ppbv		ND	1.1	ug/m3
75-01-4	Vinyl chloride	ND	0.20	ppbv		ND	0.51	ug/m3

CAS No.	Surrogate Recoveries		Limits
460-00-4	4-Bromofluorobenzene	95%	57-139%

Blank Spike Summary

Job Number: M64467
Account: GEI GEI Consultants, Inc.
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
MSJ533-BS	J10286B.D	1	05/10/07	DFT	n/a	n/a	MSJ533

The QC reported here applies to the following samples: **Method:** TO-15

M64467-1

CAS No.	Compound	Spike ppbv	BSP ppbv	BSP %	Limits
75-00-3	Chloroethane	10	10.2	102	70-130
56-23-5	Carbon tetrachloride	10	8.6	86	70-130
75-34-3	1,1-Dichloroethane	10	8.5	85	70-130
75-35-4	1,1-Dichloroethylene	10	8.7	87	70-130
107-06-2	1,2-Dichloroethane	10	8.1	81	70-130
156-60-5	trans-1,2-Dichloroethylene	10	8.8	88	70-130
156-59-2	cis-1,2-Dichloroethylene	10	8.6	86	70-130
71-55-6	1,1,1-Trichloroethane	10	8.6	86	70-130
79-34-5	1,1,2,2-Tetrachloroethane	10	8.7	87	70-130
79-00-5	1,1,2-Trichloroethane	10	8.0	80	70-130
127-18-4	Tetrachloroethylene	10	8.0	80	70-130
79-01-6	Trichloroethylene	10	8.8	88	70-130
75-01-4	Vinyl chloride	10	8.9	89	70-130

CAS No.	Surrogate Recoveries	BSP	Limits
460-00-4	4-Bromofluorobenzene	111%	57-139%

Duplicate Summary

Job Number: M64467
Account: GEI GEI Consultants, Inc.
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
M64467-1DUP	J10293.D	1	05/10/07	DFT	n/a	n/a	MSJ533
M64467-1	J10289.D	1	05/10/07	DFT	n/a	n/a	MSJ533

The QC reported here applies to the following samples: **Method:** TO-15

M64467-1

CAS No.	Compound	M64467-1 ppbv	Q	DUP ppbv	Q	RPD	Limits
75-00-3	Chloroethane	ND		ND		nc	25
56-23-5	Carbon tetrachloride	ND		ND		nc	25
75-34-3	1,1-Dichloroethane	ND		ND		nc	25
75-35-4	1,1-Dichloroethylene	ND		ND		nc	25
107-06-2	1,2-Dichloroethane	ND		ND		nc	25
156-60-5	trans-1,2-Dichloroethylene	ND		ND		nc	25
156-59-2	cis-1,2-Dichloroethylene	ND		ND		nc	25
71-55-6	1,1,1-Trichloroethane	ND		ND		nc	25
79-34-5	1,1,2,2-Tetrachloroethane	ND		ND		nc	25
79-00-5	1,1,2-Trichloroethane	ND		ND		nc	25
127-18-4	Tetrachloroethylene	8.9		11.0		21	25
79-01-6	Trichloroethylene	ND		ND		nc	25
75-01-4	Vinyl chloride	ND		ND		nc	25

CAS No.	Surrogate Recoveries	DUP	M64467-1	Limits
460-00-4	4-Bromofluorobenzene	94%	95%	57-139%

Instrument Performance Check (BFB)

Job Number: M64467
Account: GEI GEI Consultants, Inc.
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

Sample: MSJ533-BFB
Lab File ID: J10277.D
Instrument ID: GCMSJ
Injection Date: 05/09/07
Injection Time: 18:39

m/e	Ion Abundance Criteria	Raw Abundance	% Relative Abundance		Pass/Fail
50	8.0 - 40.0% of mass 95	12561	17.7		Pass
75	30.0 - 66.0% of mass 95	35015	49.5		Pass
95	Base peak, 100% relative abundance	70793	100.0		Pass
96	5.0 - 9.0% of mass 95	4662	6.6		Pass
173	Less than 2.0% of mass 174	0	0.0	(0.0) [a]	Pass
174	50.0 - 120.0% of mass 95	74325	105.0		Pass
175	4.0 - 9.0% of mass 174	5105	7.2	(6.9) [a]	Pass
176	93.0 - 101.0% of mass 174	69706	98.5	(93.8) [a]	Pass
177	5.0 - 9.0% of mass 176	5111	7.2	(7.3) [b]	Pass

(a) Value is % of mass 174
(b) Value is % of mass 176

This check applies to the following Samples, MS, MSD, Blanks, and Standards:

Lab Sample ID	Lab File ID	Date Analyzed	Time Analyzed	Hours Lapsed	Client Sample ID
MSJ533-IC533	J10277.D	05/09/07	18:39	00:00	Initial cal 0.2
MSJ533-IC533	J10278.D	05/09/07	19:29	00:50	Initial cal 0.5
MSJ533-IC533	J10279.D	05/09/07	20:18	01:39	Initial cal 40
MSJ533-IC533	J10280.D	05/09/07	21:05	02:26	Initial cal 20
MSJ533-ICC533	J10281.D	05/09/07	21:52	03:13	Initial cal 10
MSJ533-IC533	J10282.D	05/09/07	22:40	04:01	Initial cal 5
MSJ533-IC533	J10283.D	05/09/07	23:27	04:48	Initial cal 2
MSJ533-ICV533	J10286.D	05/10/07	10:33	15:54	Initial cal verification 10

Instrument Performance Check (BFB)

Job Number: M64467
Account: GEI GEI Consultants, Inc.
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

Sample: MSJ533-BFB
Lab File ID: J10286A.D
Instrument ID: GCMSJ
Injection Date: 05/10/07
Injection Time: 10:33

m/e	Ion Abundance Criteria	Raw Abundance	% Relative Abundance		Pass/Fail
50	8.0 - 40.0% of mass 95	58752	17.6		Pass
75	30.0 - 66.0% of mass 95	163184	48.9		Pass
95	Base peak, 100% relative abundance	333840	100.0		Pass
96	5.0 - 9.0% of mass 95	23018	6.9		Pass
173	Less than 2.0% of mass 174	0	0.0	(0.0) [a]	Pass
174	50.0 - 120.0% of mass 95	359152	107.6		Pass
175	4.0 - 9.0% of mass 174	15166	4.5	(4.2) [a]	Pass
176	93.0 - 101.0% of mass 174	348720	104.5	(97.1) [a]	Pass
177	5.0 - 9.0% of mass 176	21654	6.5	(6.2) [b]	Pass

(a) Value is % of mass 174
(b) Value is % of mass 176

This check applies to the following Samples, MS, MSD, Blanks, and Standards:

Lab Sample ID	Lab File ID	Date Analyzed	Time Analyzed	Hours Lapsed	Client Sample ID
MSJ533-CC533	J10286A.D	05/10/07	10:33	00:00	Continuing cal 10
MSJ533-BS	J10286B.D	05/10/07	10:33	00:00	Blank Spike
MSJ533-MB	J10288.D	05/10/07	12:14	01:41	Method Blank
MSJ533-SCC	J10288A.D	05/10/07	12:14	01:41	Summa Cleaning Certification
M64467-1	J10289.D	05/10/07	13:03	02:30	045162-97R FRANK-SS1
ZZZZZZ	J10290.D	05/10/07	13:57	03:24	(unrelated sample)
ZZZZZZ	J10291.D	05/10/07	14:51	04:18	(unrelated sample)
M64467-1DUP	J10293.D	05/10/07	16:47	06:14	Duplicate
ZZZZZZ	J10295.D	05/10/07	18:32	07:59	(unrelated sample)
ZZZZZZ	J10296.D	05/10/07	19:23	08:50	(unrelated sample)
ZZZZZZ	J10297.D	05/10/07	20:12	09:39	(unrelated sample)
ZZZZZZ	J10298.D	05/10/07	20:59	10:26	(unrelated sample)
ZZZZZZ	J10299.D	05/10/07	21:46	11:13	(unrelated sample)
ZZZZZZ	J10300.D	05/10/07	22:33	12:00	(unrelated sample)
ZZZZZZ	J10301.D	05/10/07	23:20	12:47	(unrelated sample)
ZZZZZZ	J10302.D	05/11/07	00:06	13:33	(unrelated sample)
ZZZZZZ	J10303.D	05/11/07	00:51	14:18	(unrelated sample)
MSJ533-SCC	J10305.D	05/11/07	02:21	15:48	Summa Cleaning Certification
MSJ533-SCC	J10306.D	05/11/07	03:06	16:33	Summa Cleaning Certification
MSJ533-SCC	J10307.D	05/11/07	03:50	17:17	Summa Cleaning Certification
MSJ533-SCC	J10308.D	05/11/07	04:35	18:02	Summa Cleaning Certification
MSJ533-SCC	J10309.D	05/11/07	08:22	21:49	Summa Cleaning Certification
MSJ533-SCC	J10310.D	05/11/07	09:07	22:34	Summa Cleaning Certification
MSJ533-SCC	J10311.D	05/11/07	09:52	23:19	Summa Cleaning Certification

Volatile Internal Standard Area Summary

Job Number: M64467
Account: GEI GEI Consultants, Inc.
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

Check Std:	MSJ533-CC533	**Injection Date:**	05/10/07
Lab File ID:	J10286A.D	**Injection Time:**	10:33
Instrument ID:	GCMSJ	**Method:**	TO-15

	IS 1 AREA	RT	IS 2 AREA	RT	IS 3 AREA	RT
Check Std	413189	8.62	1835396	10.44	1640141	14.67
Upper Limit [a]	578465	8.95	2569554	10.77	2296197	15.00
Lower Limit [b]	247913	8.29	1101238	10.11	984085	14.34

Lab Sample ID	IS 1 AREA	RT	IS 2 AREA	RT	IS 3 AREA	RT
MSJ533-BS	413189	8.62	1835396	10.44	1640141	14.67
MSJ533-MB	381318	8.62	1690348	10.44	1515608	14.67
MSJ533-SCC	381318	8.62	1690348	10.44	1515608	14.67
M64467-1	362149	8.62	1604778	10.43	1460059	14.67
ZZZZZZ	362860	8.62	1552293	10.43	1515758	14.66
ZZZZZZ	449251	8.61	2059368	10.43	1706065	14.69
M64467-1DUP	333619	8.60	1340934	10.42	1201491	14.66
ZZZZZZ	386660	8.63	1712562	10.44	1576861	14.67
ZZZZZZ	353894	8.63	1439066	10.44	1260168	14.67
ZZZZZZ	355456	8.66	1514660	10.48	1484790	14.68
ZZZZZZ	389440	8.66	1647308	10.47	1556498	14.67
ZZZZZZ	383722	8.66	1663801	10.46	1538544	14.67
ZZZZZZ	410880	8.62	1772361	10.44	1571230	14.67
ZZZZZZ	363537	8.62	1567838	10.44	1415679	14.67
ZZZZZZ	361622	8.62	1590815	10.44	1460787	14.67
ZZZZZZ	372402	8.62	1622994	10.44	1470779	14.67
MSJ533-SCC	308699	8.62	1227895	10.43	1099290	14.67
MSJ533-SCC	361981	8.62	1631570	10.44	1459024	14.67
MSJ533-SCC	324340	8.62	1392533	10.44	1253669	14.67
MSJ533-SCC	314587	8.62	1276511	10.43	1108567	14.67
MSJ533-SCC	294181	8.62	1243989	10.44	1158956	14.67
MSJ533-SCC	314379	8.62	1255223	10.43	1149666	14.67
MSJ533-SCC	320484	8.63	1338572	10.44	1208850	14.67

IS 1 = Bromochloromethane
IS 2 = 1,4-Difluorobenzene
IS 3 = Chlorobenzene-D5

(a) Upper Limit = +40% of check standard area; Retention time +0.33 minutes.
(b) Lower Limit = -40% of check standard area; Retention time -0.33 minutes.

Initial Calibration Retention Time/Internal Standard Area Summary

Job Number: M64467
Account: GEI GEI Consultants, Inc.
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

Sample Number	Lab File ID	Injected	By	Level	Inst ID	Method	
MSJ533-IC533	J10277.D	05/09/07 18:39	DFT	0.2	GCMSJ	TO-15	Reporting this level
MSJ533-IC533	J10278.D	05/09/07 19:29	DFT	0.5	GCMSJ	TO-15	
MSJ533-IC533	J10279.D	05/09/07 20:18	DFT	40	GCMSJ	TO-15	
MSJ533-IC533	J10280.D	05/09/07 21:05	DFT	20	GCMSJ	TO-15	
MSJ533-ICC533	J10281.D	05/09/07 21:52	DFT	10	GCMSJ	TO-15	
MSJ533-IC533	J10282.D	05/09/07 22:40	DFT	5	GCMSJ	TO-15	
MSJ533-IC533	J10283.D	05/09/07 23:27	DFT	2	GCMSJ	TO-15	

Target Compound	RT (min.)	Istd RT (min.)	Rel RT		Mean Rel RT	Rel RT Range (+/- .06)
Acetone	5.68	8.62	0.659	ok	0.657	0.597-0.717
1,3-Butadiene	4.59	8.62	0.532	ok	0.532	0.472-0.592
Benzene	10.12	10.43	0.970	ok	0.969	0.909-1.029
Bromodichloromethane	11.08	10.43	1.062	ok	1.062	1.002-1.122
Bromoform	15.39	14.67	1.049	ok	1.049	0.989-1.109
Bromomethane	4.85	8.62	0.563	ok	0.563	0.503-0.623
Bromoethene	5.38	8.62	0.624	ok	0.624	0.564-0.684
Carbon disulfide	6.85	8.62	0.795	ok	0.794	0.734-0.854
Chlorobenzene	14.71	14.67	1.003	ok	1.003	0.943-1.063
Chloroethane	5.04	8.62	0.585	ok	0.583	0.523-0.643
Chloroform	8.75	8.62	1.015	ok	1.014	0.954-1.074
Chloromethane	4.22	8.62	0.490	ok	0.490	0.430-0.550
3-Chloropropene	6.66	8.62	0.773	ok	0.772	0.712-0.832
Carbon tetrachloride	10.25	8.62	1.189	ok	1.190	1.130-1.250
Cyclohexane	10.37	10.43	0.994	ok	0.993	0.933-1.053
1,1-Dichloroethane	7.66	8.62	0.889	ok	0.888	0.828-0.948
1,1-Dichloroethylene	6.42	8.62	0.745	ok	0.745	0.685-0.805
1,2-Dibromoethane	13.57	14.67	0.925	ok	0.925	0.865-0.985
1,2-Dichloroethane	9.45	8.62	1.096	ok	1.096	1.036-1.156
1,2-Dichloropropane	10.90	10.43	1.045	ok	1.044	0.984-1.104
Dichlorodifluoromethane	4.07	8.62	0.472	ok	0.472	0.412-0.532
Dibromochloromethane	13.32	14.67	0.908	ok	0.908	0.848-0.968
trans-1,2-Dichloroethylene	7.46	8.62	0.865	ok	0.865	0.805-0.925
cis-1,2-Dichloroethylene	8.46	8.62	0.981	ok	0.981	0.921-1.041
cis-1,3-Dichloropropene	11.94	10.43	1.145	ok	1.144	1.084-1.204
trans-1,3-Dichloropropene	12.43	10.43	1.192	ok	1.192	1.132-1.252
Ethylbenzene	15.09	14.67	1.029	ok	1.029	0.969-1.089
Ethyl Acetate	8.72	8.62	1.012	ok	1.009	0.949-1.069
Freon 113	6.79	8.62	0.788	ok	0.788	0.728-0.848
Freon 114	4.32	8.62	0.501	ok	0.501	0.441-0.561
Heptane	11.36	10.43	1.089	ok	1.088	1.028-1.148
Hexane	8.63	8.62	1.001	ok	1.001	0.941-1.061
Isopropyl Alcohol	6.05	8.62	0.702	ok	0.692	0.632-0.752
Methyl ethyl ketone	8.11	8.62	0.941	ok	0.937	0.877-0.997
Methyl Tert Butyl Ether	7.76	8.62	0.900	ok	0.897	0.837-0.957
Pentane	6.14	8.62	0.712	ok	0.712	0.652-0.772
1,1,1-Trichloroethane	9.68	8.62	1.123	ok	1.123	1.063-1.183
1,1,2,2-Tetrachloroethane	15.79	14.67	1.076	ok	1.077	1.017-1.137
1,1,2-Trichloroethane	12.62	10.43	1.210	ok	1.209	1.149-1.269
2,2,4-Trimethylpentane	11.12	10.43	1.066	ok	1.065	1.005-1.125
Tetrachloroethylene	14.01	14.67	0.955	ok	0.955	0.895-1.015

Initial Calibration Retention Time/Internal Standard Area Summary

Job Number: M64467
Account: GEI GEI Consultants, Inc.
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

Sample Number	Lab File ID	Injected	By	Level	Inst ID	Method	
MSJ533-IC533	J10277.D	05/09/07 18:39	DFT	0.2	GCMSJ	TO-15	Reporting this level
MSJ533-IC533	J10278.D	05/09/07 19:29	DFT	0.5	GCMSJ	TO-15	
MSJ533-IC533	J10279.D	05/09/07 20:18	DFT	40	GCMSJ	TO-15	
MSJ533-IC533	J10280.D	05/09/07 21:05	DFT	20	GCMSJ	TO-15	
MSJ533-ICC533	J10281.D	05/09/07 21:52	DFT	10	GCMSJ	TO-15	
MSJ533-IC533	J10282.D	05/09/07 22:40	DFT	5	GCMSJ	TO-15	
MSJ533-IC533	J10283.D	05/09/07 23:27	DFT	2	GCMSJ	TO-15	

Target Compound	RT (min.)	Istd RT (min.)	Rel RT		Mean Rel RT	Rel RT Range (+/- .06)
Toluene	12.88	10.43	1.235	ok	1.234	1.174-1.294
Trichloroethylene	11.12	10.43	1.066	ok	1.065	1.005-1.125
Trichlorofluoromethane	5.79	8.62	0.672	ok	0.672	0.612-0.732
Vinyl chloride	4.46	8.62	0.517	ok	0.516	0.456-0.576
m,p-Xylene	15.29	14.67	1.042	ok	1.042	0.982-1.102

Internal Standard	RT (min.)		Mean RT(min.)	RT Range (+/- 0.33)	Area		Mean Area	Area Range (+/- 40 %)
Bromochloromethane	8.62	ok	8.62	8.29-8.95	344690	ok	371623	222974-520272
1,4-Difluorobenzene	10.43	ok	10.44	10.11-10.77	1460896	ok	1631886	979132-2284640
Chlorobenzene-D5	14.67	ok	14.67	14.34-15.00	1214337	ok	1431182	858709-2003655

Initial Calibration Retention Time/Internal Standard Area Summary

Job Number: M64467
Account: GEI GEI Consultants, Inc.
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

Sample Number	Lab File ID	Injected	By	Level	Inst ID	Method	
MSJ533-IC533	J10277.D	05/09/07 18:39	DFT	0.2	GCMSJ	TO-15	
MSJ533-IC533	J10278.D	05/09/07 19:29	DFT	0.5	GCMSJ	TO-15	Reporting this level
MSJ533-IC533	J10279.D	05/09/07 20:18	DFT	40	GCMSJ	TO-15	
MSJ533-IC533	J10280.D	05/09/07 21:05	DFT	20	GCMSJ	TO-15	
MSJ533-ICC533	J10281.D	05/09/07 21:52	DFT	10	GCMSJ	TO-15	
MSJ533-IC533	J10282.D	05/09/07 22:40	DFT	5	GCMSJ	TO-15	
MSJ533-IC533	J10283.D	05/09/07 23:27	DFT	2	GCMSJ	TO-15	

Target Compound	RT (min.)	Istd RT (min.)	Rel RT		Mean Rel RT	Rel RT Range (+/- .06)
Acetone	5.67	8.62	0.658	ok	0.657	0.597-0.717
1,3-Butadiene	4.59	8.62	0.532	ok	0.532	0.472-0.592
Benzene	10.12	10.44	0.969	ok	0.969	0.909-1.029
Bromodichloromethane	11.09	10.44	1.062	ok	1.062	1.002-1.122
Bromoform	15.39	14.67	1.049	ok	1.049	0.989-1.109
Bromomethane	4.86	8.62	0.564	ok	0.563	0.503-0.623
Bromoethene	5.38	8.62	0.624	ok	0.624	0.564-0.684
Benzyl Chloride	17.88	14.67	1.219	ok	1.219	1.159-1.279
Carbon disulfide	6.84	8.62	0.794	ok	0.794	0.734-0.854
Chlorobenzene	14.71	14.67	1.003	ok	1.003	0.943-1.063
Chloroethane	5.03	8.62	0.584	ok	0.583	0.523-0.643
Chloroform	8.74	8.62	1.014	ok	1.014	0.954-1.074
Chloromethane	4.22	8.62	0.490	ok	0.490	0.430-0.550
3-Chloropropene	6.65	8.62	0.771	ok	0.772	0.712-0.832
2-Chlorotoluene	16.97	14.67	1.157	ok	1.157	1.097-1.217
Carbon tetrachloride	10.26	8.62	1.190	ok	1.190	1.130-1.250
Cyclohexane	10.37	10.44	0.993	ok	0.993	0.933-1.053
1,1-Dichloroethane	7.66	8.62	0.889	ok	0.888	0.828-0.948
1,1-Dichloroethylene	6.42	8.62	0.745	ok	0.745	0.685-0.805
1,2-Dibromoethane	13.57	14.67	0.925	ok	0.925	0.865-0.985
1,2-Dichloroethane	9.45	8.62	1.096	ok	1.096	1.036-1.156
1,2-Dichloropropane	10.89	10.44	1.043	ok	1.044	0.984-1.104
1,4-Dioxane	11.24	10.44	1.077	ok	1.071	1.011-1.131
Dichlorodifluoromethane	4.06	8.62	0.471	ok	0.472	0.412-0.532
Dibromochloromethane	13.32	14.67	0.908	ok	0.908	0.848-0.968
trans-1,2-Dichloroethylene	7.45	8.62	0.864	ok	0.865	0.805-0.925
cis-1,2-Dichloroethylene	8.46	8.62	0.981	ok	0.981	0.921-1.041
cis-1,3-Dichloropropene	11.94	10.44	1.144	ok	1.144	1.084-1.204
m-Dichlorobenzene	17.90	14.67	1.220	ok	1.220	1.160-1.280
o-Dichlorobenzene	18.37	14.67	1.252	ok	1.252	1.192-1.312
p-Dichlorobenzene	17.98	14.67	1.226	ok	1.226	1.166-1.286
trans-1,3-Dichloropropene	12.44	10.44	1.192	ok	1.192	1.132-1.252
Ethanol	5.23	8.62	0.607	ok	0.604	0.544-0.664
Ethylbenzene	15.09	14.67	1.029	ok	1.029	0.969-1.089
Ethyl Acetate	8.70	8.62	1.009	ok	1.009	0.949-1.069
4-Ethyltoluene	17.17	14.67	1.170	ok	1.170	1.110-1.230
Freon 113	6.79	8.62	0.788	ok	0.788	0.728-0.848
Freon 114	4.32	8.62	0.501	ok	0.501	0.441-0.561
Heptane	11.36	10.44	1.088	ok	1.088	1.028-1.148
Hexachlorobutadiene	20.77	14.67	1.416	ok	1.415	1.355-1.475
Hexane	8.63	8.62	1.001	ok	1.001	0.941-1.061

Initial Calibration Retention Time/Internal Standard Area Summary

Job Number: M64467
Account: GEI GEI Consultants, Inc.
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

Sample Number	Lab File ID	Injected	By	Level	Inst ID	Method	
MSJ533-IC533	J10277.D	05/09/07 18:39	DFT	0.2	GCMSJ	TO-15	
MSJ533-IC533	J10278.D	05/09/07 19:29	DFT	0.5	GCMSJ	TO-15	Reporting this level
MSJ533-IC533	J10279.D	05/09/07 20:18	DFT	40	GCMSJ	TO-15	
MSJ533-IC533	J10280.D	05/09/07 21:05	DFT	20	GCMSJ	TO-15	
MSJ533-ICC533	J10281.D	05/09/07 21:52	DFT	10	GCMSJ	TO-15	
MSJ533-IC533	J10282.D	05/09/07 22:40	DFT	5	GCMSJ	TO-15	
MSJ533-IC533	J10283.D	05/09/07 23:27	DFT	2	GCMSJ	TO-15	

Target Compound	RT (min.)	Istd RT (min.)	Rel RT		Mean Rel RT	Rel RT Range (+/- .06)
2-Hexanone	13.20	14.67	0.900	ok	0.898	0.838-0.958
Isopropylbenzene	16.43	14.67	1.120	ok	1.120	1.060-1.180
Isopropyl Alcohol	5.98	8.62	0.694	ok	0.692	0.632-0.752
Methylene chloride	6.55	8.62	0.760	ok	0.760	0.700-0.820
Methyl ethyl ketone	8.08	8.62	0.937	ok	0.937	0.877-0.997
Methyl Isobutyl Ketone	12.01	10.44	1.150	ok	1.149	1.089-1.209
Methyl Tert Butyl Ether	7.74	8.62	0.898	ok	0.897	0.837-0.957
Nonane	15.97	14.67	1.089	ok	1.089	1.029-1.149
Pentane	6.14	8.62	0.712	ok	0.712	0.652-0.772
Propylene	3.99	8.62	0.463	ok	0.463	0.403-0.523
Styrene	15.66	14.67	1.067	ok	1.068	1.008-1.128
1,1,1-Trichloroethane	9.68	8.62	1.123	ok	1.123	1.063-1.183
1,1,2,2-Tetrachloroethane	15.79	14.67	1.076	ok	1.077	1.017-1.137
1,1,2-Trichloroethane	12.62	10.44	1.209	ok	1.209	1.149-1.269
1,2,4-Trichlorobenzene	20.28	14.67	1.382	ok	1.382	1.322-1.442
1,2,4-Trimethylbenzene	17.72	14.67	1.208	ok	1.208	1.148-1.268
1,3,5-Trimethylbenzene	17.25	14.67	1.176	ok	1.176	1.116-1.236
2,2,4-Trimethylpentane	11.12	10.44	1.065	ok	1.065	1.005-1.125
Tertiary Butyl Alcohol	6.60	8.62	0.766	ok	0.760	0.700-0.820
Tetrachloroethylene	14.01	14.67	0.955	ok	0.955	0.895-1.015
Tetrahydrofuran	9.23	8.62	1.071	ok	1.066	1.006-1.126
Toluene	12.89	10.44	1.235	ok	1.234	1.174-1.294
Trichloroethylene	11.11	10.44	1.064	ok	1.065	1.005-1.125
Trichlorofluoromethane	5.80	8.62	0.673	ok	0.672	0.612-0.732
Vinyl chloride	4.44	8.62	0.515	ok	0.516	0.456-0.576
Vinyl Acetate	7.79	8.62	0.904	ok	0.905	0.845-0.965
m,p-Xylene	15.27	14.67	1.041	ok	1.042	0.982-1.102
o-Xylene	15.77	14.67	1.075	ok	1.076	1.016-1.136

Internal Standard	RT (min.)		Mean RT(min.)	RT Range (+/- 0.33)	Area		Mean Area	Area Range (+/- 40 %)
Bromochloromethane	8.62	ok	8.62	8.29-8.95	303253	ok	371623	222974-520272
1,4-Difluorobenzene	10.44	ok	10.44	10.11-10.77	1271837	ok	1631886	979132-2284640
Chlorobenzene-D5	14.67	ok	14.67	14.34-15.00	1022579	ok	1431182	858709-2003655

Initial Calibration Retention Time/Internal Standard Area Summary

Job Number: M64467
Account: GEI GEI Consultants, Inc.
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

Sample Number	Lab File ID	Injected	By	Level	Inst ID	Method	
MSJ533-IC533	J10277.D	05/09/07 18:39	DFT	0.2	GCMSJ	TO-15	
MSJ533-IC533	J10278.D	05/09/07 19:29	DFT	0.5	GCMSJ	TO-15	
MSJ533-IC533	J10279.D	05/09/07 20:18	DFT	40	GCMSJ	TO-15	Reporting this level
MSJ533-IC533	J10280.D	05/09/07 21:05	DFT	20	GCMSJ	TO-15	
MSJ533-ICC533	J10281.D	05/09/07 21:52	DFT	10	GCMSJ	TO-15	
MSJ533-IC533	J10282.D	05/09/07 22:40	DFT	5	GCMSJ	TO-15	
MSJ533-IC533	J10283.D	05/09/07 23:27	DFT	2	GCMSJ	TO-15	

Target Compound	RT (min.)	Istd RT (min.)	Rel RT		Mean Rel RT	Rel RT Range (+/- .06)
Acetone	5.65	8.63	0.655	ok	0.657	0.597-0.717
1,3-Butadiene	4.58	8.63	0.531	ok	0.532	0.472-0.592
Benzene	10.13	10.45	0.969	ok	0.969	0.909-1.029
Bromodichloromethane	11.11	10.45	1.063	ok	1.062	1.002-1.122
Bromoform	15.42	14.68	1.050	ok	1.049	0.989-1.109
Bromomethane	4.86	8.63	0.563	ok	0.563	0.503-0.623
Bromoethene	5.38	8.63	0.623	ok	0.624	0.564-0.684
Benzyl Chloride	17.90	14.68	1.219	ok	1.219	1.159-1.279
Carbon disulfide	6.85	8.63	0.794	ok	0.794	0.734-0.854
Chlorobenzene	14.73	14.68	1.003	ok	1.003	0.943-1.063
Chloroethane	5.02	8.63	0.582	ok	0.583	0.523-0.643
Chloroform	8.76	8.63	1.015	ok	1.014	0.954-1.074
Chloromethane	4.23	8.63	0.490	ok	0.490	0.430-0.550
3-Chloropropene	6.66	8.63	0.772	ok	0.772	0.712-0.832
2-Chlorotoluene	17.00	14.68	1.158	ok	1.157	1.097-1.217
Carbon tetrachloride	10.26	8.63	1.189	ok	1.190	1.130-1.250
Cyclohexane	10.38	10.45	0.993	ok	0.993	0.933-1.053
1,1-Dichloroethane	7.66	8.63	0.888	ok	0.888	0.828-0.948
1,1-Dichloroethylene	6.43	8.63	0.745	ok	0.745	0.685-0.805
1,2-Dibromoethane	13.59	14.68	0.926	ok	0.925	0.865-0.985
1,2-Dichloroethane	9.47	8.63	1.097	ok	1.096	1.036-1.156
1,2-Dichloropropane	10.91	10.45	1.044	ok	1.044	0.984-1.104
1,4-Dioxane	11.18	10.45	1.070	ok	1.071	1.011-1.131
Dichlorodifluoromethane	4.06	8.63	0.470	ok	0.472	0.412-0.532
Dibromochloromethane	13.35	14.68	0.909	ok	0.908	0.848-0.968
trans-1,2-Dichloroethylene	7.46	8.63	0.864	ok	0.865	0.805-0.925
cis-1,2-Dichloroethylene	8.47	8.63	0.981	ok	0.981	0.921-1.041
cis-1,3-Dichloropropene	11.95	10.45	1.144	ok	1.144	1.084-1.204
m-Dichlorobenzene	17.92	14.68	1.221	ok	1.220	1.160-1.280
o-Dichlorobenzene	18.38	14.68	1.252	ok	1.252	1.192-1.312
p-Dichlorobenzene	17.99	14.68	1.225	ok	1.226	1.166-1.286
trans-1,3-Dichloropropene	12.46	10.45	1.192	ok	1.192	1.132-1.252
Ethanol	5.20	8.63	0.603	ok	0.604	0.544-0.664
Ethylbenzene	15.11	14.68	1.029	ok	1.029	0.969-1.089
Ethyl Acetate	8.70	8.63	1.008	ok	1.009	0.949-1.069
4-Ethyltoluene	17.18	14.68	1.170	ok	1.170	1.110-1.230
Freon 113	6.79	8.63	0.787	ok	0.788	0.728-0.848
Freon 114	4.32	8.63	0.501	ok	0.501	0.441-0.561
Heptane	11.37	10.45	1.088	ok	1.088	1.028-1.148
Hexachlorobutadiene	20.76	14.68	1.414	ok	1.415	1.355-1.475
Hexane	8.63	8.63	1.000	ok	1.001	0.941-1.061

Initial Calibration Retention Time/Internal Standard Area Summary

Job Number: M64467
Account: GEI GEI Consultants, Inc.
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

Sample Number	Lab File ID	Injected	By	Level	Inst ID	Method	
MSJ533-IC533	J10277.D	05/09/07 18:39	DFT	0.2	GCMSJ	TO-15	
MSJ533-IC533	J10278.D	05/09/07 19:29	DFT	0.5	GCMSJ	TO-15	
MSJ533-IC533	J10279.D	05/09/07 20:18	DFT	40	GCMSJ	TO-15	Reporting this level
MSJ533-IC533	J10280.D	05/09/07 21:05	DFT	20	GCMSJ	TO-15	
MSJ533-ICC533	J10281.D	05/09/07 21:52	DFT	10	GCMSJ	TO-15	
MSJ533-IC533	J10282.D	05/09/07 22:40	DFT	5	GCMSJ	TO-15	
MSJ533-IC533	J10283.D	05/09/07 23:27	DFT	2	GCMSJ	TO-15	

Target Compound	RT (min.)	Istd RT (min.)	Rel RT		Mean Rel RT	Rel RT Range (+/- .06)
2-Hexanone	13.18	14.68	0.898	ok	0.898	0.838-0.958
Isopropylbenzene	16.44	14.68	1.120	ok	1.120	1.060-1.180
Isopropyl Alcohol	5.97	8.63	0.692	ok	0.692	0.632-0.752
Methylene chloride	6.55	8.63	0.759	ok	0.760	0.700-0.820
Methyl ethyl ketone	8.08	8.63	0.936	ok	0.937	0.877-0.997
Methyl Isobutyl Ketone	12.01	10.45	1.149	ok	1.149	1.089-1.209
Methyl Tert Butyl Ether	7.73	8.63	0.896	ok	0.897	0.837-0.957
Nonane	15.99	14.68	1.089	ok	1.089	1.029-1.149
Pentane	6.14	8.63	0.711	ok	0.712	0.652-0.772
Propylene	3.99	8.63	0.462	ok	0.463	0.403-0.523
Styrene	15.69	14.68	1.069	ok	1.068	1.008-1.128
1,1,1-Trichloroethane	9.69	8.63	1.123	ok	1.123	1.063-1.183
1,1,2,2-Tetrachloroethane	15.82	14.68	1.078	ok	1.077	1.017-1.137
1,1,2-Trichloroethane	12.65	10.45	1.211	ok	1.209	1.149-1.269
1,2,4-Trichlorobenzene	20.27	14.68	1.381	ok	1.382	1.322-1.442
1,2,4-Trimethylbenzene	17.73	14.68	1.208	ok	1.208	1.148-1.268
1,3,5-Trimethylbenzene	17.27	14.68	1.176	ok	1.176	1.116-1.236
2,2,4-Trimethylpentane	11.13	10.45	1.065	ok	1.065	1.005-1.125
Tertiary Butyl Alcohol	6.57	8.63	0.761	ok	0.760	0.700-0.820
Tetrachloroethylene	14.02	14.68	0.955	ok	0.955	0.895-1.015
Tetrahydrofuran	9.18	8.63	1.064	ok	1.066	1.006-1.126
Toluene	12.90	10.45	1.234	ok	1.234	1.174-1.294
Trichloroethylene	11.13	10.45	1.065	ok	1.065	1.005-1.125
Trichlorofluoromethane	5.80	8.63	0.672	ok	0.672	0.612-0.732
Vinyl chloride	4.45	8.63	0.516	ok	0.516	0.456-0.576
Vinyl Acetate	7.81	8.63	0.905	ok	0.905	0.845-0.965
m,p-Xylene	15.29	14.68	1.042	ok	1.042	0.982-1.102
o-Xylene	15.80	14.68	1.076	ok	1.076	1.016-1.136

Internal Standard	RT (min.)		Mean RT(min.)	RT Range (+/- 0.33)	Area		Mean Area	Area Range (+/- 40 %)
Bromochloromethane	8.63	ok	8.62	8.29-8.95	404357	ok	371623	222974-520272
1,4-Difluorobenzene	10.45	ok	10.44	10.11-10.77	1793963	ok	1631886	979132-2284640
Chlorobenzene-D5	14.68	ok	14.67	14.34-15.00	1690369	ok	1431182	858709-2003655

Initial Calibration Retention Time/Internal Standard Area Summary

Job Number: M64467
Account: GEI GEI Consultants, Inc.
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

Sample Number	Lab File ID	Injected	By	Level	Inst ID	Method	
MSJ533-IC533	J10277.D	05/09/07 18:39	DFT	0.2	GCMSJ	TO-15	
MSJ533-IC533	J10278.D	05/09/07 19:29	DFT	0.5	GCMSJ	TO-15	
MSJ533-IC533	J10279.D	05/09/07 20:18	DFT	40	GCMSJ	TO-15	
MSJ533-IC533	J10280.D	05/09/07 21:05	DFT	20	GCMSJ	TO-15	Reporting this level
MSJ533-ICC533	J10281.D	05/09/07 21:52	DFT	10	GCMSJ	TO-15	
MSJ533-IC533	J10282.D	05/09/07 22:40	DFT	5	GCMSJ	TO-15	
MSJ533-IC533	J10283.D	05/09/07 23:27	DFT	2	GCMSJ	TO-15	

Target Compound	RT (min.)	Istd RT (min.)	Rel RT		Mean Rel RT	Rel RT Range (+/- .06)
Acetone	5.65	8.63	0.655	ok	0.657	0.597-0.717
1,3-Butadiene	4.59	8.63	0.532	ok	0.532	0.472-0.592
Benzene	10.13	10.45	0.969	ok	0.969	0.909-1.029
Bromodichloromethane	11.09	10.45	1.061	ok	1.062	1.002-1.122
Bromoform	15.41	14.68	1.050	ok	1.049	0.989-1.109
Bromomethane	4.86	8.63	0.563	ok	0.563	0.503-0.623
Bromoethene	5.39	8.63	0.625	ok	0.624	0.564-0.684
Benzyl Chloride	17.89	14.68	1.219	ok	1.219	1.159-1.279
Carbon disulfide	6.85	8.63	0.794	ok	0.794	0.734-0.854
Chlorobenzene	14.72	14.68	1.003	ok	1.003	0.943-1.063
Chloroethane	5.03	8.63	0.583	ok	0.583	0.523-0.643
Chloroform	8.75	8.63	1.014	ok	1.014	0.954-1.074
Chloromethane	4.23	8.63	0.490	ok	0.490	0.430-0.550
3-Chloropropene	6.66	8.63	0.772	ok	0.772	0.712-0.832
2-Chlorotoluene	16.99	14.68	1.157	ok	1.157	1.097-1.217
Carbon tetrachloride	10.26	8.63	1.189	ok	1.190	1.130-1.250
Cyclohexane	10.37	10.45	0.992	ok	0.993	0.933-1.053
1,1-Dichloroethane	7.66	8.63	0.888	ok	0.888	0.828-0.948
1,1-Dichloroethylene	6.43	8.63	0.745	ok	0.745	0.685-0.805
1,2-Dibromoethane	13.58	14.68	0.925	ok	0.925	0.865-0.985
1,2-Dichloroethane	9.46	8.63	1.096	ok	1.096	1.036-1.156
1,2-Dichloropropane	10.90	10.45	1.043	ok	1.044	0.984-1.104
1,4-Dioxane	11.17	10.45	1.069	ok	1.071	1.011-1.131
Dichlorodifluoromethane	4.07	8.63	0.472	ok	0.472	0.412-0.532
Dibromochloromethane	13.33	14.68	0.908	ok	0.908	0.848-0.968
trans-1,2-Dichloroethylene	7.46	8.63	0.864	ok	0.865	0.805-0.925
cis-1,2-Dichloroethylene	8.46	8.63	0.980	ok	0.981	0.921-1.041
cis-1,3-Dichloropropene	11.94	10.45	1.143	ok	1.144	1.084-1.204
m-Dichlorobenzene	17.91	14.68	1.220	ok	1.220	1.160-1.280
o-Dichlorobenzene	18.38	14.68	1.252	ok	1.252	1.192-1.312
p-Dichlorobenzene	17.99	14.68	1.225	ok	1.226	1.166-1.286
trans-1,3-Dichloropropene	12.45	10.45	1.191	ok	1.192	1.132-1.252
Ethanol	5.20	8.63	0.603	ok	0.604	0.544-0.664
Ethylbenzene	15.10	14.68	1.029	ok	1.029	0.969-1.089
Ethyl Acetate	8.69	8.63	1.007	ok	1.009	0.949-1.069
4-Ethyltoluene	17.18	14.68	1.170	ok	1.170	1.110-1.230
Freon 113	6.79	8.63	0.787	ok	0.788	0.728-0.848
Freon 114	4.33	8.63	0.502	ok	0.501	0.441-0.561
Heptane	11.37	10.45	1.088	ok	1.088	1.028-1.148
Hexachlorobutadiene	20.77	14.68	1.415	ok	1.415	1.355-1.475
Hexane	8.64	8.63	1.001	ok	1.001	0.941-1.061

Initial Calibration Retention Time/Internal Standard Area Summary

Job Number: M64467
Account: GEI GEI Consultants, Inc.
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

Sample Number	Lab File ID	Injected	By	Level	Inst ID	Method	
MSJ533-IC533	J10277.D	05/09/07 18:39	DFT	0.2	GCMSJ	TO-15	
MSJ533-IC533	J10278.D	05/09/07 19:29	DFT	0.5	GCMSJ	TO-15	
MSJ533-IC533	J10279.D	05/09/07 20:18	DFT	40	GCMSJ	TO-15	
MSJ533-IC533	J10280.D	05/09/07 21:05	DFT	20	GCMSJ	TO-15	Reporting this level
MSJ533-ICC533	J10281.D	05/09/07 21:52	DFT	10	GCMSJ	TO-15	
MSJ533-IC533	J10282.D	05/09/07 22:40	DFT	5	GCMSJ	TO-15	
MSJ533-IC533	J10283.D	05/09/07 23:27	DFT	2	GCMSJ	TO-15	

Target Compound	RT (min.)	Istd RT (min.)	Rel RT		Mean Rel RT	Rel RT Range (+/- .06)
2-Hexanone	13.16	14.68	0.896	ok	0.898	0.838-0.958
Isopropylbenzene	16.43	14.68	1.119	ok	1.120	1.060-1.180
Isopropyl Alcohol	5.94	8.63	0.688	ok	0.692	0.632-0.752
Methylene chloride	6.55	8.63	0.759	ok	0.760	0.700-0.820
Methyl ethyl ketone	8.07	8.63	0.935	ok	0.937	0.877-0.997
Methyl Isobutyl Ketone	11.99	10.45	1.147	ok	1.149	1.089-1.209
Methyl Tert Butyl Ether	7.73	8.63	0.896	ok	0.897	0.837-0.957
Nonane	15.98	14.68	1.089	ok	1.089	1.029-1.149
Pentane	6.14	8.63	0.711	ok	0.712	0.652-0.772
Propylene	3.99	8.63	0.462	ok	0.463	0.403-0.523
Styrene	15.68	14.68	1.068	ok	1.068	1.008-1.128
1,1,1-Trichloroethane	9.69	8.63	1.123	ok	1.123	1.063-1.183
1,1,2,2-Tetrachloroethane	15.80	14.68	1.076	ok	1.077	1.017-1.137
1,1,2-Trichloroethane	12.63	10.45	1.209	ok	1.209	1.149-1.269
1,2,4-Trichlorobenzene	20.27	14.68	1.381	ok	1.382	1.322-1.442
1,2,4-Trimethylbenzene	17.73	14.68	1.208	ok	1.208	1.148-1.268
1,3,5-Trimethylbenzene	17.26	14.68	1.176	ok	1.176	1.116-1.236
2,2,4-Trimethylpentane	11.12	10.45	1.064	ok	1.065	1.005-1.125
Tertiary Butyl Alcohol	6.54	8.63	0.758	ok	0.760	0.700-0.820
Tetrachloroethylene	14.02	14.68	0.955	ok	0.955	0.895-1.015
Tetrahydrofuran	9.17	8.63	1.063	ok	1.066	1.006-1.126
Toluene	12.89	10.45	1.233	ok	1.234	1.174-1.294
Trichloroethylene	11.12	10.45	1.064	ok	1.065	1.005-1.125
Trichlorofluoromethane	5.79	8.63	0.671	ok	0.672	0.612-0.732
Vinyl chloride	4.45	8.63	0.516	ok	0.516	0.456-0.576
Vinyl Acetate	7.80	8.63	0.904	ok	0.905	0.845-0.965
m,p-Xylene	15.29	14.68	1.042	ok	1.042	0.982-1.102
o-Xylene	15.79	14.68	1.076	ok	1.076	1.016-1.136

Internal Standard	RT (min.)		Mean RT(min.)	RT Range (+/- 0.33)	Area		Mean Area	Area Range (+/- 40 %)
Bromochloromethane	8.63	ok	8.62	8.29-8.95	404934	ok	371623	222974-520272
1,4-Difluorobenzene	10.45	ok	10.44	10.11-10.77	1815477	ok	1631886	979132-2284640
Chlorobenzene-D5	14.68	ok	14.67	14.34-15.00	1676330	ok	1431182	858709-2003655

Initial Calibration Retention Time/Internal Standard Area Summary

Job Number: M64467
Account: GEI GEI Consultants, Inc.
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

Sample Number	Lab File ID	Injected	By	Level	Inst ID	Method	
MSJ533-IC533	J10277.D	05/09/07 18:39	DFT	0.2	GCMSJ	TO-15	
MSJ533-IC533	J10278.D	05/09/07 19:29	DFT	0.5	GCMSJ	TO-15	
MSJ533-IC533	J10279.D	05/09/07 20:18	DFT	40	GCMSJ	TO-15	
MSJ533-IC533	J10280.D	05/09/07 21:05	DFT	20	GCMSJ	TO-15	
MSJ533-ICC533	J10281.D	05/09/07 21:52	DFT	10	GCMSJ	TO-15	Reporting this level
MSJ533-IC533	J10282.D	05/09/07 22:40	DFT	5	GCMSJ	TO-15	
MSJ533-IC533	J10283.D	05/09/07 23:27	DFT	2	GCMSJ	TO-15	

Target Compound	RT (min.)	Istd RT (min.)	Rel RT		Mean Rel RT	Rel RT Range (+/- .06)
Acetone	5.66	8.62	0.657	ok	0.657	0.597-0.717
1,3-Butadiene	4.58	8.62	0.531	ok	0.532	0.472-0.592
Benzene	10.12	10.44	0.969	ok	0.969	0.909-1.029
Bromodichloromethane	11.09	10.44	1.062	ok	1.062	1.002-1.122
Bromoform	15.40	14.67	1.050	ok	1.049	0.989-1.109
Bromomethane	4.86	8.62	0.564	ok	0.563	0.503-0.623
Bromoethene	5.38	8.62	0.624	ok	0.624	0.564-0.684
Benzyl Chloride	17.89	14.67	1.219	ok	1.219	1.159-1.279
Carbon disulfide	6.85	8.62	0.795	ok	0.794	0.734-0.854
Chlorobenzene	14.72	14.67	1.003	ok	1.003	0.943-1.063
Chloroethane	5.03	8.62	0.584	ok	0.583	0.523-0.643
Chloroform	8.74	8.62	1.014	ok	1.014	0.954-1.074
Chloromethane	4.23	8.62	0.491	ok	0.490	0.430-0.550
3-Chloropropene	6.66	8.62	0.773	ok	0.772	0.712-0.832
2-Chlorotoluene	16.98	14.67	1.157	ok	1.157	1.097-1.217
Carbon tetrachloride	10.26	8.62	1.190	ok	1.190	1.130-1.250
Cyclohexane	10.37	10.44	0.993	ok	0.993	0.933-1.053
1,1-Dichloroethane	7.66	8.62	0.889	ok	0.888	0.828-0.948
1,1-Dichloroethylene	6.42	8.62	0.745	ok	0.745	0.685-0.805
1,2-Dibromoethane	13.57	14.67	0.925	ok	0.925	0.865-0.985
1,2-Dichloroethane	9.45	8.62	1.096	ok	1.096	1.036-1.156
1,2-Dichloropropane	10.90	10.44	1.044	ok	1.044	0.984-1.104
1,4-Dioxane	11.17	10.44	1.070	ok	1.071	1.011-1.131
Dichlorodifluoromethane	4.06	8.62	0.471	ok	0.472	0.412-0.532
Dibromochloromethane	13.33	14.67	0.909	ok	0.908	0.848-0.968
trans-1,2-Dichloroethylene	7.45	8.62	0.864	ok	0.865	0.805-0.925
cis-1,2-Dichloroethylene	8.46	8.62	0.981	ok	0.981	0.921-1.041
cis-1,3-Dichloropropene	11.94	10.44	1.144	ok	1.144	1.084-1.204
m-Dichlorobenzene	17.91	14.67	1.221	ok	1.220	1.160-1.280
o-Dichlorobenzene	18.37	14.67	1.252	ok	1.252	1.192-1.312
p-Dichlorobenzene	17.98	14.67	1.226	ok	1.226	1.166-1.286
trans-1,3-Dichloropropene	12.44	10.44	1.192	ok	1.192	1.132-1.252
Ethanol	5.21	8.62	0.604	ok	0.604	0.544-0.664
Ethylbenzene	15.09	14.67	1.029	ok	1.029	0.969-1.089
Ethyl Acetate	8.69	8.62	1.008	ok	1.009	0.949-1.069
4-Ethyltoluene	17.17	14.67	1.170	ok	1.170	1.110-1.230
Freon 113	6.79	8.62	0.788	ok	0.788	0.728-0.848
Freon 114	4.32	8.62	0.501	ok	0.501	0.441-0.561
Heptane	11.36	10.44	1.088	ok	1.088	1.028-1.148
Hexachlorobutadiene	20.76	14.67	1.415	ok	1.415	1.355-1.475
Hexane	8.63	8.62	1.001	ok	1.001	0.941-1.061

Initial Calibration Retention Time/Internal Standard Area Summary

Job Number: M64467
Account: GEI GEI Consultants, Inc.
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

Sample Number	Lab File ID	Injected	By	Level	Inst ID	Method	
MSJ533-IC533	J10277.D	05/09/07 18:39	DFT	0.2	GCMSJ	TO-15	
MSJ533-IC533	J10278.D	05/09/07 19:29	DFT	0.5	GCMSJ	TO-15	
MSJ533-IC533	J10279.D	05/09/07 20:18	DFT	40	GCMSJ	TO-15	
MSJ533-IC533	J10280.D	05/09/07 21:05	DFT	20	GCMSJ	TO-15	
MSJ533-ICC533	J10281.D	05/09/07 21:52	DFT	10	GCMSJ	TO-15	Reporting this level
MSJ533-IC533	J10282.D	05/09/07 22:40	DFT	5	GCMSJ	TO-15	
MSJ533-IC533	J10283.D	05/09/07 23:27	DFT	2	GCMSJ	TO-15	

Target Compound	RT (min.)	Istd RT (min.)	Rel RT		Mean Rel RT	Rel RT Range (+/- .06)
2-Hexanone	13.16	14.67	0.897	ok	0.898	0.838-0.958
Isopropylbenzene	16.43	14.67	1.120	ok	1.120	1.060-1.180
Isopropyl Alcohol	5.92	8.62	0.687	ok	0.692	0.632-0.752
Methylene chloride	6.55	8.62	0.760	ok	0.760	0.700-0.820
Methyl ethyl ketone	8.07	8.62	0.936	ok	0.937	0.877-0.997
Methyl Isobutyl Ketone	11.99	10.44	1.148	ok	1.149	1.089-1.209
Methyl Tert Butyl Ether	7.73	8.62	0.897	ok	0.897	0.837-0.957
Nonane	15.97	14.67	1.089	ok	1.089	1.029-1.149
Pentane	6.14	8.62	0.712	ok	0.712	0.652-0.772
Propylene	3.99	8.62	0.463	ok	0.463	0.403-0.523
Styrene	15.67	14.67	1.068	ok	1.068	1.008-1.128
1,1,1-Trichloroethane	9.68	8.62	1.123	ok	1.123	1.063-1.183
1,1,2,2-Tetrachloroethane	15.79	14.67	1.076	ok	1.077	1.017-1.137
1,1,2-Trichloroethane	12.62	10.44	1.209	ok	1.209	1.149-1.269
1,2,4-Trichlorobenzene	20.27	14.67	1.382	ok	1.382	1.322-1.442
1,2,4-Trimethylbenzene	17.72	14.67	1.208	ok	1.208	1.148-1.268
1,3,5-Trimethylbenzene	17.26	14.67	1.177	ok	1.176	1.116-1.236
2,2,4-Trimethylpentane	11.12	10.44	1.065	ok	1.065	1.005-1.125
Tertiary Butyl Alcohol	6.53	8.62	0.758	ok	0.760	0.700-0.820
Tetrachloroethylene	14.01	14.67	0.955	ok	0.955	0.895-1.015
Tetrahydrofuran	9.18	8.62	1.065	ok	1.066	1.006-1.126
Toluene	12.89	10.44	1.235	ok	1.234	1.174-1.294
Trichloroethylene	11.12	10.44	1.065	ok	1.065	1.005-1.125
Trichlorofluoromethane	5.79	8.62	0.672	ok	0.672	0.612-0.732
Vinyl chloride	4.45	8.62	0.516	ok	0.516	0.456-0.576
Vinyl Acetate	7.80	8.62	0.905	ok	0.905	0.845-0.965
m,p-Xylene	15.29	14.67	1.042	ok	1.042	0.982-1.102
o-Xylene	15.79	14.67	1.076	ok	1.076	1.016-1.136

Internal Standard	RT (min.)		Mean RT(min.)	RT Range (+/- 0.33)	Area		Mean Area	Area Range (+/- 40 %)
Bromochloromethane	8.62	ok	8.62	8.29-8.95	412358	ok	371623	222974-520272
1,4-Difluorobenzene	10.44	ok	10.44	10.11-10.77	1821050	ok	1631886	979132-2284640
Chlorobenzene-D5	14.67	ok	14.67	14.34-15.00	1621470	ok	1431182	858709-2003655

Initial Calibration Retention Time/Internal Standard Area Summary

Job Number: M64467
Account: GEI GEI Consultants, Inc.
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

Sample Number	Lab File ID	Injected	By	Level	Inst ID	Method	
MSJ533-IC533	J10277.D	05/09/07 18:39	DFT	0.2	GCMSJ	TO-15	
MSJ533-IC533	J10278.D	05/09/07 19:29	DFT	0.5	GCMSJ	TO-15	
MSJ533-IC533	J10279.D	05/09/07 20:18	DFT	40	GCMSJ	TO-15	
MSJ533-IC533	J10280.D	05/09/07 21:05	DFT	20	GCMSJ	TO-15	
MSJ533-ICC533	J10281.D	05/09/07 21:52	DFT	10	GCMSJ	TO-15	
MSJ533-IC533	J10282.D	05/09/07 22:40	DFT	5	GCMSJ	TO-15	Reporting this level
MSJ533-IC533	J10283.D	05/09/07 23:27	DFT	2	GCMSJ	TO-15	

Target Compound	RT (min.)	Istd RT (min.)	Rel RT		Mean Rel RT	Rel RT Range (+/- .06)
Acetone	5.66	8.62	0.657	ok	0.657	0.597-0.717
1,3-Butadiene	4.58	8.62	0.531	ok	0.532	0.472-0.592
Benzene	10.12	10.44	0.969	ok	0.969	0.909-1.029
Bromodichloromethane	11.08	10.44	1.061	ok	1.062	1.002-1.122
Bromoform	15.39	14.67	1.049	ok	1.049	0.989-1.109
Bromomethane	4.86	8.62	0.564	ok	0.563	0.503-0.623
Bromoethene	5.39	8.62	0.625	ok	0.624	0.564-0.684
Benzyl Chloride	17.89	14.67	1.219	ok	1.219	1.159-1.279
Carbon disulfide	6.85	8.62	0.795	ok	0.794	0.734-0.854
Chlorobenzene	14.72	14.67	1.003	ok	1.003	0.943-1.063
Chloroethane	5.03	8.62	0.584	ok	0.583	0.523-0.643
Chloroform	8.74	8.62	1.014	ok	1.014	0.954-1.074
Chloromethane	4.24	8.62	0.492	ok	0.490	0.430-0.550
3-Chloropropene	6.66	8.62	0.773	ok	0.772	0.712-0.832
2-Chlorotoluene	16.98	14.67	1.157	ok	1.157	1.097-1.217
Carbon tetrachloride	10.26	8.62	1.190	ok	1.190	1.130-1.250
Cyclohexane	10.37	10.44	0.993	ok	0.993	0.933-1.053
1,1-Dichloroethane	7.65	8.62	0.887	ok	0.888	0.828-0.948
1,1-Dichloroethylene	6.43	8.62	0.746	ok	0.745	0.685-0.805
1,2-Dibromoethane	13.57	14.67	0.925	ok	0.925	0.865-0.985
1,2-Dichloroethane	9.45	8.62	1.096	ok	1.096	1.036-1.156
1,2-Dichloropropane	10.90	10.44	1.044	ok	1.044	0.984-1.104
1,4-Dioxane	11.18	10.44	1.071	ok	1.071	1.011-1.131
Dichlorodifluoromethane	4.07	8.62	0.472	ok	0.472	0.412-0.532
Dibromochloromethane	13.32	14.67	0.908	ok	0.908	0.848-0.968
trans-1,2-Dichloroethylene	7.46	8.62	0.865	ok	0.865	0.805-0.925
cis-1,2-Dichloroethylene	8.46	8.62	0.981	ok	0.981	0.921-1.041
cis-1,3-Dichloropropene	11.94	10.44	1.144	ok	1.144	1.084-1.204
m-Dichlorobenzene	17.90	14.67	1.220	ok	1.220	1.160-1.280
o-Dichlorobenzene	18.37	14.67	1.252	ok	1.252	1.192-1.312
p-Dichlorobenzene	17.98	14.67	1.226	ok	1.226	1.166-1.286
trans-1,3-Dichloropropene	12.44	10.44	1.192	ok	1.192	1.132-1.252
Ethanol	5.20	8.62	0.603	ok	0.604	0.544-0.664
Ethylbenzene	15.09	14.67	1.029	ok	1.029	0.969-1.089
Ethyl Acetate	8.69	8.62	1.008	ok	1.009	0.949-1.069
4-Ethyltoluene	17.17	14.67	1.170	ok	1.170	1.110-1.230
Freon 113	6.79	8.62	0.788	ok	0.788	0.728-0.848
Freon 114	4.33	8.62	0.502	ok	0.501	0.441-0.561
Heptane	11.36	10.44	1.088	ok	1.088	1.028-1.148
Hexachlorobutadiene	20.76	14.67	1.415	ok	1.415	1.355-1.475
Hexane	8.63	8.62	1.001	ok	1.001	0.941-1.061

Initial Calibration Retention Time/Internal Standard Area Summary

Job Number: M64467
Account: GEI GEI Consultants, Inc.
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

Sample Number	Lab File ID	Injected	By	Level	Inst ID	Method	
MSJ533-IC533	J10277.D	05/09/07 18:39	DFT	0.2	GCMSJ	TO-15	
MSJ533-IC533	J10278.D	05/09/07 19:29	DFT	0.5	GCMSJ	TO-15	
MSJ533-IC533	J10279.D	05/09/07 20:18	DFT	40	GCMSJ	TO-15	
MSJ533-IC533	J10280.D	05/09/07 21:05	DFT	20	GCMSJ	TO-15	
MSJ533-ICC533	J10281.D	05/09/07 21:52	DFT	10	GCMSJ	TO-15	
MSJ533-IC533	J10282.D	05/09/07 22:40	DFT	5	GCMSJ	TO-15	Reporting this level
MSJ533-IC533	J10283.D	05/09/07 23:27	DFT	2	GCMSJ	TO-15	

Target Compound	RT (min.)	Istd RT (min.)	Rel RT		Mean Rel RT	Rel RT Range (+/- .06)
2-Hexanone	13.16	14.67	0.897	ok	0.898	0.838-0.958
Isopropylbenzene	16.42	14.67	1.119	ok	1.120	1.060-1.180
Isopropyl Alcohol	5.93	8.62	0.688	ok	0.692	0.632-0.752
Methylene chloride	6.55	8.62	0.760	ok	0.760	0.700-0.820
Methyl ethyl ketone	8.07	8.62	0.936	ok	0.937	0.877-0.997
Methyl Isobutyl Ketone	11.99	10.44	1.148	ok	1.149	1.089-1.209
Methyl Tert Butyl Ether	7.73	8.62	0.897	ok	0.897	0.837-0.957
Nonane	15.97	14.67	1.089	ok	1.089	1.029-1.149
Pentane	6.14	8.62	0.712	ok	0.712	0.652-0.772
Propylene	4.00	8.62	0.464	ok	0.463	0.403-0.523
Styrene	15.66	14.67	1.067	ok	1.068	1.008-1.128
1,1,1-Trichloroethane	9.68	8.62	1.123	ok	1.123	1.063-1.183
1,1,2,2-Tetrachloroethane	15.79	14.67	1.076	ok	1.077	1.017-1.137
1,1,2-Trichloroethane	12.62	10.44	1.209	ok	1.209	1.149-1.269
1,2,4-Trichlorobenzene	20.27	14.67	1.382	ok	1.382	1.322-1.442
1,2,4-Trimethylbenzene	17.72	14.67	1.208	ok	1.208	1.148-1.268
1,3,5-Trimethylbenzene	17.25	14.67	1.176	ok	1.176	1.116-1.236
2,2,4-Trimethylpentane	11.12	10.44	1.065	ok	1.065	1.005-1.125
Tertiary Butyl Alcohol	6.53	8.62	0.758	ok	0.760	0.700-0.820
Tetrachloroethylene	14.01	14.67	0.955	ok	0.955	0.895-1.015
Tetrahydrofuran	9.18	8.62	1.065	ok	1.066	1.006-1.126
Toluene	12.88	10.44	1.234	ok	1.234	1.174-1.294
Trichloroethylene	11.12	10.44	1.065	ok	1.065	1.005-1.125
Trichlorofluoromethane	5.80	8.62	0.673	ok	0.672	0.612-0.732
Vinyl chloride	4.45	8.62	0.516	ok	0.516	0.456-0.576
Vinyl Acetate	7.80	8.62	0.905	ok	0.905	0.845-0.965
m,p-Xylene	15.27	14.67	1.041	ok	1.042	0.982-1.102
o-Xylene	15.78	14.67	1.076	ok	1.076	1.016-1.136

Internal Standard	RT (min.)		Mean RT(min.)	RT Range (+/- 0.33)	Area		Mean Area	Area Range (+/- 40 %)
Bromochloromethane	8.62	ok	8.62	8.29-8.95	366773	ok	371623	222974-520272
1,4-Difluorobenzene	10.44	ok	10.44	10.11-10.77	1563527	ok	1631886	979132-2284640
Chlorobenzene-D5	14.67	ok	14.67	14.34-15.00	1351983	ok	1431182	858709-2003655

Initial Calibration Retention Time/Internal Standard Area Summary

Job Number: M64467
Account: GEI GEI Consultants, Inc.
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

Sample Number	Lab File ID	Injected	By	Level	Inst ID	Method	
MSJ533-IC533	J10277.D	05/09/07 18:39	DFT	0.2	GCMSJ	TO-15	
MSJ533-IC533	J10278.D	05/09/07 19:29	DFT	0.5	GCMSJ	TO-15	
MSJ533-IC533	J10279.D	05/09/07 20:18	DFT	40	GCMSJ	TO-15	
MSJ533-IC533	J10280.D	05/09/07 21:05	DFT	20	GCMSJ	TO-15	
MSJ533-ICC533	J10281.D	05/09/07 21:52	DFT	10	GCMSJ	TO-15	
MSJ533-IC533	J10282.D	05/09/07 22:40	DFT	5	GCMSJ	TO-15	
MSJ533-IC533	J10283.D	05/09/07 23:27	DFT	2	GCMSJ	TO-15	Reporting this level

Target Compound	RT (min.)	Istd RT (min.)	Rel RT		Mean Rel RT	Rel RT Range (+/- .06)
Acetone	5.67	8.62	0.658	ok	0.657	0.597-0.717
1,3-Butadiene	4.59	8.62	0.532	ok	0.532	0.472-0.592
Benzene	10.12	10.44	0.969	ok	0.969	0.909-1.029
Bromodichloromethane	11.08	10.44	1.061	ok	1.062	1.002-1.122
Bromoform	15.39	14.67	1.049	ok	1.049	0.989-1.109
Bromomethane	4.86	8.62	0.564	ok	0.563	0.503-0.623
Bromoethene	5.38	8.62	0.624	ok	0.624	0.564-0.684
Benzyl Chloride	17.89	14.67	1.219	ok	1.219	1.159-1.279
Carbon disulfide	6.85	8.62	0.795	ok	0.794	0.734-0.854
Chlorobenzene	14.72	14.67	1.003	ok	1.003	0.943-1.063
Chloroethane	5.03	8.62	0.584	ok	0.583	0.523-0.643
Chloroform	8.74	8.62	1.014	ok	1.014	0.954-1.074
Chloromethane	4.23	8.62	0.491	ok	0.490	0.430-0.550
3-Chloropropene	6.67	8.62	0.774	ok	0.772	0.712-0.832
2-Chlorotoluene	16.98	14.67	1.157	ok	1.157	1.097-1.217
Carbon tetrachloride	10.26	8.62	1.190	ok	1.190	1.130-1.250
Cyclohexane	10.37	10.44	0.993	ok	0.993	0.933-1.053
1,1-Dichloroethane	7.65	8.62	0.887	ok	0.888	0.828-0.948
1,1-Dichloroethylene	6.42	8.62	0.745	ok	0.745	0.685-0.805
1,2-Dibromoethane	13.57	14.67	0.925	ok	0.925	0.865-0.985
1,2-Dichloroethane	9.45	8.62	1.096	ok	1.096	1.036-1.156
1,2-Dichloropropane	10.90	10.44	1.044	ok	1.044	0.984-1.104
1,4-Dioxane	11.20	10.44	1.073	ok	1.071	1.011-1.131
Dichlorodifluoromethane	4.07	8.62	0.472	ok	0.472	0.412-0.532
Dibromochloromethane	13.32	14.67	0.908	ok	0.908	0.848-0.968
trans-1,2-Dichloroethylene	7.46	8.62	0.865	ok	0.865	0.805-0.925
cis-1,2-Dichloroethylene	8.46	8.62	0.981	ok	0.981	0.921-1.041
cis-1,3-Dichloropropene	11.93	10.44	1.143	ok	1.144	1.084-1.204
m-Dichlorobenzene	17.90	14.67	1.220	ok	1.220	1.160-1.280
o-Dichlorobenzene	18.37	14.67	1.252	ok	1.252	1.192-1.312
p-Dichlorobenzene	17.98	14.67	1.226	ok	1.226	1.166-1.286
trans-1,3-Dichloropropene	12.44	10.44	1.192	ok	1.192	1.132-1.252
Ethanol	5.22	8.62	0.606	ok	0.604	0.544-0.664
Ethylbenzene	15.09	14.67	1.029	ok	1.029	0.969-1.089
Ethyl Acetate	8.70	8.62	1.009	ok	1.009	0.949-1.069
4-Ethyltoluene	17.17	14.67	1.170	ok	1.170	1.110-1.230
Freon 113	6.80	8.62	0.789	ok	0.788	0.728-0.848
Freon 114	4.32	8.62	0.501	ok	0.501	0.441-0.561
Heptane	11.36	10.44	1.088	ok	1.088	1.028-1.148
Hexachlorobutadiene	20.76	14.67	1.415	ok	1.415	1.355-1.475
Hexane	8.63	8.62	1.001	ok	1.001	0.941-1.061

Initial Calibration Retention Time/Internal Standard Area Summary

Job Number: M64467
Account: GEI GEI Consultants, Inc.
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

Sample Number	Lab File ID	Injected	By	Level	Inst ID	Method	
MSJ533-IC533	J10277.D	05/09/07 18:39	DFT	0.2	GCMSJ	TO-15	
MSJ533-IC533	J10278.D	05/09/07 19:29	DFT	0.5	GCMSJ	TO-15	
MSJ533-IC533	J10279.D	05/09/07 20:18	DFT	40	GCMSJ	TO-15	
MSJ533-IC533	J10280.D	05/09/07 21:05	DFT	20	GCMSJ	TO-15	
MSJ533-ICC533	J10281.D	05/09/07 21:52	DFT	10	GCMSJ	TO-15	
MSJ533-IC533	J10282.D	05/09/07 22:40	DFT	5	GCMSJ	TO-15	
MSJ533-IC533	J10283.D	05/09/07 23:27	DFT	2	GCMSJ	TO-15	Reporting this level

Target Compound	RT (min.)	Istd RT (min.)	Rel RT		Mean Rel RT	Rel RT Range (+/- .06)
2-Hexanone	13.18	14.67	0.898	ok	0.898	0.838-0.958
Isopropylbenzene	16.42	14.67	1.119	ok	1.120	1.060-1.180
Isopropyl Alcohol	5.95	8.62	0.690	ok	0.692	0.632-0.752
Methylene chloride	6.55	8.62	0.760	ok	0.760	0.700-0.820
Methyl ethyl ketone	8.08	8.62	0.937	ok	0.937	0.877-0.997
Methyl Isobutyl Ketone	12.00	10.44	1.149	ok	1.149	1.089-1.209
Methyl Tert Butyl Ether	7.74	8.62	0.898	ok	0.897	0.837-0.957
Nonane	15.97	14.67	1.089	ok	1.089	1.029-1.149
Pentane	6.14	8.62	0.712	ok	0.712	0.652-0.772
Propylene	3.99	8.62	0.463	ok	0.463	0.403-0.523
Styrene	15.67	14.67	1.068	ok	1.068	1.008-1.128
1,1,1-Trichloroethane	9.68	8.62	1.123	ok	1.123	1.063-1.183
1,1,2,2-Tetrachloroethane	15.79	14.67	1.076	ok	1.077	1.017-1.137
1,1,2-Trichloroethane	12.62	10.44	1.209	ok	1.209	1.149-1.269
1,2,4-Trichlorobenzene	20.27	14.67	1.382	ok	1.382	1.322-1.442
1,2,4-Trimethylbenzene	17.72	14.67	1.208	ok	1.208	1.148-1.268
1,3,5-Trimethylbenzene	17.25	14.67	1.176	ok	1.176	1.116-1.236
2,2,4-Trimethylpentane	11.12	10.44	1.065	ok	1.065	1.005-1.125
Tertiary Butyl Alcohol	6.56	8.62	0.761	ok	0.760	0.700-0.820
Tetrachloroethylene	14.01	14.67	0.955	ok	0.955	0.895-1.015
Tetrahydrofuran	9.20	8.62	1.067	ok	1.066	1.006-1.126
Toluene	12.89	10.44	1.235	ok	1.234	1.174-1.294
Trichloroethylene	11.11	10.44	1.064	ok	1.065	1.005-1.125
Trichlorofluoromethane	5.80	8.62	0.673	ok	0.672	0.612-0.732
Vinyl chloride	4.44	8.62	0.515	ok	0.516	0.456-0.576
Vinyl Acetate	7.81	8.62	0.906	ok	0.905	0.845-0.965
m,p-Xylene	15.28	14.67	1.042	ok	1.042	0.982-1.102
o-Xylene	15.78	14.67	1.076	ok	1.076	1.016-1.136

Internal Standard	RT (min.)		Mean RT(min.)	RT Range (+/- 0.33)	Area		Mean Area	Area Range (+/- 40 %)
Bromochloromethane	8.62	ok	8.62	8.29-8.95	364997	ok	371623	222974-520272
1,4-Difluorobenzene	10.44	ok	10.44	10.11-10.77	1696449	ok	1631886	979132-2284640
Chlorobenzene-D5	14.67	ok	14.67	14.34-15.00	1441208	ok	1431182	858709-2003655

Volatile Surrogate Recovery Summary

Job Number: M64467
Account: GEI GEI Consultants, Inc.
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

Method: TO-15 **Matrix:** AIR

Samples and QC shown here apply to the above method

Lab Sample ID	Lab File ID	S1
M64467-1	J10289.D	95.0
M64467-1DUP	J10293.D	94.0
MSJ533-BS	J10286B.D	111.0
MSJ533-MB	J10288.D	95.0

Surrogate Compounds	Recovery Limits
S1 = 4-Bromofluorobenzene	57-139%

Initial Calibration Summary

Job Number: M64467 **Sample:** MSJ533-ICC533
Account: GEI GEI Consultants, Inc. **Lab FileID:** J10281.D
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

Response Factor Report MSJ

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 11:09:51 2007
Response via : Initial Calibration

Calibration Files
40 =J10279.D 20 =J10280.D 10 =J10281.D 5 =J10282.D
2 =J10283.D .5 =J10278.D .2 =J10277.D 4 =J10094.D

Compound	40	20	10	5	2	.5	.2	4	Avg	%RSD
1) I BROMOCHLOROMETHANE	----------ISTD----------									
2) DICHLORODIFL	3.120	3.273	3.238	3.596	3.635	3.947	3.789		3.514	8.80
3) PROPYLENE	0.441	0.464	0.450	0.457	0.379	0.218			0.402	23.69
4) FREON 114	2.706	2.781	2.678	2.746	3.317	3.295	3.221		2.963	10.03
5) CHLOROMETHAN	0.602	0.633	0.604	0.652	0.731	0.673	0.672		0.652	6.91
6) VINYL CHLORI	0.763	0.812	0.777	0.773	0.816	0.980	0.812		0.819	9.06
7) 1,3-BUTADIEN	0.552	0.559	0.516	0.517	0.534	0.488	0.294		0.494	18.50
8) BROMOMETHANE	0.932	0.971	0.917	0.948	1.092	1.087	0.996		0.992	7.20
9) CHLOROETHANE	0.367	0.372	0.350	0.338	0.318	0.384	0.206		0.334	18.18
10) TRICHLOROFLU	3.269	3.266	3.181	3.251	3.916	4.248	3.920		3.579	12.17
11) ISOPROPYL AL	0.852	0.902	0.781	0.676	0.501	0.568	1.000		0.754	24.06
12) ACETONE	0.914	0.907	0.813	0.792	0.851	1.346	2.038		1.094	41.70
----- Linear regression ----- Coefficient = 0.9988										
Response Ratio = -0.02062 + 0.91416 *A										
13) PENTANE	0.602	0.607	0.593	0.591	0.677	0.660	0.483		0.602	10.38
14) 1,1-DICHLORO	0.932	0.952	0.903	0.862	0.934	0.828	0.857		0.895	5.26
15) CARBON DISUL	2.468	2.554	2.452	2.551	2.910	3.098	2.935		2.710	9.72
16) ETHANOL	0.144	0.147	0.124	0.110	0.115	0.070			0.118	23.84
17) BROMOETHENE	0.969	0.996	0.918	0.892	0.954	0.837	0.751		0.902	9.44
18) METHYLENE CH	0.823	0.825	0.773	0.789	0.859	6.693			1.794	133.82
----- Linear regression ----- Coefficient = 0.9903										
Response Ratio = 0.07890 + 0.79243 *A										
19) 3-CHLOROPROP	0.659	0.593	0.533	0.498	0.531	0.664	0.574		0.579	11.10
20) FREON 113	2.043	2.040	1.931	1.994	2.235	2.559	2.531		2.191	11.84
21) TRANS-1,2-DI	0.908	0.903	0.825	0.807	0.859	0.957	0.832		0.870	6.22
22) TERTIARY BUT	1.256	1.300	1.094	1.012	0.875	0.656			1.032	23.45
23) METHYL TERTI	1.838	1.828	1.586	1.502	1.538	1.685	1.111		1.584	15.61
24) TETRAHYDROFU	0.549	0.538	0.453	0.402	0.375	0.293			0.435	22.71
25) HEXANE	0.917	0.875	0.828	0.809	0.877	1.203	1.259		0.967	19.10
26) VINYL ACETAT	1.179	1.133	0.954	0.870	0.810	0.857			0.967	15.94
27) 1,1-DICHLORO	1.453	1.376	1.294	1.257	1.424	1.694	1.367		1.409	10.13
28) METHYL ETHYL	1.012	1.019	0.886	0.829	0.883	0.918	0.734		0.897	11.12
29) cis-1,2-DICH	0.907	0.846	0.777	0.735	0.847	0.866	0.685		0.809	9.80
30) ETHYL ACETAT	1.132	1.126	0.988	0.901	0.970	1.112	0.808		1.005	12.42
31) CHLOROFORM	2.115	1.982	1.867	1.838	2.157	2.447	2.117		2.075	9.97
32) 1,1,1-TRICHL	2.596	2.404	2.194	2.122	2.383	2.753	2.400		2.408	9.02
33) CARBON TETRA	3.205	3.035	2.890	2.934	3.171	3.638	3.284		3.165	8.01
34) 1,2-DICHLORO	1.269	1.157	1.028	0.997	1.178	1.294	1.081		1.143	10.00
35) I 1,4-DIFLUOROBENZENE	----------ISTD----------									
36) BENZENE	0.525	0.465	0.415	0.389	0.388	0.445	0.359		0.427	13.18
37) CYCLOHEXANE	0.241	0.229	0.214	0.211	0.193	0.169	0.132		0.198	18.90
38) TRICHLOROETH	0.276	0.251	0.230	0.235	0.258	0.259	0.186		0.242	12.03
39) 1,2-DICHLORO	0.169	0.149	0.133	0.131	0.131	0.170	0.122		0.144	13.52

Initial Calibration Summary

Job Number: M64467 **Sample:** MSJ533-ICC533
Account: GEI GEI Consultants, Inc. **Lab FileID:** J10281.D
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

```
40) BROMODICHLOR 0.492 0.444 0.405 0.408 0.448 0.563 0.462       0.460  11.77
41) 2,2,4-TRIMET 0.788 0.745 0.684 0.633 0.568 0.527 0.374       0.617  22.93
42) 1,4-DIOXANE  0.097 0.092 0.076 0.061 0.040 0.013             0.063  51.22
            ----- Linear regression -----  Coefficient =  0.9978
              Response Ratio = -0.01465 + 0.10021 *A

43) HEPTANE      0.240 0.218 0.196 0.184 0.178 0.155 0.100       0.182  25.10
44) METHYL ISOBU 0.304 0.275 0.233 0.210 0.182 0.148             0.225  25.80
45) cis-1,3-DICH 0.282 0.253 0.212 0.189 0.169 0.164 0.117       0.198  28.22
46) TOLUENE      0.398 0.350 0.298 0.274 0.246 0.223 0.170       0.280  27.54
47) trans-1,3-DI 0.280 0.250 0.204 0.177 0.156 0.117             0.197  30.59
            ----- Linear regression -----  Coefficient =  0.9953
              Response Ratio = -0.04344 + 0.28517 *A

48) 1,1,2-TRICHL 0.188 0.169 0.150 0.147 0.156 0.183 0.151       0.163  10.27

49) I   CHLOROBENZENE-D5      ----------------ISTD--------------------
50) 2-HEXANONE   0.289 0.258 0.216 0.198 0.146 0.117             0.204  32.03
            ----- Linear regression -----  Coefficient =  0.9959
              Response Ratio = -0.04283 + 0.29406 *A

51) TETRACHLOROE 0.385 0.340 0.301 0.302 0.320 0.420 0.296       0.338  14.20
52) DIBROMOCHLOR 0.573 0.521 0.476 0.490 0.541 0.706 0.498       0.544  14.49
53) 1,2-DIBROMOE 0.406 0.369 0.322 0.314 0.328 0.353 0.255       0.335  14.15
54) CHLOROBENZEN 0.597 0.553 0.492 0.494 0.550 0.591 0.443       0.531  10.70
55) ETHYLBENZENE 0.945 0.860 0.750 0.711 0.615 0.528 0.353       0.680  29.58
56) m,p-XYLENE   0.391 0.352 0.315 0.301 0.274 0.205 0.142       0.283  30.28
            ----- Linear regression -----  Coefficient =  0.9966
              Response Ratio = -0.06858 + 0.39166 *A

57) o-XYLENE     0.400 0.354 0.316 0.301 0.264 0.212 0.137       0.283  31.16
            ----- Linear regression -----  Coefficient =  0.9958
              Response Ratio = -0.03784 + 0.39955 *A

58) STYRENE      0.576 0.505 0.413 0.338 0.255 0.174             0.377  40.25
            ----- Linear regression -----  Coefficient =  0.9947
              Response Ratio = -0.10022 + 0.58768 *A

59) NONANE       0.364 0.328 0.297 0.276 0.251 0.212             0.288  18.86
60) BROMOFORM    0.637 0.571 0.504 0.489 0.522 0.646 0.501       0.553  11.93
61) 4-BROMOFLUOR 0.708 0.670 0.649 0.564 0.542 0.487 0.487       0.587  15.22
62) 1,1,2,2-TETR 0.536 0.487 0.433 0.429 0.437 0.596 0.365       0.469  16.53
63) ISOPROPYLBEN 1.158 1.034 0.905 0.830 0.739 0.644             0.885  21.41
64) 2-CHLOROTOLU 0.799 0.703 0.597 0.525 0.456 0.418             0.583  25.21
65) 4-ETHYLTOLUE 1.085 0.941 0.781 0.679 0.565 0.410             0.744  33.16
            ----- Linear regression -----  Coefficient =  0.9942
              Response Ratio = -0.17841 + 1.10272 *A

66) 1,3,5-TRIMET 0.968 0.848 0.721 0.632 0.548 0.460             0.696  27.23
67) 1,2,4-TRIMET 0.924 0.779 0.636 0.559 0.448 0.335             0.613  35.12
            ----- Linear regression -----  Coefficient =  0.9923
              Response Ratio = -0.16618 + 0.93823 *A

68) m-DICHLOROBE 0.684 0.567 0.466 0.455 0.452 0.359             0.497  22.70
69) BENZYL CHLOR 0.586 0.470 0.377 0.332 0.253 0.218             0.373  36.91
         ----- Quadratic regression -----                Coefficient =  0.9998
              Response Ratio = -0.01976 + 0.35387 *A + 0.05932 *A^2

70) p-DICHLOROBE 0.640 0.534 0.441 0.429 0.383 0.301             0.455  26.06
71) o-DICHLOROBE 0.613 0.502 0.413 0.384 0.368 0.379             0.443  21.74
```

Initial Calibration Summary

Job Number: M64467 **Sample:** MSJ533-ICC533
Account: GEI GEI Consultants, Inc. **Lab FileID:** J10281.D
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

```
72) HEXACHLOROBU 0.348 0.292 0.256 0.267 0.259 1.126            0.425  81.32
         ----- Quadratic regression -----                 Coefficient =  0.9993
              Response Ratio = 0.02574 + 0.20482 *A + 0.03437 *A^2

73) 1,2,4-TRICHL 0.206 0.171 0.135 0.119 0.090 0.144            0.144  28.01
74) NAPHTHALENE                                                 0.000  -1.00
---------------------------------------------------------------------------
(#) = Out of Range  ###  Number of calibration levels exceeded format  ###

          J050907T.M        Thu May 10 11:10:50 2007
```

5.8

5

Continuing Calibration Summary

Job Number: M64467 **Sample:** MSJ533-CC533
Account: GEI GEI Consultants, Inc. **Lab FileID:** J10286A.D
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\J10286A.D Vial: 4
Acq On : 10 May 2007 10:33 am Operator: danat
Sample : CC533-10 (M054) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 11:09:51 2007
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 40% Max. R.T. Dev 0.33min
Max. RRF Dev : 30% Max. Rel. Area : 180%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)	R.T.
1	I	BROMOCHLOROMETHANE	1.000	1.000	0.0	100	0.00	8.62
2	m	DICHLORODIFLUOROMETHANE	3.514	3.160	10.1	98	0.00	4.07
3	m	PROPYLENE	0.402	0.432	-7.5	96	0.00	3.99
4	m	FREON 114	2.963	2.556	13.7	96	0.00	4.32
5	m	CHLOROMETHANE	0.652	0.598	8.3	99	0.00	4.23
6	m	VINYL CHLORIDE	0.819	0.730	10.9	94	0.00	4.45
7	m	1,3-BUTADIENE	0.494	0.495	-0.2	96	0.00	4.58
8	m	BROMOMETHANE	0.992	0.872	12.1	95	0.00	4.85
9	m	CHLOROETHANE	0.334	0.341	-2.1	97	0.00	5.03
10	m	TRICHLOROFLUOROMETHANE	3.579	3.057	14.6	96	0.00	5.79
11	m	ISOPROPYL ALCOHOL	0.754	0.729	3.3	94	0.00	5.92
		----------	True	Calc.	% Drift	----------		
12	m	ACETONE	10.000	7.722	22.8	84	0.00	5.65
		----------	AvgRF	CCRF	% Dev	----------		
13	m	PENTANE	0.602	0.577	4.2	97	0.00	6.13
14	m	1,1-DICHLOROETHYLENE	0.895	0.778	13.1	86	0.00	6.43
15	m	CARBON DISULFIDE	2.710	2.238	17.4	91	0.00	6.85
16	m	ETHANOL	0.118	0.120	-1.7	97	0.00	5.20
17	m	BROMOETHENE	0.902	0.839	7.0	92	0.00	5.38
		----------	True	Calc.	% Drift	----------		
18	m	METHYLENE CHLORIDE	10.000	7.732	22.7	90	0.00	6.54
		----------	AvgRF	CCRF	% Dev	----------		
19	m	3-CHLOROPROPENE	0.579	0.532	8.1	100	0.00	6.66
20	m	FREON 113	2.191	1.736	20.8	90	0.00	6.79
21	m	TRANS-1,2-DICHLOROETHYLEN	0.870	0.768	11.7	93	0.00	7.46
22	m	TERTIARY BUTYL ALCOHOL	1.032	0.991	4.0	91	0.00	6.52
23	m	METHYL TERTIARY BUTYL ETH	1.584	1.292	18.4	82	0.00	7.73
24	m	TETRAHYDROFURAN	0.435	0.372	14.5	82	0.00	9.17
25	m	HEXANE	0.967	0.727	24.8	88	0.00	8.63
26	m	VINYL ACETATE	0.967	0.780	19.3	82	0.00	7.80
27	m	1,1-DICHLOROETHANE	1.409	1.192	15.4	92	0.00	7.66
28	m	METHYL ETHYL KETONE	0.897	0.737	17.8	83	0.00	8.07
29	m	cis-1,2-DICHLOROETHYLENE	0.809	0.692	14.5	89	0.00	8.46
30	m	ETHYL ACETATE	1.005	0.785	21.9	80	0.00	8.68
31	m	CHLOROFORM	2.075	1.724	16.9	93	0.00	8.74
32	m	1,1,1-TRICHLOROETHANE	2.408	2.068	14.1	94	0.00	9.68
33	m	CARBON TETRACHLORIDE	3.165	2.729	13.8	95	0.00	10.26
34	m	1,2-DICHLOROETHANE	1.143	0.925	19.1	90	0.00	9.45

Continuing Calibration Summary

Job Number: M64467 **Sample:** MSJ533-CC533
Account: GEI GEI Consultants, Inc. **Lab FileID:** J10286A.D
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

35	I	1,4-DIFLUOROBENZENE	1.000	1.000	0.0	101	0.00	10.44
36	m	BENZENE	0.427	0.365	14.5	89	0.00	10.12
37	m	CYCLOHEXANE	0.198	0.193	2.5	91	0.00	10.37
38	m	TRICHLOROETHYLENE	0.242	0.212	12.4	93	0.00	11.12
39	m	1,2-DICHLOROPROPANE	0.144	0.117	18.7	88	0.00	10.90
40	m	BROMODICHLOROMETHANE	0.460	0.357	22.4	89	0.00	11.08
41	m	2,2,4-TRIMETHYLPENTANE	0.617	0.647	-4.9	95	0.00	11.12
		----------	True	Calc.	% Drift	----------		
42	m	1,4-DIOXANE	10.000	8.136	18.6	89	0.00	11.16
		----------	AvgRF	CCRF	% Dev	----------		
43	m	HEPTANE	0.182	0.176	3.3	91	0.00	11.36
44	m	METHYL ISOBUTYL KETONE	0.225	0.200	11.1	86	0.00	11.98
45	m	cis-1,3-DICHLOROPROPENE	0.198	0.190	4.0	91	0.00	11.94
46	m	TOLUENE	0.280	0.252	10.0	85	0.00	12.89
		----------	True	Calc.	% Drift	----------		
47	m	trans-1,3-DICHLOROPROPENE	10.000	7.072	29.3	78	0.00	12.44
		----------	AvgRF	CCRF	% Dev	----------		
48	m	1,1,2-TRICHLOROETHANE	0.163	0.131	19.6	88	0.00	12.62
49	I	CHLOROBENZENE-D5	1.000	1.000	0.0	101	0.00	14.67
		----------	True	Calc.	% Drift	----------		
50	m	2-HEXANONE	10.000	7.381	26.2	82	0.00	13.15
		----------	AvgRF	CCRF	% Dev	----------		
51	m	TETRACHLOROETHYLENE	0.338	0.272	19.5	91	0.00	14.01
52	m	DIBROMOCHLOROMETHANE	0.544	0.414	23.9	88	0.00	13.33
53	m	1,2-DIBROMOETHANE	0.335	0.284	15.2	89	0.00	13.57
54	m	CHLOROBENZENE	0.465	0.436	6.2	90	0.00	14.71
55	m	ETHYLBENZENE	0.680	0.654	3.8	88	0.00	15.09
		----------	True	Calc.	% Drift	----------		
56	m	m,p-XYLENE	20.000	15.871	20.6	89	0.00	15.28
57	m	o-XYLENE	10.000	7.994	20.1	90	0.00	15.78
58	m	STYRENE	10.000	8.041	19.6	91	0.00	15.67
		----------	AvgRF	CCRF	% Dev	----------		
59	m	NONANE	0.288	0.279	3.1	95	0.00	15.97
60	m	BROMOFORM	0.553	0.446	19.3	90	0.00	15.40
61	S	4-BROMOFLUOROBENZENE	0.513	0.649	-26.5	101	0.00	16.29
62	m	1,1,2,2-TETRACHLOROETHANE	0.469	0.409	12.8	96	0.00	15.79
63	m	ISOPROPYLBENZENE	0.885	0.823	7.0	92	0.00	16.43
64	m	2-CHLOROTOLUENE	0.583	0.559	4.1	95	0.00	16.98
		----------	True	Calc.	% Drift	----------		
65	m	4-ETHYLTOLUENE	10.000	7.775	22.2	88	0.00	17.17
		----------	AvgRF	CCRF	% Dev	----------		
66	m	1,3,5-TRIMETHYLBENZENE	0.696	0.670	3.7	94	0.00	17.25
		----------	True	Calc.	% Drift	----------		
67	m	1,2,4-TRIMETHYLBENZENE	10.000	8.257	17.4	97	0.00	17.72
		----------	AvgRF	CCRF	% Dev	----------		
68	m	m-DICHLOROBENZENE	0.497	0.486	2.2	106	0.00	17.91

Continuing Calibration Summary

Job Number: M64467 **Sample:** MSJ533-CC533
Account: GEI GEI Consultants, Inc. **Lab FileID:** J10286A.D
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

```
            ----------------------- True    Calc.   % Drift  ------------
69 m   BENZYL CHLORIDE            10.000   9.896      1.0  104   0.00   17.89

            ----------------------- AvgRF   CCRF     % Dev   -------------
70 m   p-DICHLOROBENZENE           0.455   0.457     -0.4  105   0.00   17.98
71 m   o-DICHLOROBENZENE           0.443   0.431      2.7  105   0.00   18.37

            ----------------------- True    Calc.   % Drift  ------------
72 m   HEXACHLOROBUTADIENE        10.000  11.870    -18.7  126   0.00   20.76

            ----------------------- AvgRF   CCRF     % Dev   -------------
73 m   1,2,4-TRICHLOROBENZENE      0.144   0.201    -39.6# 151   0.00   20.27
74 m   NAPHTHALENE                          ----------NA----------
-------------------------------------------------------------------------------

 (  1.4 %)  1 of  70 compounds'%D > 30
-------------------------------------------------------------------------------

   (#) = Out of Range               SPCC's out = 0  CCC's out = 0
    J10281.D   J050907T.M       Thu May 10 14:49:31 2007
```

5.8

5

Initial Calibration Verification

Job Number: M64467 **Sample:** MSJ533-ICV533
Account: GEI GEI Consultants, Inc. **Lab FileID:** J10286.D
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\J10286.D Vial: 4
Acq On : 10 May 2007 10:33 am Operator: danat
Sample : ICV533-10 (M054) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 11:09:51 2007
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 40% Max. R.T. Dev 0.33min
Max. RRF Dev : 30% Max. Rel. Area : 180%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)	R.T.
1	I	BROMOCHLOROMETHANE	1.000	1.000	0.0	100	0.00	8.62
2	m	DICHLORODIFLUOROMETHANE	3.514	3.160	10.1	98	0.00	4.07
3	m	PROPYLENE	0.402	0.432	-7.5	96	0.00	3.99
4	m	FREON 114	2.963	2.556	13.7	96	0.00	4.32
5	m	CHLOROMETHANE	0.652	0.598	8.3	99	0.00	4.23
6	m	VINYL CHLORIDE	0.819	0.730	10.9	94	0.00	4.45
7	m	1,3-BUTADIENE	0.494	0.495	-0.2	96	0.00	4.58
8	m	BROMOMETHANE	0.992	0.872	12.1	95	0.00	4.85
9	m	CHLOROETHANE	0.334	0.341	-2.1	97	0.00	5.03
10	m	TRICHLOROFLUOROMETHANE	3.579	3.057	14.6	96	0.00	5.79
11	m	ISOPROPYL ALCOHOL	0.754	0.729	3.3	94	0.00	5.92
		----------	True	Calc.	% Drift	----------		
12	m	ACETONE	10.000	7.722	22.8	84	0.00	5.65
		----------	AvgRF	CCRF	% Dev	----------		
13	m	PENTANE	0.602	0.577	4.2	97	0.00	6.13
14	m	1,1-DICHLOROETHYLENE	0.895	0.778	13.1	86	0.00	6.43
15	m	CARBON DISULFIDE	2.710	2.238	17.4	91	0.00	6.85
16	m	ETHANOL	0.118	0.120	-1.7	97	0.00	5.20
17	m	BROMOETHENE	0.902	0.839	7.0	92	0.00	5.38
		----------	True	Calc.	% Drift	----------		
18	m	METHYLENE CHLORIDE	10.000	7.732	22.7	90	0.00	6.54
		----------	AvgRF	CCRF	% Dev	----------		
19	m	3-CHLOROPROPENE	0.579	0.532	8.1	100	0.00	6.66
20	m	FREON 113	2.191	1.736	20.8	90	0.00	6.79
21	m	TRANS-1,2-DICHLOROETHYLEN	0.870	0.768	11.7	93	0.00	7.46
22	m	TERTIARY BUTYL ALCOHOL	1.032	0.991	4.0	91	0.00	6.52
23	m	METHYL TERTIARY BUTYL ETH	1.584	1.292	18.4	82	0.00	7.73
24	m	TETRAHYDROFURAN	0.435	0.372	14.5	82	0.00	9.17
25	m	HEXANE	0.967	0.727	24.8	88	0.00	8.63
26	m	VINYL ACETATE	0.967	0.780	19.3	82	0.00	7.80
27	m	1,1-DICHLOROETHANE	1.409	1.192	15.4	92	0.00	7.66
28	m	METHYL ETHYL KETONE	0.897	0.737	17.8	83	0.00	8.07
29	m	cis-1,2-DICHLOROETHYLENE	0.809	0.692	14.5	89	0.00	8.46
30	m	ETHYL ACETATE	1.005	0.785	21.9	80	0.00	8.68
31	m	CHLOROFORM	2.075	1.724	16.9	93	0.00	8.74
32	m	1,1,1-TRICHLOROETHANE	2.408	2.068	14.1	94	0.00	9.68
33	m	CARBON TETRACHLORIDE	3.165	2.729	13.8	95	0.00	10.26
34	m	1,2-DICHLOROETHANE	1.143	0.925	19.1	90	0.00	9.45

Initial Calibration Verification

Job Number: M64467 **Sample:** MSJ533-ICV533
Account: GEI GEI Consultants, Inc. **Lab FileID:** J10286.D
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

35	I	1,4-DIFLUOROBENZENE	1.000	1.000	0.0	101	0.00	10.44
36	m	BENZENE	0.427	0.365	14.5	89	0.00	10.12
37	m	CYCLOHEXANE	0.198	0.193	2.5	91	0.00	10.37
38	m	TRICHLOROETHYLENE	0.242	0.212	12.4	93	0.00	11.12
39	m	1,2-DICHLOROPROPANE	0.144	0.117	18.7	88	0.00	10.90
40	m	BROMODICHLOROMETHANE	0.460	0.357	22.4	89	0.00	11.08
41	m	2,2,4-TRIMETHYLPENTANE	0.617	0.647	-4.9	95	0.00	11.12
		----------	True	Calc.	% Drift	----------		
42	m	1,4-DIOXANE	10.000	8.136	18.6	89	0.00	11.16
		----------	AvgRF	CCRF	% Dev	----------		
43	m	HEPTANE	0.182	0.176	3.3	91	0.00	11.36
44	m	METHYL ISOBUTYL KETONE	0.225	0.200	11.1	86	0.00	11.98
45	m	cis-1,3-DICHLOROPROPENE	0.198	0.190	4.0	91	0.00	11.94
46	m	TOLUENE	0.280	0.252	10.0	85	0.00	12.89
		----------	True	Calc.	% Drift	----------		
47	m	trans-1,3-DICHLOROPROPENE	10.000	7.072	29.3	78	0.00	12.44
		----------	AvgRF	CCRF	% Dev	----------		
48	m	1,1,2-TRICHLOROETHANE	0.163	0.131	19.6	88	0.00	12.62
49	I	CHLOROBENZENE-D5	1.000	1.000	0.0	101	0.00	14.67
		----------	True	Calc.	% Drift	----------		
50	m	2-HEXANONE	10.000	7.381	26.2	82	0.00	13.15
		----------	AvgRF	CCRF	% Dev	----------		
51	m	TETRACHLOROETHYLENE	0.338	0.272	19.5	91	0.00	14.01
52	m	DIBROMOCHLOROMETHANE	0.544	0.414	23.9	88	0.00	13.33
53	m	1,2-DIBROMOETHANE	0.335	0.284	15.2	89	0.00	13.57
54	m	CHLOROBENZENE	0.465	0.436	6.2	90	0.00	14.71
55	m	ETHYLBENZENE	0.680	0.654	3.8	88	0.00	15.09
		----------	True	Calc.	% Drift	----------		
56	m	m,p-XYLENE	20.000	15.871	20.6	89	0.00	15.28
57	m	o-XYLENE	10.000	7.994	20.1	90	0.00	15.78
58	m	STYRENE	10.000	8.041	19.6	91	0.00	15.67
		----------	AvgRF	CCRF	% Dev	----------		
59	m	NONANE	0.288	0.279	3.1	95	0.00	15.97
60	m	BROMOFORM	0.553	0.446	19.3	90	0.00	15.40
61	S	4-BROMOFLUOROBENZENE	0.513	0.649	-26.5	101	0.00	16.29
62	m	1,1,2,2-TETRACHLOROETHANE	0.469	0.409	12.8	96	0.00	15.79
63	m	ISOPROPYLBENZENE	0.885	0.823	7.0	92	0.00	16.43
64	m	2-CHLOROTOLUENE	0.583	0.559	4.1	95	0.00	16.98
		----------	True	Calc.	% Drift	----------		
65	m	4-ETHYLTOLUENE	10.000	7.775	22.2	88	0.00	17.17
		----------	AvgRF	CCRF	% Dev	----------		
66	m	1,3,5-TRIMETHYLBENZENE	0.696	0.670	3.7	94	0.00	17.25
		----------	True	Calc.	% Drift	----------		
67	m	1,2,4-TRIMETHYLBENZENE	10.000	8.257	17.4	97	0.00	17.72
		----------	AvgRF	CCRF	% Dev	----------		
68	m	m-DICHLOROBENZENE	0.497	0.486	2.2	106	0.00	17.91

Initial Calibration Verification

Job Number: M64467 **Sample:** MSJ533-ICV533
Account: GEI GEI Consultants, Inc. **Lab FileID:** J10286.D
Project: Indoor & Outdoor Air Samples Tufts St., Somerville MA

```
            ----------------------- True    Calc.   % Drift  ------------
69 m   BENZYL CHLORIDE            10.000   9.896      1.0  104   0.00   17.89

            ----------------------- AvgRF   CCRF     % Dev   -------------
70 m   p-DICHLOROBENZENE           0.455   0.457     -0.4  105   0.00   17.98
71 m   o-DICHLOROBENZENE           0.443   0.431      2.7  105   0.00   18.37

            ----------------------- True    Calc.   % Drift  ------------
72 m   HEXACHLOROBUTADIENE        10.000  11.870    -18.7  126   0.00   20.76

            ----------------------- AvgRF   CCRF     % Dev   -------------
73 m   1,2,4-TRICHLOROBENZENE      0.144   0.201    -39.6# 151   0.00   20.27
74 m   NAPHTHALENE                          ----------NA----------
-------------------------------------------------------------------------------

 (  1.4 %)  1 of  70 compounds'%D > 30
-------------------------------------------------------------------------------

   (#) = Out of Range               SPCC's out = 0  CCC's out = 0
    J10281.D   J050907T.M       Thu May 10 11:11:41 2007
```

GC/MS Volatiles

Raw Data

Manual Integrations APPROVED (compounds with "m" flag)
Philip Bassignani 05/11/07 16:59

Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\
Data File : J10289.D
Acq On : 10 May 2007 1:03 pm
Operator : danat
Sample : M64467-1 (M126)
Misc : MS13540, MSJ533,,,,,1
ALS Vial : 6 Sample Multiplier: 1

Quant Time: May 11 09:56:25 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 11:09:51 2007
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) BROMOCHLOROMETHANE	8.62	128	362149	10.00	PPBV	0.00
35) 1,4-DIFLUOROBENZENE	10.43	114	1604778	10.00	PPBV	-0.01
49) CHLOROBENZENE-D5	14.67	117	1460059	10.00	PPBV	0.00

System Monitoring Compounds

61) 4-BROMOFLUOROBENZENE	16.28	95	406174	4.74	PPBV	0.00

Spiked Amount 5.000 Range 51 - 142 Recovery = 94.80%

Target Compounds						Qvalue
2) DICHLORODIFLUOROMETHANE	4.07	85	70402	0.55	PPBV	99
10) TRICHLOROFLUOROMETHANE	5.81	101	38498	0.30	PPBV	93
12) ACETONE	5.67	43	110987	3.58	PPBV	99
13) PENTANE	6.15	42	2767m	0.13	PPBV	
15) CARBON DISULFIDE	6.85	76	22809	0.23	PPBV	91
28) METHYL ETHYL KETONE	8.11	43	9379m	0.29	PPBV	
31) CHLOROFORM	8.74	83	7804	0.10	PPBV	79
36) BENZENE	10.11	78	16628	0.24	PPBV	96
46) TOLUENE	12.89	92	4431	0.10	PPBV	90
51) TETRACHLOROETHYLENE	14.01	164	437867	8.88	PPBV	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\
Data File : J10289.D
Acq On : 10 May 2007 1:03 pm
Operator : danat
Sample : M64467-1 (M126)
Misc : MS13540, MSJ533,,,,,1
ALS Vial : 6 Sample Multiplier: 1

Quant Time: May 11 09:56:25 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 11:09:51 2007
Response via : Initial Calibration





#2
DICHLORODIFLUOROMETHANE
Concen: 0.55 PPBV
RT: 4.07 min Scan# 841
Delta R.T. 0.01 min
Lab File: J10289.D
Acq: 10 May 2007 1:03 pm

Tgt Ion: 85 Resp: 70402

Ion	Ratio	Lower	Upper
85	100		
87	32.7	12.7	52.7
50	11.8	0.0	31.0





#10
TRICHLOROFLUOROMETHANE
Concen: 0.30 PPBV
RT: 5.81 min Scan# 1819
Delta R.T. 0.01 min
Lab File: J10289.D
Acq: 10 May 2007 1:03 pm

Tgt Ion:101 Resp: 38498

Ion	Ratio	Lower	Upper
101	100		
103	70.5	44.8	84.8
105	9.6	0.0	30.5





#12
ACETONE
Concen: 3.58 PPBV
RT: 5.67 min Scan# 1744
Delta R.T. 0.02 min
Lab File: J10289.D
Acq: 10 May 2007 1:03 pm

Tgt Ion: 43 Resp: 110987

Ion	Ratio	Lower	Upper
43	100		
58	26.3	6.8	46.8





#13
PENTANE
Concen: 0.13 PPBV m
RT: 6.15 min Scan# 2014
Delta R.T. 0.02 min
Lab File: J10289.D
Acq: 10 May 2007 1:03 pm

Tgt Ion: 42 Resp: 2767

Ion	Ratio	Lower	Upper
42	100		
41	0.0	78.8	118.8#
57	0.0	7.4	47.4#





#15
CARBON DISULFIDE
Concen: 0.23 PPBV
RT: 6.85 min Scan# 2410
Delta R.T. 0.01 min
Lab File: J10289.D
Acq: 10 May 2007 1:03 pm

Tgt Ion: 76 Resp: 22809

Ion	Ratio	Lower	Upper
76	100		
78	5.5	0.0	29.2
44	18.2	0.0	34.9





#28
METHYL ETHYL KETONE
Concen: 0.29 PPBV m
RT: 8.11 min Scan# 3120
Delta R.T. 0.04 min
Lab File: J10289.D
Acq: 10 May 2007 1:03 pm

Tgt Ion: 43 Resp: 9379

Ion	Ratio	Lower	Upper
43	100		
57	0.0	0.0	28.8
72	11.1	7.6	47.6





#31
CHLOROFORM
Concen: 0.10 PPBV
RT: 8.74 min Scan# 3478
Delta R.T. 0.00 min
Lab File: J10289.D
Acq: 10 May 2007 1:03 pm

Tgt Ion: 83 Resp: 7804

Ion	Ratio	Lower	Upper
83	100		
85	45.5	44.8	84.8
47	26.9	1.7	41.7





#36
BENZENE
Concen: 0.24 PPBV
RT: 10.11 min Scan# 4250
Delta R.T. -0.01 min
Lab File: J10289.D
Acq: 10 May 2007 1:03 pm

Tgt Ion: 78 Resp: 16628

Ion	Ratio	Lower	Upper
78	100		
77	21.4	4.3	44.3
52	13.5	0.0	34.2





#46
TOLUENE
Concen: 0.10 PPBV
RT: 12.89 min Scan# 5816
Delta R.T. -0.00 min
Lab File: J10289.D
Acq: 10 May 2007 1:03 pm

Tgt Ion: 92 Resp: 4431

Ion	Ratio	Lower	Upper
92	100		
91	177.7	146.0	186.0
65	6.6	0.0	38.6





#51
TETRACHLOROETHYLENE
Concen: 8.88 PPBV
RT: 14.01 min Scan# 6450
Delta R.T. -0.00 min
Lab File: J10289.D
Acq: 10 May 2007 1:03 pm

Tgt Ion:164 Resp: 437867

Ion	Ratio	Lower	Upper
164	100		
129	85.6	67.6	107.6
168	62.8	42.8	82.8
131	84.2	64.5	104.5



Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10289.D Vial: 6
Acq On : 10 May 2007 1:03 pm Operator: danat
Sample : M64467-1 (M126) Inst : MSJ
Misc : MS13540, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 14:44:06 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 11:09:51 2007
Response via : Multiple Level Calibration







(28) METHYL ETHYL KETONE (m)

8.11min 0.16PPBV

response 5119

Ion	Exp%	Act%
43.00	100	100
57.00	8.80	0.00
72.00	27.60	11.08
0.00	0.00	0.00

J10289.D J050907T.M Thu May 10 14:45:10 2007

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10288.D
Acq On : 10 May 2007 12:14 pm
Operator : danat
Sample : MB (M009)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 5 Sample Multiplier: 1

Quant Time: May 10 14:41:18 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 11:09:51 2007
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) BROMOCHLOROMETHANE	8.62	128	381318	10.00	PPBV	0.00
35) 1,4-DIFLUOROBENZENE	10.44	114	1690348	10.00	PPBV	0.00
49) CHLOROBENZENE-D5	14.67	117	1515608	10.00	PPBV	0.00

System Monitoring Compounds

61) 4-BROMOFLUOROBENZENE	16.29	95	423304	4.76	PPBV	0.00

Spiked Amount 5.000 Range 51 - 142 Recovery = 95.20%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

6.2.1

6

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10288.D
Acq On : 10 May 2007 12:14 pm
Operator : danat
Sample : MB (M009)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 5 Sample Multiplier: 1

Quant Time: May 10 14:41:18 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 11:09:51 2007
Response via : Initial Calibration



Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10286B.D
Acq On : 10 May 2007 10:33 am
Operator : danat
Sample : BS (M054)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 4 Sample Multiplier: 1

Quant Time: May 10 11:10:20 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 11:09:51 2007
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) BROMOCHLOROMETHANE	8.62	128	413189	10.00	PPBV	0.00
35) 1,4-DIFLUOROBENZENE	10.44	114	1835396	10.00	PPBV	0.00
49) CHLOROBENZENE-D5	14.67	117	1640141	10.00	PPBV	0.00

System Monitoring Compounds

61) 4-BROMOFLUOROBENZENE	16.29	95	531829	5.53	PPBV	0.00

Spiked Amount 5.000 Range 51 - 142 Recovery = 110.60%

Target Compounds						Qvalue
2) DICHLORODIFLUOROMETHANE	4.07	85	1305850	8.99	PPBV	100
3) PROPYLENE	3.99	41	178641	10.76	PPBV #	87
4) FREON 114	4.32	85	1056037	8.62	PPBV	99
5) CHLOROMETHANE	4.23	50	247208	9.17	PPBV	97
6) VINYL CHLORIDE	4.45	62	301533	8.91	PPBV	100
7) 1,3-BUTADIENE	4.58	39	204602	10.02	PPBV	97
8) BROMOMETHANE	4.85	94	360144	8.79	PPBV	99
9) CHLOROETHANE	5.03	64	140741	10.21	PPBV	99
10) TRICHLOROFLUOROMETHANE	5.79	101	1263256	8.54	PPBV	100
11) ISOPROPYL ALCOHOL	5.92	45	301406	9.67	PPBV	97
12) ACETONE	5.65	43	283156	7.72	PPBV	99
13) PENTANE	6.13	42	238500	9.59	PPBV	98
14) 1,1-DICHLOROETHYLENE	6.43	96	321653	8.69	PPBV	96
15) CARBON DISULFIDE	6.85	76	924747	8.26	PPBV	99
16) ETHANOL	5.20	45	49752	10.19	PPBV	96
17) BROMOETHENE	5.38	106	346747	9.30	PPBV	100
18) METHYLENE CHLORIDE	6.54	84	285766	7.73	PPBV	97
19) 3-CHLOROPROPENE	6.66	39	219659	9.18	PPBV	96
20) FREON 113	6.79	151	717493	7.93	PPBV	99
21) TRANS-1,2-DICHLOROETHYLENE	7.46	96	317265	8.82	PPBV	98
22) TERTIARY BUTYL ALCOHOL	6.52	59	409671	9.61	PPBV	99
23) METHYL TERTIARY BUTYL ETHE	7.73	73	534027	8.16	PPBV	98
24) TETRAHYDROFURAN	9.17	42	153701	8.55	PPBV	98
25) HEXANE	8.63	57	300535	7.52	PPBV	97
26) VINYL ACETATE	7.80	43	322349	8.07	PPBV	98
27) 1,1-DICHLOROETHANE	7.66	63	492394	8.46	PPBV	100
28) METHYL ETHYL KETONE	8.07	43	304589	8.22	PPBV	94
29) cis-1,2-DICHLOROETHYLENE	8.46	96	285978	8.56	PPBV	97
30) ETHYL ACETATE	8.68	43	324353	7.81	PPBV	99
31) CHLOROFORM	8.74	83	712177	8.31	PPBV	99
32) 1,1,1-TRICHLOROETHANE	9.68	97	854284	8.59	PPBV	99
33) CARBON TETRACHLORIDE	10.26	117	1127397	8.62	PPBV	99
34) 1,2-DICHLOROETHANE	9.45	62	382396	8.09	PPBV	100
36) BENZENE	10.12	78	669804	8.55	PPBV	100
37) CYCLOHEXANE	10.37	84	353671	9.72	PPBV	99
38) TRICHLOROETHYLENE	11.12	95	389434	8.76	PPBV	97
39) 1,2-DICHLOROPROPANE	10.90	63	214272	8.13	PPBV	98
40) BROMODICHLOROMETHANE	11.08	83	655320	7.76	PPBV	99
41) 2,2,4-TRIMETHYLPENTANE	11.12	57	1187978	10.49	PPBV	98
42) 1,4-DIOXANE	11.16	88	122751	8.14	PPBV	95
43) HEPTANE	11.36	43	323287	9.70	PPBV	99

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10286B.D
Acq On : 10 May 2007 10:33 am
Operator : danat
Sample : BS (M054)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 4 Sample Multiplier: 1

Quant Time: May 10 11:10:20 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 11:09:51 2007
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) METHYL ISOBUTYL KETONE	11.98	43	366654	8.87	PPBV	98
45) cis-1,3-DICHLOROPROPENE	11.94	75	349534	9.62	PPBV	98
46) TOLUENE	12.89	92	462783	9.02	PPBV	99
47) trans-1,3-DICHLOROPROPENE	12.44	75	290402	7.07	PPBV	99
48) 1,1,2-TRICHLOROETHANE	12.62	83	240426	8.02	PPBV	99
50) 2-HEXANONE	13.15	43	285755	7.38	PPBV	98
51) TETRACHLOROETHYLENE	14.01	164	445675	8.05	PPBV	99
52) DIBROMOCHLOROMETHANE	13.33	129	679264	7.62	PPBV	100
53) 1,2-DIBROMOETHANE	13.57	107	465149	8.46	PPBV	99
54) CHLOROBENZENE	14.71	112	714978	8.20	PPBV	98
55) ETHYLBENZENE	15.09	91	1072352	9.61	PPBV	99
56) m,p-XYLENE	15.28	106	907029	15.87	PPBV	100
57) o-XYLENE	15.78	106	461835	7.99	PPBV	100
58) STYRENE	15.67	104	610702	8.04	PPBV	100
59) NONANE	15.97	43	458405	9.71	PPBV	99
60) BROMOFORM	15.40	173	731566	8.07	PPBV	100
62) 1,1,2,2-TETRACHLOROETHANE	15.79	83	671390	8.73	PPBV	100
63) ISOPROPYLBENZENE	16.43	105	1349019	9.29	PPBV	99
64) 2-CHLOROTOLUENE	16.98	91	917142	9.59	PPBV	99
65) 4-ETHYLTOLUENE	17.17	105	1113566	7.77	PPBV	99
66) 1,3,5-TRIMETHYLBENZENE	17.25	105	1099713	9.63	PPBV	99
67) 1,2,4-TRIMETHYLBENZENE	17.72	105	998081	8.26	PPBV	99
68) m-DICHLOROBENZENE	17.91	146	797605	9.78	PPBV	99
69) BENZYL CHLORIDE	17.89	91	637252	9.90	PPBV	100
70) p-DICHLOROBENZENE	17.98	146	749201	10.05	PPBV	98
71) o-DICHLOROBENZENE	18.37	146	706204	9.72	PPBV	100
72) HEXACHLOROBUTADIENE	20.76	225	520415	11.87	PPBV	100
73) 1,2,4-TRICHLOROBENZENE	20.27	180	330262	13.97	PPBV	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10286B.D
Acq On : 10 May 2007 10:33 am
Operator : danat
Sample : BS (M054)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 4 Sample Multiplier: 1

Quant Time: May 10 11:10:20 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 11:09:51 2007
Response via : Initial Calibration



Manual Integrations APPROVED (compounds with "m" flag)
Philip Bassignani
05/11/07 17:00

Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\
Data File : J10293.D
Acq On : 10 May 2007 4:47 pm
Operator : danat
Sample : M64467-1DUP (M126)
Misc : MS13540, MSJ533,,,,,1
ALS Vial : 6 Sample Multiplier: 1

Quant Time: May 11 09:57:20 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 11:09:51 2007
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) BROMOCHLOROMETHANE	8.60	128	333619	10.00	PPBV	-0.02
35) 1,4-DIFLUOROBENZENE	10.42	114	1340934	10.00	PPBV	-0.03
49) CHLOROBENZENE-D5	14.66	117	1201491	10.00	PPBV	-0.01

System Monitoring Compounds

61) 4-BROMOFLUOROBENZENE	16.28	95	330512	4.69	PPBV	-0.01

Spiked Amount 5.000 Range 51 - 142 Recovery = 93.80%

Target Compounds						Qvalue
2) DICHLORODIFLUOROMETHANE	4.06	85	72531	0.62	PPBV	96
10) TRICHLOROFLUOROMETHANE	5.77	101	40395	0.34	PPBV	96
12) ACETONE	5.65	43	151158	5.18	PPBV	97
13) PENTANE	6.10	42	3077	0.15	PPBV #	77
15) CARBON DISULFIDE	6.83	76	24355	0.27	PPBV	91
28) METHYL ETHYL KETONE	8.09	43	9046m	0.30	PPBV	
31) CHLOROFORM	8.72	83	6584m	0.10	PPBV	
36) BENZENE	10.09	78	17289	0.30	PPBV	95
46) TOLUENE	12.88	92	4585	0.12	PPBV	94
51) TETRACHLOROETHYLENE	14.00	164	448375	11.05	PPBV	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ACCUTEST Laboratories

Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\
Data File : J10293.D
Acq On : 10 May 2007 4:47 pm
Operator : danat
Sample : M64467-1DUP (M126)
Misc : MS13540, MSJ533,,,,,1
ALS Vial : 6 Sample Multiplier: 1

Quant Time: May 11 09:57:20 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 11:09:51 2007
Response via : Initial Calibration



J050907T.M Fri May 11 09:57:47 2007 Page: 2



#2
DICHLORODIFLUOROMETHANE
Concen: 0.62 PPBV
RT: 4.06 min Scan# 833
Delta R.T. -0.00 min
Lab File: J10293.D
Acq: 10 May 2007 4:47 pm

Tgt Ion: 85 Resp: 72531

Ion	Ratio	Lower	Upper
85	100		
87	33.5	12.7	52.7
50	6.1	0.0	31.0





#10
TRICHLOROFLUOROMETHANE
Concen: 0.34 PPBV
RT: 5.77 min Scan# 1798
Delta R.T. -0.02 min
Lab File: J10293.D
Acq: 10 May 2007 4:47 pm

Tgt Ion:101 Resp: 40395

Ion	Ratio	Lower	Upper
101	100		
103	63.1	44.8	84.8
105	4.7	0.0	30.5





#12
ACETONE
Concen: 5.18 PPBV
RT: 5.65 min Scan# 1730
Delta R.T. -0.01 min
Lab File: J10293.D
Acq: 10 May 2007 4:47 pm

Tgt Ion: 43 Resp: 151158

Ion	Ratio	Lower	Upper
43	100		
58	25.5	6.8	46.8





#13
PENTANE
Concen: 0.15 PPBV
RT: 6.10 min Scan# 1987
Delta R.T. -0.03 min
Lab File: J10293.D
Acq: 10 May 2007 4:47 pm

Tgt Ion: 42 Resp: 3077

Ion	Ratio	Lower	Upper
42	100		
41	77.6	78.8	118.8#
57	11.3	7.4	47.4





#15
CARBON DISULFIDE
Concen: 0.27 PPBV
RT: 6.83 min Scan# 2397
Delta R.T. -0.02 min
Lab File: J10293.D
Acq: 10 May 2007 4:47 pm

Tgt Ion: 76 Resp: 24355

Ion	Ratio	Lower	Upper
76	100		
78	2.1	0.0	29.2
44	16.1	0.0	34.9





#28
METHYL ETHYL KETONE
Concen: 0.30 PPBV m
RT: 8.09 min Scan# 3107
Delta R.T. 0.02 min
Lab File: J10293.D
Acq: 10 May 2007 4:47 pm

Tgt Ion: 43 Resp: 9046

Ion	Ratio	Lower	Upper
43	100		
57	8.7	0.0	28.8
72	25.7	7.6	47.6





#31
CHLOROFORM
Concen: 0.10 PPBV m
RT: 8.72 min Scan# 3464
Delta R.T. -0.02 min
Lab File: J10293.D
Acq: 10 May 2007 4:47 pm

Tgt Ion: 83 Resp: 6584

Ion	Ratio	Lower	Upper
83	100		
85	0.0	44.8	84.8#
47	0.0	1.7	41.7#





#36
BENZENE
Concen: 0.30 PPBV
RT: 10.09 min Scan# 4240
Delta R.T. -0.03 min
Lab File: J10293.D
Acq: 10 May 2007 4:47 pm

Tgt Ion: 78 Resp: 17289

Ion	Ratio	Lower	Upper
78	100		
77	26.1	4.3	44.3
52	11.1	0.0	34.2





#46
TOLUENE
Concen: 0.12 PPBV
RT: 12.88 min Scan# 5810
Delta R.T. -0.01 min
Lab File: J10293.D
Acq: 10 May 2007 4:47 pm

Tgt Ion: 92 Resp: 4585

Ion	Ratio	Lower	Upper
92	100		
91	159.8	146.0	186.0
65	10.1	0.0	38.6





#51
TETRACHLOROETHYLENE
Concen: 11.05 PPBV
RT: 14.00 min Scan# 6446
Delta R.T. -0.01 min
Lab File: J10293.D
Acq: 10 May 2007 4:47 pm

Tgt Ion:164 Resp: 448375

Ion	Ratio	Lower	Upper
164	100		
129	84.3	67.6	107.6
168	60.8	42.8	82.8
131	83.7	64.5	104.5



Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10293.D Vial: 6
Acq On : 10 May 2007 4:47 pm Operator: danat
Sample : M64467-1DUP (M126) Inst : MSJ
Misc : MS13540, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 17:14:10 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 11:09:51 2007
Response via : Multiple Level Calibration







(28) METHYL ETHYL KETONE (m)

8.09min 0.21PPBV

response 6209

Ion	Exp%	Act%
43.00	100	100
57.00	8.80	8.74
72.00	27.60	25.70
0.00	0.00	0.00

J10293.D J050907T.M Thu May 10 17:15:32 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10293.D Vial: 6
Acq On : 10 May 2007 4:47 pm Operator: danat
Sample : M64467-1DUP (M126) Inst : MSJ
Misc : MS13540, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 17:14:10 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 11:09:51 2007
Response via : Multiple Level Calibration







(28) METHYL ETHYL KETONE (m)

8.09min 0.30PPBV m

response 9046

Ion	Exp%	Act%
43.00	100	100
57.00	8.80	8.74
72.00	27.60	25.70
0.00	0.00	0.00

J10293.D J050907T.M Thu May 10 17:15:45 2007

BFB

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um





Spectrum Information: Average of 16.263 to 16.281 min.

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	17.7	12561	PASS
75	95	30	66	49.5	35015	PASS
95	95	100	100	100.0	70793	PASS
96	95	5	9	6.6	4662	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	105.0	74325	PASS
175	174	4	9	6.9	5105	PASS
176	174	93	101	93.8	69706	PASS
177	176	5	9	7.3	5111	PASS

J10277.D J050907T.M Thu May 10 12:01:54 2007

Average of 16.263 to 16.281 min.: J10277.D
IC533-0.2 (M058)

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
35.50	59	40.05	2707	50.15	12561	60.15	636
35.90	54	44.05	1187	51.15	4218	61.15	3432
36.00	73	45.20	443	51.90	42	62.10	4086
36.10	359	45.70	51	52.15	88	63.15	3087
36.20	106	46.40	41	52.85	131	64.05	186
36.30	86	46.80	38	53.20	55	64.25	367
36.50	76	46.95	237	54.90	37	65.15	1110
37.10	86	47.25	436	55.80	47	66.20	45
37.25	2989	47.80	83	56.10	787	67.10	37
38.15	3048	48.15	249	57.10	2251	68.10	7805
39.15	1478	49.15	3504	59.95	433	69.05	7535

Average of 16.263 to 16.281 min.: J10277.D
IC533-0.2 (M058)

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
69.80	50	77.35	108	88.90	51	103.70	64
70.00	148	77.90	197	90.95	220	104.05	294
70.15	313	79.00	3245	91.20	107	104.80	44
71.10	52	80.10	925	92.05	2174	105.10	41
71.85	140	80.95	3533	93.05	3251	105.70	87
72.15	355	82.00	848	94.10	10204	106.00	297
73.10	3854	86.00	61	95.10	70793	107.20	43
74.15	11943	86.15	152	96.05	4662	110.30	67
75.10	35015	86.90	283	96.90	41	111.90	50
76.10	3357	87.05	2199	97.20	90	112.90	44
77.00	239	88.05	2442	103.30	48	114.80	80

Average of 16.263 to 16.281 min.: J10277.D
IC533-0.2 (M058)

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
116.05	235	134.80	65	148.60	48	170.80	44
117.00	621	136.60	39	149.80	47	171.05	92
117.20	101	137.00	54	153.60	57	171.95	1392
117.90	296	140.75	197	153.90	55	172.30	103
118.95	524	140.95	721	154.80	84	174.00	74325
122.90	38	142.90	1780	156.85	159	175.05	5105
127.90	193	143.90	110	158.75	119	176.00	69706
129.20	45	144.90	88	159.00	36	176.95	5111
129.80	54	145.05	610	160.70	44	178.00	75
130.00	67	145.90	130	160.90	76	191.70	38
131.00	46	147.95	326	170.10	40	192.10	60

Average of 16.263 to 16.281 min.: J10277.D
IC533-0.2 (M058)

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
206.85	95						
253.80	41						
267.05	91						
268.10	52						
280.20	38						
281.80	38						
284.90	37						
294.70	36						

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration



(3) PROPYLENE (m)

3.99min 0.04PPBV

response 567

Ion	Exp%	Act%
41.00	100	100
39.00	71.40	0.00#
42.00	62.70	0.00#
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:35:35 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration



(3) PROPYLENE (m)

4.00min 0.09PPBV m

response 1240

Ion	Exp%	Act%
41.00	100	100
39.00	71.40	45.77#
42.00	62.70	64.25
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:35:44 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration







(5) CHLOROMETHANE (m)

4.22min 0.12PPBV

response 2652

Ion	Exp%	Act%
50.00	100	100
52.00	32.60	29.05
0.00	0.00	0.00
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:35:58 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration







(5) CHLOROMETHANE (m)

4.22min 0.21PPBV m

response 4633

Ion	Exp%	Act%
50.00	100	100
52.00	32.60	34.59
0.00	0.00	0.00
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:36:03 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration







(9) CHLOROETHANE (m)

5.02min 0.05PPBV

response 557

Ion	Exp%	Act%
64.00	100	100
66.00	33.10	22.08
49.00	30.20	11.49
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:36:16 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration



(9) CHLOROETHANE (m)

5.04min 0.12PPBV m

response 1417

Ion	Exp%	Act%
64.00	100	100
66.00	33.10	8.68#
49.00	30.20	4.52#
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:36:22 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration



(11) ISOPROPYL ALCOHOL (m)

6.01min 0.01PPBV

response 160

Ion	Exp%	Act%
45.00	100	100
59.00	3.80	0.00
43.00	25.40	0.00#
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:36:30 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration







(11) ISOPROPYL ALCOHOL (m)

6.05min 0.28PPBV m

response 6895

Ion	Exp%	Act%
45.00	100	100
59.00	3.80	0.00
43.00	25.40	0.00#
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:36:36 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration



(14) 1,1-DICHLOROETHYLENE (m)

6.42min 0.11PPBV

response 3441

Ion	Exp%	Act%
96.00	100	100
61.00	157.70	284.42#
63.00	52.40	62.05
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:36:49 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration



(14) 1,1-DICHLOROETHYLENE (m)

6.42min 0.19PPBV m

response 5906

Ion	Exp%	Act%
96.00	100	100
61.00	157.70	165.71
63.00	52.40	36.15
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:36:54 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration







(30) ETHYL ACETATE (m)

8.67min 0.00PPBV

response 151

Ion	Exp%	Act%
43.00	100	100
62.00	0.30	34.44#
88.00	4.70	0.00
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:37:49 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration







(30) ETHYL ACETATE (m)

8.72min 0.16PPBV m

response 5570

Ion	Exp%	Act%
43.00	100	100
62.00	0.30	0.93
88.00	4.70	0.00
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:38:02 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration







(56) m,p-XYLENE (m)

15.29min 0.10PPBV

response 3552

Ion	Exp%	Act%
106.00	100	100
91.00	187.40	113.31#
77.00	26.20	23.79
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:39:14 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration







(56) m,p-XYLENE (m)

15.29min 0.18PPBV m

response 6881

Ion	Exp%	Act%
106.00	100	100
91.00	187.40	113.31#
77.00	26.20	23.79
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:39:20 2007

BFB

Data File : C:\MSDChem\1\data\J10286A.D
Acq On : 10 May 2007 10:33 am
Sample : CC533-10 (M054)
Misc : MS13434, MSJ533,,,,,1
MS Integration Params: LSCINT.P

Vial: 4
Operator: danat
Inst : MSJ
Multiplr: 1.00

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um





Spectrum Information: Average of 16.288 to 16.293 min.

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	17.6	58752	PASS
75	95	30	66	48.9	163184	PASS
95	95	100	100	100.0	333840	PASS
96	95	5	9	6.9	23018	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	107.6	359152	PASS
175	174	4	9	4.2	15166	PASS
176	174	93	101	97.1	348720	PASS
177	176	5	9	6.2	21654	PASS

J10286A.D J050907T.M Thu May 10 14:38:50 2007

Average of 16.288 to 16.293 min.: J10286A.D
ICV533-10 (M054)

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
36.20	3491	47.15	2284	60.15	3179	73.15	15163
37.15	13954	48.15	2073	61.10	14780	74.15	57210
38.15	12524	49.20	13885	62.15	13302	75.10	163184
39.10	4889	50.15	58752	63.15	12600	76.10	15193
40.05	2916	51.15	17862	64.15	1730	77.05	1904
41.00	170	52.10	1231	65.10	2439	78.15	1231
43.20	112	53.15	211	68.10	35584	79.00	9849
44.10	3082	55.10	1094	69.10	31824	80.05	3326
45.15	2603	56.15	5036	70.05	2813	81.05	11335
45.80	108	57.15	9032	71.20	116	82.00	2878
46.05	279	58.10	164	72.20	2264	83.00	124

Average of 16.288 to 16.293 min.: J10286A.D
ICV533-10 (M054)

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
86.00	649	102.10	100	113.10	146	126.10	190
87.05	11920	103.30	276	115.00	134	127.95	1534
88.10	11211	104.00	1596	115.90	800	128.80	136
91.05	1552	105.05	681	116.05	863	128.95	486
92.10	10183	106.00	1918	117.00	2454	129.95	1221
93.10	16963	107.05	547	117.95	1349	131.00	751
94.15	43468	108.20	137	118.70	187	134.85	468
95.10	333840	109.00	109	118.95	1608	135.10	215
96.10	23018	109.90	117	122.00	180	136.85	563
97.00	147	111.00	126	122.90	103	138.95	501
97.20	162	112.05	266	124.00	127	140.00	105

Average of 16.288 to 16.293 min.: J10286A.D
ICV533-10 (M054)

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
141.00	4418	150.05	366	174.00	359152	251.75	314
141.80	323	151.75	293	175.10	15166	259.40	141
142.00	322	152.95	309	176.00	348720	276.90	116
143.00	5192	153.70	118	176.95	21654	290.30	157
143.90	103	154.80	146	177.90	794		
144.10	281	155.00	675	192.20	187		
144.95	865	156.85	604	200.70	102		
145.85	299	158.90	528	202.80	105		
147.95	1139	160.95	470	207.10	273		
148.80	246	170.40	107	235.80	124		
149.10	101	171.95	3601	241.60	135		

Manual Integrations APPROVED (compounds with "m" flag)
Philip Bassignani
05/11/07 16:37

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10277.D
Acq On : 9 May 2007 6:39 pm
Operator : danat
Sample : IC533-0.2 (M058)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 2 Sample Multiplier: 1

Quant Time: May 10 08:40:19 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 08:35:02 2007
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) BROMOCHLOROMETHANE	8.62	128	344690	10.00	PPBV	0.00
35) 1,4-DIFLUOROBENZENE	10.43	114	1460896	10.00	PPBV	0.00
49) CHLOROBENZENE-D5	14.67	117	1214337	10.00	PPBV	0.00

System Monitoring Compounds

61) 4-BROMOFLUOROBENZENE	16.29	95	295885	4.04	PPBV	0.00

Spiked Amount 5.000 Range 51 - 142 Recovery = 80.80%

Target Compounds						Qvalue
2) DICHLORODIFLUOROMETHANE	4.07	85	26123	0.22	PPBV #	52
3) PROPYLENE	4.00	41	1240m	0.09	PPBV	
4) FREON 114	4.32	85	22207	0.22	PPBV	98
5) CHLOROMETHANE	4.22	50	4633m	0.21	PPBV	
6) VINYL CHLORIDE	4.46	62	5596	0.20	PPBV #	66
7) 1,3-BUTADIENE	4.59	39	2028	0.11	PPBV	91
8) BROMOMETHANE	4.85	94	6867	0.20	PPBV #	70
9) CHLOROETHANE	5.04	64	1417m	0.12	PPBV	
10) TRICHLOROFLUOROMETHANE	5.79	101	27025	0.22	PPBV	96
11) ISOPROPYL ALCOHOL	6.05	45	6895m	0.28	PPBV	
12) ACETONE	5.68	43	14047	0.43	PPBV	97
13) PENTANE	6.14	42	3329	0.16	PPBV #	60
14) 1,1-DICHLOROETHYLENE	6.42	96	5906m	0.19	PPBV	
15) CARBON DISULFIDE	6.85	76	20235	0.22	PPBV #	77
17) BROMOETHENE	5.38	106	5175	0.16	PPBV #	94
19) 3-CHLOROPROPENE	6.66	39	3960	0.20	PPBV #	58
20) FREON 113	6.79	151	17449	0.24	PPBV	92
21) TRANS-1,2-DICHLOROETHYLENE	7.46	96	5737	0.19	PPBV	95
23) METHYL TERTIARY BUTYL ETHE	7.76	73	7657	0.13	PPBV	86
25) HEXANE	8.63	57	8678	0.27	PPBV	96
27) 1,1-DICHLOROETHANE	7.66	63	9424	0.19	PPBV	97
28) METHYL ETHYL KETONE	8.11	43	5063	0.16	PPBV #	54
29) cis-1,2-DICHLOROETHYLENE	8.46	96	4723	0.17	PPBV	94
30) ETHYL ACETATE	8.72	43	5570m	0.16	PPBV	
31) CHLOROFORM	8.75	83	14593	0.20	PPBV	95
32) 1,1,1-TRICHLOROETHANE	9.68	97	16547	0.20	PPBV	100
33) CARBON TETRACHLORIDE	10.25	117	22639	0.21	PPBV	99
34) 1,2-DICHLOROETHANE	9.45	62	7449	0.19	PPBV #	67
36) BENZENE	10.12	78	10499	0.16	PPBV	77
37) CYCLOHEXANE	10.37	84	3863	0.13	PPBV	92
38) TRICHLOROETHYLENE	11.12	95	5440	0.15	PPBV	96
39) 1,2-DICHLOROPROPANE	10.90	63	3567	0.17	PPBV	87
40) BROMODICHLOROMETHANE	11.08	83	13502	0.20	PPBV	100
41) 2,2,4-TRIMETHYLPENTANE	11.12	57	10934	0.11	PPBV	74
43) HEPTANE	11.36	43	2910	0.10	PPBV #	72
45) cis-1,3-DICHLOROPROPENE	11.94	75	3432	0.11	PPBV	90
46) TOLUENE	12.88	92	4977	0.11	PPBV #	74
47) trans-1,3-DICHLOROPROPENE	12.43	75	2652	0.09	PPBV #	66
48) 1,1,2-TRICHLOROETHANE	12.62	83	4402	0.18	PPBV	97
51) TETRACHLOROETHYLENE	14.01	164	7177	0.17	PPBV	96
52) DIBROMOCHLOROMETHANE	13.32	129	12097	0.18	PPBV	99

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10277.D
Acq On : 9 May 2007 6:39 pm
Operator : danat
Sample : IC533-0.2 (M058)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 2 Sample Multiplier: 1

Quant Time: May 10 08:40:19 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 08:35:02 2007
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
53) 1,2-DIBROMOETHANE	13.57	107	6198	0.15	PPBV	90
54) CHLOROBENZENE	14.71	112	10764	0.16	PPBV	82
55) ETHYLBENZENE	15.09	91	8563	0.10	PPBV	92
56) m,p-XYLENE	15.29	106	6881m	0.18	PPBV	
57) o-XYLENE	15.78	106	3329	0.09	PPBV	91
60) BROMOFORM	15.39	173	12175	0.18	PPBV	92
62) 1,1,2,2-TETRACHLOROETHANE	15.79	83	8853	0.15	PPBV	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

6.6.1

6

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10277.D
Acq On : 9 May 2007 6:39 pm
Operator : danat
Sample : IC533-0.2 (M058)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 2 Sample Multiplier: 1

Quant Time: May 10 08:40:19 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 08:35:02 2007
Response via : Initial Calibration



Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration



(3) PROPYLENE (m)

3.99min 0.04PPBV

response 567

Ion	Exp%	Act%
41.00	100	100
39.00	71.40	0.00#
42.00	62.70	0.00#
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:35:35 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration







(3) PROPYLENE (m)

4.00min 0.09PPBV m

response 1240

Ion	Exp%	Act%
41.00	100	100
39.00	71.40	45.77#
42.00	62.70	64.25
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:35:44 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration







(5) CHLOROMETHANE (m)

4.22min 0.12PPBV

response 2652

Ion	Exp%	Act%
50.00	100	100
52.00	32.60	29.05
0.00	0.00	0.00
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:35:58 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration







(5) CHLOROMETHANE (m)

4.22min 0.21PPBV m

response 4633

Ion	Exp%	Act%
50.00	100	100
52.00	32.60	34.59
0.00	0.00	0.00
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:36:03 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration







(9) CHLOROETHANE (m)

5.02min 0.05PPBV

response 557

Ion	Exp%	Act%
64.00	100	100
66.00	33.10	22.08
49.00	30.20	11.49
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:36:16 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration







(9) CHLOROETHANE (m)

5.04min 0.12PPBV m

response 1417

Ion	Exp%	Act%
64.00	100	100
66.00	33.10	8.68#
49.00	30.20	4.52#
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:36:22 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration



(11) ISOPROPYL ALCOHOL (m)

6.01min 0.01PPBV

response 160

Ion	Exp%	Act%
45.00	100	100
59.00	3.80	0.00
43.00	25.40	0.00#
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:36:30 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration



(11) ISOPROPYL ALCOHOL (m)

6.05min 0.28PPBV m

response 6895

Ion	Exp%	Act%
45.00	100	100
59.00	3.80	0.00
43.00	25.40	0.00#
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:36:36 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration



(14) 1,1-DICHLOROETHYLENE (m)

6.42min 0.11PPBV

response 3441

Ion	Exp%	Act%
96.00	100	100
61.00	157.70	284.42#
63.00	52.40	62.05
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:36:49 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration



(14) 1,1-DICHLOROETHYLENE (m)

6.42min 0.19PPBV m

response 5906

Ion	Exp%	Act%
96.00	100	100
61.00	157.70	165.71
63.00	52.40	36.15
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:36:54 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration



(30) ETHYL ACETATE (m)

8.67min 0.00PPBV

response 151

Ion	Exp%	Act%
43.00	100	100
62.00	0.30	34.44#
88.00	4.70	0.00
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:37:49 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration



(30) ETHYL ACETATE (m)

8.72min 0.16PPBV m

response 5570

Ion	Exp%	Act%
43.00	100	100
62.00	0.30	0.93
88.00	4.70	0.00
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:38:02 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration







(56) m,p-XYLENE (m)

15.29min 0.10PPBV

response 3552

Ion	Exp%	Act%
106.00	100	100
91.00	187.40	113.31#
77.00	26.20	23.79
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:39:14 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10277.D Vial: 2
Acq On : 9 May 2007 6:39 pm Operator: danat
Sample : IC533-0.2 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:35:17 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:35:02 2007
Response via : Multiple Level Calibration







(56) m,p-XYLENE (m)

15.29min 0.18PPBV m

response 6881

Ion	Exp%	Act%
106.00	100	100
91.00	187.40	113.31#
77.00	26.20	23.79
0.00	0.00	0.00

J10277.D J050907T.M Thu May 10 08:39:20 2007

Manual Integrations APPROVED (compounds with "m" flag)
Philip Bassignani
05/11/07 16:35

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10278.D
Acq On : 9 May 2007 7:29 pm
Operator : danat
Sample : IC533-0.5 (M058)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 2 Sample Multiplier: 1

Quant Time: May 10 08:34:15 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 08:30:48 2007
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) BROMOCHLOROMETHANE	8.62	128	303253	10.00	PPBV	0.00
35) 1,4-DIFLUOROBENZENE	10.44	114	1271837	10.00	PPBV	0.00
49) CHLOROBENZENE-D5	14.67	117	1022579	10.00	PPBV	0.00

System Monitoring Compounds

61) 4-BROMOFLUOROBENZENE	16.29	95	248904	3.88	PPBV	0.00

Spiked Amount 5.000 Range 51 - 142 Recovery = 77.60%

Target Compounds							Qvalue
2) DICHLORODIFLUOROMETHANE	4.06	85	59841	0.59	PPBV		99
3) PROPYLENE	3.99	41	3304	0.25	PPBV	#	64
4) FREON 114	4.32	85	49964	0.58	PPBV		100
5) CHLOROMETHANE	4.22	50	10199	0.52	PPBV		72
6) VINYL CHLORIDE	4.44	62	14859	0.62	PPBV		92
7) 1,3-BUTADIENE	4.59	39	7393	0.46	PPBV	#	58
8) BROMOMETHANE	4.86	94	16476	0.56	PPBV	#	14
9) CHLOROETHANE	5.03	64	5817	0.55	PPBV	#	64
10) TRICHLOROFLUOROMETHANE	5.80	101	64408	0.63	PPBV		94
11) ISOPROPYL ALCOHOL	5.98	45	8617	0.38	PPBV		71
12) ACETONE	5.67	43	20402	0.79	PPBV		90
13) PENTANE	6.14	42	10013	0.54	PPBV		96
14) 1,1-DICHLOROETHYLENE	6.42	96	12550	0.45	PPBV	#	82
15) CARBON DISULFIDE	6.84	76	46974	0.60	PPBV		96
16) ETHANOL	5.23	45	1054	0.27	PPBV		88
17) BROMOETHENE	5.38	106	12685	0.44	PPBV		99
18) METHYLENE CHLORIDE	6.55	84	101487	4.11	PPBV		96
19) 3-CHLOROPROPENE	6.65	39	10064	0.59	PPBV	#	11
20) FREON 113	6.79	151	38798	0.62	PPBV		98
21) TRANS-1,2-DICHLOROETHYLENE	7.45	96	14508	0.56	PPBV		97
22) TERTIARY BUTYL ALCOHOL	6.60	59	9940	0.30	PPBV		76
23) METHYL TERTIARY BUTYL ETHE	7.74	73	25551	0.51	PPBV		80
24) TETRAHYDROFURAN	9.23	42	4448m	0.32	PPBV		
25) HEXANE	8.63	57	18242	0.70	PPBV		91
26) VINYL ACETATE	7.79	43	12990	0.43	PPBV		86
27) 1,1-DICHLOROETHANE	7.66	63	25680	0.62	PPBV		98
28) METHYL ETHYL KETONE	8.08	43	13917	0.50	PPBV		66
29) cis-1,2-DICHLOROETHYLENE	8.46	96	13138	0.53	PPBV		98
30) ETHYL ACETATE	8.70	43	16867	0.54	PPBV		95
31) CHLOROFORM	8.74	83	37104	0.61	PPBV		99
32) 1,1,1-TRICHLOROETHANE	9.68	97	41748	0.59	PPBV		99
33) CARBON TETRACHLORIDE	10.26	117	55167	0.60	PPBV		96
34) 1,2-DICHLOROETHANE	9.45	62	19616	0.57	PPBV		94
36) BENZENE	10.12	78	28276	0.51	PPBV		98
37) CYCLOHEXANE	10.37	84	10724	0.39	PPBV		95
38) TRICHLOROETHYLENE	11.11	95	16450	0.52	PPBV		96
39) 1,2-DICHLOROPROPANE	10.89	63	10804	0.60	PPBV	#	85
40) BROMODICHLOROMETHANE	11.09	83	35777	0.64	PPBV		99
41) 2,2,4-TRIMETHYLPENTANE	11.12	57	33493	0.39	PPBV		89
42) 1,4-DIOXANE	11.24	88	808m	0.09	PPBV		
43) HEPTANE	11.36	43	9865	0.38	PPBV		94

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10278.D
Acq On : 9 May 2007 7:29 pm
Operator : danat
Sample : IC533-0.5 (M058)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 2 Sample Multiplier: 1

Quant Time: May 10 08:34:15 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 08:30:48 2007
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) METHYL ISOBUTYL KETONE	12.01	43	9402	0.31	PPBV		83
45) cis-1,3-DICHLOROPROPENE	11.94	75	10454	0.37	PPBV		94
46) TOLUENE	12.89	92	14159	0.36	PPBV		91
47) trans-1,3-DICHLOROPROPENE	12.44	75	7429	0.27	PPBV		94
48) 1,1,2-TRICHLOROETHANE	12.62	83	11654	0.57	PPBV		91
50) 2-HEXANONE	13.20	43	5962m	0.26	PPBV		
51) TETRACHLOROETHYLENE	14.01	164	21493	0.64	PPBV		96
52) DIBROMOCHLOROMETHANE	13.32	129	36088	0.68	PPBV		94
53) 1,2-DIBROMOETHANE	13.57	107	18069	0.51	PPBV		95
54) CHLOROBENZENE	14.71	112	30195	0.55	PPBV		98
55) ETHYLBENZENE	15.09	91	26982	0.34	PPBV		96
56) m,p-XYLENE	15.27	106	20949	0.63	PPBV	#	56
57) o-XYLENE	15.77	106	10835	0.32	PPBV		98
58) STYRENE	15.66	104	8882	0.21	PPBV		96
59) NONANE	15.97	43	10837	0.35	PPBV		93
60) BROMOFORM	15.39	173	33030	0.59	PPBV		95
62) 1,1,2,2-TETRACHLOROETHANE	15.79	83	30492	0.64	PPBV		98
63) ISOPROPYLBENZENE	16.43	105	32943	0.35	PPBV		98
64) 2-CHLOROTOLUENE	16.97	91	21360	0.34	PPBV		99
65) 4-ETHYLTOLUENE	17.17	105	20980	0.25	PPBV		98
66) 1,3,5-TRIMETHYLBENZENE	17.25	105	23503	0.31	PPBV		97
67) 1,2,4-TRIMETHYLBENZENE	17.72	105	17132	0.25	PPBV		97
68) m-DICHLOROBENZENE	17.90	146	18344	0.34	PPBV		95
69) BENZYL CHLORIDE	17.88	91	11159	0.27	PPBV		86
70) p-DICHLOROBENZENE	17.98	146	15394	0.31	PPBV		93
71) o-DICHLOROBENZENE	18.37	146	19372	0.42	PPBV		97
72) HEXACHLOROBUTADIENE	20.77	225	57566	1.98	PPBV		97
73) 1,2,4-TRICHLOROBENZENE	20.28	180	7369	0.50	PPBV		98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10278.D
Acq On : 9 May 2007 7:29 pm
Operator : danat
Sample : IC533-0.5 (M058)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 2 Sample Multiplier: 1

Quant Time: May 10 08:34:15 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 08:30:48 2007
Response via : Initial Calibration



Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10278.D Vial: 2
Acq On : 9 May 2007 7:29 pm Operator: danat
Sample : IC533-0.5 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:31:42 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:30:48 2007
Response via : Multiple Level Calibration



(24) TETRAHYDROFURAN (m)

9.21min 0.18PPBV

response 2563

Ion	Exp%	Act%
42.00	100	100
72.00	51.30	27.08#
71.00	52.50	65.90
0.00	0.00	0.00

J10278.D J050907T.M Thu May 10 08:32:39 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10278.D Vial: 2
Acq On : 9 May 2007 7:29 pm Operator: danat
Sample : IC533-0.5 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:31:42 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:30:48 2007
Response via : Multiple Level Calibration





(24) TETRAHYDROFURAN (m)

9.23min 0.32PPBV m

response 4448

Ion	Exp%	Act%
42.00	100	100
72.00	51.30	15.60#
71.00	52.50	37.97
0.00	0.00	0.00

J10278.D J050907T.M Thu May 10 08:32:46 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10278.D Vial: 2
Acq On : 9 May 2007 7:29 pm Operator: danat
Sample : IC533-0.5 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:31:42 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:30:48 2007
Response via : Multiple Level Calibration





(42) 1,4-DIOXANE (m)

11.22min 0.01PPBV

response 55

Ion	Exp%	Act%
88.00	100	100
57.00	26.70	0.00#
58.00	51.70	0.00#
43.00	24.80	0.00#

J10278.D J050907T.M Thu May 10 08:33:15 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10278.D Vial: 2
Acq On : 9 May 2007 7:29 pm Operator: danat
Sample : IC533-0.5 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:31:42 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:30:48 2007
Response via : Multiple Level Calibration





(50) 2-HEXANONE (m)

13.18min 0.13PPBV

response 2999

Ion	Exp%	Act%
43.00	100	100
58.00	52.90	72.66
100.00	11.70	5.40
0.00	0.00	0.00

J10278.D J050907T.M Thu May 10 08:33:47 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10278.D Vial: 2
Acq On : 9 May 2007 7:29 pm Operator: danat
Sample : IC533-0.5 (M058) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:31:42 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:30:48 2007
Response via : Multiple Level Calibration





(50) 2-HEXANONE (m)

13.20min 0.26PPBV m

response 5962

Ion	Exp%	Act%
43.00	100	100
58.00	52.90	36.55
100.00	11.70	2.72
0.00	0.00	0.00

J10278.D J050907T.M Thu May 10 08:33:53 2007

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10279.D
Acq On : 9 May 2007 8:18 pm
Operator : danat
Sample : IC533-40 (M079)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 3 Sample Multiplier: 1

Quant Time: May 10 08:26:33 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 08:25:12 2007
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) BROMOCHLOROMETHANE	8.63	128	404357	10.00	PPBV	0.01
35) 1,4-DIFLUOROBENZENE	10.45	114	1793963	10.00	PPBV	0.01
49) CHLOROBENZENE-D5	14.68	117	1690369	10.00	PPBV	0.01

System Monitoring Compounds

61) 4-BROMOFLUOROBENZENE	16.31	95	598108	5.36	PPBV	0.02

Spiked Amount 5.000 Range 51 - 142 Recovery = 107.20%

Target Compounds						Qvalue
2) DICHLORODIFLUOROMETHANE	4.06	85	5045782	38.33	PPBV	100
3) PROPYLENE	3.99	41	713566	38.59	PPBV	89
4) FREON 114	4.32	85	4377290	39.66	PPBV	99
5) CHLOROMETHANE	4.23	50	974089	38.95	PPBV	96
6) VINYL CHLORIDE	4.45	62	1234612	38.44	PPBV	98
7) 1,3-BUTADIENE	4.58	39	893531	41.12	PPBV	98
8) BROMOMETHANE	4.86	94	1506649	39.48	PPBV	99
9) CHLOROETHANE	5.02	64	594297	40.67	PPBV	100
10) TRICHLOROFLUOROMETHANE	5.80	101	5287371	40.57	PPBV	100
11) ISOPROPYL ALCOHOL	5.97	45	1378514	40.51	PPBV	98
12) ACETONE	5.65	43	1478585	42.51	PPBV	99
13) PENTANE	6.14	42	973868	40.12	PPBV	98
14) 1,1-DICHLOROETHYLENE	6.43	96	1508099	40.21	PPBV	97
15) CARBON DISULFIDE	6.85	76	3991509	39.44	PPBV	99
16) ETHANOL	5.20	45	232995	42.55	PPBV #	57
17) BROMOETHENE	5.38	106	1567714	40.51	PPBV	100
18) METHYLENE CHLORIDE	6.55	84	1330636	41.19	PPBV	99
19) 3-CHLOROPROPENE	6.66	39	1065920	46.84	PPBV	97
20) FREON 113	6.79	151	3304804	41.16	PPBV	99
21) TRANS-1,2-DICHLOROETHYLENE	7.46	96	1468282	42.02	PPBV	98
22) TERTIARY BUTYL ALCOHOL	6.57	59	2031691	41.97	PPBV	94
23) METHYL TERTIARY BUTYL ETHE	7.73	73	2972098	43.06	PPBV	100
24) TETRAHYDROFURAN	9.18	42	887741	44.28	PPBV	99
25) HEXANE	8.63	57	1482683	43.06	PPBV	97
26) VINYL ACETATE	7.81	43	1907349	45.21	PPBV	98
27) 1,1-DICHLOROETHANE	7.66	63	2350712	43.53	PPBV	100
28) METHYL ETHYL KETONE	8.08	43	1636032	42.50	PPBV	97
29) cis-1,2-DICHLOROETHYLENE	8.47	96	1467409	44.74	PPBV	99
30) ETHYL ACETATE	8.70	43	1831614	42.86	PPBV	99
31) CHLOROFORM	8.76	83	3420136	43.96	PPBV	99
32) 1,1,1-TRICHLOROETHANE	9.69	97	4199585	45.18	PPBV	99
33) CARBON TETRACHLORIDE	10.26	117	5183665	43.27	PPBV	100
34) 1,2-DICHLOROETHANE	9.47	62	2052096	46.44	PPBV	99
36) BENZENE	10.13	78	3765517	47.71	PPBV	100
37) CYCLOHEXANE	10.38	84	1728156	43.53	PPBV	100
38) TRICHLOROETHYLENE	11.13	95	1979928	45.92	PPBV	97
39) 1,2-DICHLOROPROPANE	10.91	63	1215349	48.02	PPBV	99
40) BROMODICHLOROMETHANE	11.11	83	3527334	46.31	PPBV	100
41) 2,2,4-TRIMETHYLPENTANE	11.13	57	5658069	44.13	PPBV	100
42) 1,4-DIOXANE	11.18	88	699433	46.50	PPBV	98
43) HEPTANE	11.37	43	1725520	46.42	PPBV	98

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10279.D
Acq On : 9 May 2007 8:18 pm
Operator : danat
Sample : IC533-40 (M079)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 3 Sample Multiplier: 1

Quant Time: May 10 08:26:33 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 08:25:12 2007
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) METHYL ISOBUTYL KETONE	12.01	43	2181947	47.85	PPBV	100
45) cis-1,3-DICHLOROPROPENE	11.95	75	2021787	48.53	PPBV	99
46) TOLUENE	12.90	92	2853663	49.13	PPBV	100
47) trans-1,3-DICHLOROPROPENE	12.46	75	2009963	49.38	PPBV	100
48) 1,1,2-TRICHLOROETHANE	12.65	83	1349208	47.18	PPBV	100
50) 2-HEXANONE	13.18	43	1955501	48.86	PPBV	98
51) TETRACHLOROETHYLENE	14.02	164	2604938	48.09	PPBV	98
52) DIBROMOCHLOROMETHANE	13.35	129	3875669	45.99	PPBV	100
53) 1,2-DIBROMOETHANE	13.59	107	2742121	47.00	PPBV	99
54) CHLOROBENZENE	14.73	112	4037572	45.72	PPBV	100
55) ETHYLBENZENE	15.11	91	6386800	46.94	PPBV	99
56) m,p-XYLENE	15.29	106	5293566	93.91	PPBV	98
57) o-XYLENE	15.80	106	2702230	47.71	PPBV	99
58) STYRENE	15.69	104	3892158	50.18	PPBV	99
59) NONANE	15.99	43	2462505	46.64	PPBV	98
60) BROMOFORM	15.42	173	4304858	47.41	PPBV	99
62) 1,1,2,2-TETRACHLOROETHANE	15.82	83	3627252	46.66	PPBV	99
63) ISOPROPYLBENZENE	16.44	105	7832569	47.78	PPBV	99
64) 2-CHLOROTOLUENE	17.00	91	5399300	49.15	PPBV	99
65) 4-ETHYLTOLUENE	17.18	105	7335819	50.40	PPBV	97
66) 1,3,5-TRIMETHYLBENZENE	17.27	105	6548425	49.39	PPBV	99
67) 1,2,4-TRIMETHYLBENZENE	17.73	105	6245441	52.23	PPBV	99
68) m-DICHLOROBENZENE	17.92	146	4624539	52.96	PPBV	98
69) BENZYL CHLORIDE	17.90	91	3959343	55.29	PPBV	99
70) p-DICHLOROBENZENE	17.99	146	4326368	52.50	PPBV	100
71) o-DICHLOROBENZENE	18.38	146	4142730	53.59	PPBV	99
72) HEXACHLOROBUTADIENE	20.76	225	2354646	50.90	PPBV	99
73) 1,2,4-TRICHLOROBENZENE	20.27	180	1390577	53.70	PPBV	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10279.D
Acq On : 9 May 2007 8:18 pm
Operator : danat
Sample : IC533-40 (M079)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 3 Sample Multiplier: 1

Quant Time: May 10 08:26:33 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 08:25:12 2007
Response via : Initial Calibration



ACCUTEST Laboratories
M64467

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10280.D
Acq On : 9 May 2007 9:05 pm
Operator : danat
Sample : IC533-20 (M079)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 3 Sample Multiplier: 1

Quant Time: May 10 08:25:03 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 08:23:49 2007
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) BROMOCHLOROMETHANE	8.63	128	404934	10.00	PPBV	0.01
35) 1,4-DIFLUOROBENZENE	10.45	114	1815477	10.00	PPBV	0.00
49) CHLOROBENZENE-D5	14.68	117	1676330	10.00	PPBV	0.00
System Monitoring Compounds						
61) 4-BROMOFLUOROBENZENE	16.30	95	561983	5.17	PPBV	0.00
Spiked Amount 5.000	Range 51 - 142		Recovery = 103.40%			
Target Compounds						Qvalue
2) DICHLORODIFLUOROMETHANE	4.07	85	2650639	20.22	PPBV	99
3) PROPYLENE	3.99	41	376167	20.64	PPBV	93
4) FREON 114	4.33	85	2252269	20.77	PPBV	100
5) CHLOROMETHANE	4.23	50	512360	20.94	PPBV	100
6) VINYL CHLORIDE	4.45	62	657359	20.90	PPBV	99
7) 1,3-BUTADIENE	4.59	39	452627	21.67	PPBV	99
8) BROMOMETHANE	4.86	94	786147	21.17	PPBV	99
9) CHLOROETHANE	5.03	64	301582	21.26	PPBV	99
10) TRICHLOROFLUOROMETHANE	5.79	101	2644990	20.54	PPBV	99
11) ISOPROPYL ALCOHOL	5.94	45	730444	23.09	PPBV	99
12) ACETONE	5.65	43	734728	22.31	PPBV	99
13) PENTANE	6.14	42	491846	20.47	PPBV	99
14) 1,1-DICHLOROETHYLENE	6.43	96	771388	21.10	PPBV	99
15) CARBON DISULFIDE	6.85	76	2068017	20.83	PPBV	99
16) ETHANOL	5.20	45	118793	23.62	PPBV #	54
17) BROMOETHENE	5.39	106	806417	21.69	PPBV	99
18) METHYLENE CHLORIDE	6.55	84	668158	21.36	PPBV	98
19) 3-CHLOROPROPENE	6.66	39	480167	22.26	PPBV	98
20) FREON 113	6.79	151	1652492	21.13	PPBV	99
21) TRANS-1,2-DICHLOROETHYLENE	7.46	96	731482	21.90	PPBV	99
22) TERTIARY BUTYL ALCOHOL	6.54	59	1052857	23.77	PPBV	95
23) METHYL TERTIARY BUTYL ETHE	7.73	73	1480582	23.06	PPBV	100
24) TETRAHYDROFURAN	9.17	42	435946	23.75	PPBV	98
25) HEXANE	8.64	57	708987	21.16	PPBV	98
26) VINYL ACETATE	7.80	43	917657	23.76	PPBV	98
27) 1,1-DICHLOROETHANE	7.66	63	1114776	21.27	PPBV	100
28) METHYL ETHYL KETONE	8.07	43	825016	23.01	PPBV	94
29) cis-1,2-DICHLOROETHYLENE	8.46	96	685038	21.79	PPBV	99
30) ETHYL ACETATE	8.69	43	911778	22.80	PPBV #	87
31) CHLOROFORM	8.75	83	1604838	21.23	PPBV	100
32) 1,1,1-TRICHLOROETHANE	9.69	97	1946543	21.91	PPBV	99
33) CARBON TETRACHLORIDE	10.26	117	2458241	21.01	PPBV	100
34) 1,2-DICHLOROETHANE	9.46	62	937386	22.52	PPBV	100
36) BENZENE	10.13	78	1689378	22.44	PPBV	100
37) CYCLOHEXANE	10.37	84	830372	21.38	PPBV	100
38) TRICHLOROETHYLENE	11.12	95	912002	21.88	PPBV	99
39) 1,2-DICHLOROPROPANE	10.90	63	539801	22.27	PPBV	99
40) BROMODICHLOROMETHANE	11.09	83	1611690	21.90	PPBV	100
41) 2,2,4-TRIMETHYLPENTANE	11.12	57	2706384	21.79	PPBV	100
42) 1,4-DIOXANE	11.17	88	334145	24.33	PPBV	97
43) HEPTANE	11.37	43	792620	22.26	PPBV	99

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10280.D
Acq On : 9 May 2007 9:05 pm
Operator : danat
Sample : IC533-20 (M079)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 3 Sample Multiplier: 1

Quant Time: May 10 08:25:03 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 08:23:49 2007
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) METHYL ISOBUTYL KETONE	11.99	43	998239	23.56	PPBV	99
45) cis-1,3-DICHLOROPROPENE	11.94	75	916926	23.83	PPBV	99
46) TOLUENE	12.89	92	1270141	23.50	PPBV	100
47) trans-1,3-DICHLOROPROPENE	12.45	75	906659	24.47	PPBV	100
48) 1,1,2-TRICHLOROETHANE	12.63	83	613035	22.52	PPBV	99
50) 2-HEXANONE	13.16	43	863718	23.87	PPBV	99
51) TETRACHLOROETHYLENE	14.02	164	1140031	22.61	PPBV	98
52) DIBROMOCHLOROMETHANE	13.33	129	1748302	21.93	PPBV	100
53) 1,2-DIBROMOETHANE	13.58	107	1236020	22.93	PPBV	100
54) CHLOROBENZENE	14.72	112	1854324	22.50	PPBV	99
55) ETHYLBENZENE	15.10	91	2884297	22.96	PPBV	99
56) m,p-XYLENE	15.29	106	2360655	44.73	PPBV	95
57) o-XYLENE	15.79	106	1187661	22.43	PPBV	99
58) STYRENE	15.68	104	1691471	24.42	PPBV	99
59) NONANE	15.98	43	1099191	22.09	PPBV	98
60) BROMOFORM	15.41	173	1913327	22.66	PPBV	100
62) 1,1,2,2-TETRACHLOROETHANE	15.80	83	1633305	22.53	PPBV	100
63) ISOPROPYLBENZENE	16.43	105	3466671	22.84	PPBV	100
64) 2-CHLOROTOLUENE	16.99	91	2358153	23.58	PPBV	99
65) 4-ETHYLTOLUENE	17.18	105	3154305	24.09	PPBV	99
66) 1,3,5-TRIMETHYLBENZENE	17.26	105	2843131	23.53	PPBV	100
67) 1,2,4-TRIMETHYLBENZENE	17.73	105	2612283	24.52	PPBV	99
68) m-DICHLOROBENZENE	17.91	146	1901266	24.34	PPBV	98
69) BENZYL CHLORIDE	17.89	91	1575414	24.90	PPBV	100
70) p-DICHLOROBENZENE	17.99	146	1791624	24.26	PPBV	100
71) o-DICHLOROBENZENE	18.38	146	1681726	24.29	PPBV	100
72) HEXACHLOROBUTADIENE	20.77	225	977770	22.81	PPBV	98
73) 1,2,4-TRICHLOROBENZENE	20.27	180	573837	25.31	PPBV	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10280.D
Acq On : 9 May 2007 9:05 pm
Operator : danat
Sample : IC533-20 (M079)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 3 Sample Multiplier: 1

Quant Time: May 10 08:25:03 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 08:23:49 2007
Response via : Initial Calibration



ACCUTEST Laboratories
M64467

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10281.D
Acq On : 9 May 2007 9:52 pm
Operator : danat
Sample : ICc533-10 (M079)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 3 Sample Multiplier: 1

Quant Time: May 10 08:23:23 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Wed May 09 17:38:43 2007
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) BROMOCHLOROMETHANE	8.62	128	412358	10.00	PPBV	0.00
35) 1,4-DIFLUOROBENZENE	10.44	114	1821050	10.00	PPBV	0.01
49) CHLOROBENZENE-D5	14.67	117	1621470	10.00	PPBV	0.00
System Monitoring Compounds						
61) 4-BROMOFLUOROBENZENE	16.29	95	525842	5.13	PPBV	0.00
Spiked Amount 5.000	Range 51 - 142		Recovery	= 102.60%		
Target Compounds						Qvalue
2) DICHLORODIFLUOROMETHANE	4.06	85	1335034	9.08	PPBV	99
3) PROPYLENE	3.99	41	185611	9.31	PPBV	97
4) FREON 114	4.32	85	1104094	8.84	PPBV	100
5) CHLOROMETHANE	4.23	50	249129	8.69	PPBV	97
6) VINYL CHLORIDE	4.45	62	320310	8.89	PPBV #	89
7) 1,3-BUTADIENE	4.58	39	212715	9.09	PPBV	100
8) BROMOMETHANE	4.86	94	378088	8.86	PPBV	100
9) CHLOROETHANE	5.03	64	144488	8.85	PPBV	98
10) TRICHLOROFLUOROMETHANE	5.79	101	1311549	9.02	PPBV	99
11) ISOPROPYL ALCOHOL	5.92	45	322164	9.12	PPBV	97
12) ACETONE	5.66	43	335368	8.76	PPBV	99
13) PENTANE	6.14	42	244642	9.20	PPBV	99
14) 1,1-DICHLOROETHYLENE	6.42	96	372254	9.11	PPBV	99
15) CARBON DISULFIDE	6.85	76	1011086	8.74	PPBV	99
16) ETHANOL	5.21	45	51209	10.38	PPBV	75
17) BROMOETHENE	5.38	106	378606	9.00	PPBV	99
18) METHYLENE CHLORIDE	6.55	84	318607	8.17	PPBV	97
19) 3-CHLOROPROPENE	6.66	39	219616	8.54	PPBV	99
20) FREON 113	6.79	151	796267	8.81	PPBV	99
21) TRANS-1,2-DICHLOROETHYLENE	7.45	96	340162	8.61	PPBV	99
22) TERTIARY BUTYL ALCOHOL	6.53	59	451125	9.09	PPBV	99
23) METHYL TERTIARY BUTYL ETHE	7.73	73	653883	9.01	PPBV	99
24) TETRAHYDROFURAN	9.18	42	186932	8.98	PPBV	98
25) HEXANE	8.63	57	341232	8.84	PPBV	98
26) VINYL ACETATE	7.80	43	393252	8.94	PPBV	99
27) 1,1-DICHLOROETHANE	7.66	63	533713	8.80	PPBV	100
28) METHYL ETHYL KETONE	8.07	43	365150	8.93	PPBV	98
29) cis-1,2-DICHLOROETHYLENE	8.46	96	320213	8.58	PPBV	99
30) ETHYL ACETATE	8.69	43	407296	8.95	PPBV	100
31) CHLOROFORM	8.74	83	769763	8.75	PPBV	100
32) 1,1,1-TRICHLOROETHANE	9.68	97	904749	8.63	PPBV	99
33) CARBON TETRACHLORIDE	10.26	117	1191513	8.88	PPBV	99
34) 1,2-DICHLOROETHANE	9.45	62	423903	8.57	PPBV	100
36) BENZENE	10.12	78	755065	8.40	PPBV	99
37) CYCLOHEXANE	10.37	84	389500	8.84	PPBV	99
38) TRICHLOROETHYLENE	11.12	95	418019	8.34	PPBV	99
39) 1,2-DICHLOROPROPANE	10.90	63	243106	8.40	PPBV	99
40) BROMODICHLOROMETHANE	11.09	83	738122	8.49	PPBV	99
41) 2,2,4-TRIMETHYLPENTANE	11.12	57	1245861	8.95	PPBV	99
42) 1,4-DIOXANE	11.17	88	137760	8.54	PPBV	95
43) HEPTANE	11.36	43	357154	8.77	PPBV	98

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10281.D
Acq On : 9 May 2007 9:52 pm
Operator : danat
Sample : ICc533-10 (M079)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 3 Sample Multiplier: 1

Quant Time: May 10 08:23:23 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Wed May 09 17:38:43 2007
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) METHYL ISOBUTYL KETONE	11.99	43	425071	8.67	PPBV	100
45) cis-1,3-DICHLOROPROPENE	11.94	75	385895	8.76	PPBV	100
46) TOLUENE	12.89	92	542169	8.61	PPBV	99
47) trans-1,3-DICHLOROPROPENE	12.44	75	371618	8.76	PPBV	99
48) 1,1,2-TRICHLOROETHANE	12.62	83	273074	8.42	PPBV	98
50) 2-HEXANONE	13.16	43	350058	8.42	PPBV	100
51) TETRACHLOROETHYLENE	14.01	164	487777	8.31	PPBV	99
52) DIBROMOCHLOROMETHANE	13.33	129	771223	8.59	PPBV	100
53) 1,2-DIBROMOETHANE	13.57	107	521511	8.51	PPBV	100
54) CHLOROBENZENE	14.72	112	797302	8.58	PPBV	99
55) ETHYLBENZENE	15.09	91	1215370	8.87	PPBV	100
56) m,p-XYLENE	15.29	106	1021055	17.90	PPBV	94
57) o-XYLENE	15.79	106	512115	8.99	PPBV	99
58) STYRENE	15.67	104	670103	9.09	PPBV	99
59) NONANE	15.97	43	481225	9.03	PPBV	99
60) BROMOFORM	15.40	173	816576	8.47	PPBV	100
62) 1,1,2,2-TETRACHLOROETHANE	15.79	83	701321	8.41	PPBV	100
63) ISOPROPYLBENZENE	16.43	105	1468071	8.91	PPBV	100
64) 2-CHLOROTOLUENE	16.98	91	967207	8.90	PPBV	99
65) 4-ETHYLTOLUENE	17.17	105	1266694	8.82	PPBV	99
66) 1,3,5-TRIMETHYLBENZENE	17.26	105	1168837	8.75	PPBV	100
67) 1,2,4-TRIMETHYLBENZENE	17.72	105	1030474	8.66	PPBV	99
68) m-DICHLOROBENZENE	17.91	146	755613	8.19	PPBV	99
69) BENZYL CHLORIDE	17.89	91	611993	8.55	PPBV	99
70) p-DICHLOROBENZENE	17.98	146	714462	8.36	PPBV	99
71) o-DICHLOROBENZENE	18.37	146	669831	8.13	PPBV	99
72) HEXACHLOROBUTADIENE	20.76	225	414611	5.84	PPBV	98
73) 1,2,4-TRICHLOROBENZENE	20.27	180	219261	7.02	PPBV	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10281.D
Acq On : 9 May 2007 9:52 pm
Operator : danat
Sample : ICc533-10 (M079)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 3 Sample Multiplier: 1

Quant Time: May 10 08:23:23 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Wed May 09 17:38:43 2007
Response via : Initial Calibration



ACCUTEST Laboratories M64467

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10282.D
Acq On : 9 May 2007 10:40 pm
Operator : danat
Sample : IC533-5 (M079)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 3 Sample Multiplier: 1

Quant Time: May 10 08:28:02 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 08:26:43 2007
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) BROMOCHLOROMETHANE	8.62	128	366773	10.00	PPBV	0.00
35) 1,4-DIFLUOROBENZENE	10.44	114	1563527	10.00	PPBV	0.00
49) CHLOROBENZENE-D5	14.67	117	1351983	10.00	PPBV	0.00
System Monitoring Compounds						
61) 4-BROMOFLUOROBENZENE	16.29	95	381322	4.17	PPBV	0.00
Spiked Amount 5.000	Range 51 - 142		Recovery = 83.40%			
Target Compounds						Qvalue
2) DICHLORODIFLUOROMETHANE	4.07	85	659515	5.60	PPBV	100
3) PROPYLENE	4.00	41	83889	5.06	PPBV	89
4) FREON 114	4.33	85	503560	5.04	PPBV	98
5) CHLOROMETHANE	4.24	50	119629	5.32	PPBV	99
6) VINYL CHLORIDE	4.45	62	141719	4.93	PPBV	97
7) 1,3-BUTADIENE	4.58	39	94862	4.77	PPBV	97
8) BROMOMETHANE	4.86	94	173852	5.04	PPBV	98
9) CHLOROETHANE	5.03	64	61935	4.65	PPBV	100
10) TRICHLOROFLUOROMETHANE	5.80	101	596208	5.02	PPBV	99
11) ISOPROPYL ALCOHOL	5.93	45	124059	4.00	PPBV	97
12) ACETONE	5.66	43	145315	4.51	PPBV	98
13) PENTANE	6.14	42	108338	4.92	PPBV	99
14) 1,1-DICHLOROETHYLENE	6.43	96	158133	4.64	PPBV	97
15) CARBON DISULFIDE	6.85	76	467845	5.12	PPBV	100
16) ETHANOL	5.20	45	20144	3.97	PPBV	77
17) BROMOETHENE	5.39	106	163524	4.64	PPBV	99
18) METHYLENE CHLORIDE	6.55	84	144762	4.89	PPBV	98
19) 3-CHLOROPROPENE	6.66	39	91378	4.19	PPBV	99
20) FREON 113	6.79	151	365624	4.97	PPBV	100
21) TRANS-1,2-DICHLOROETHYLENE	7.46	96	148077	4.59	PPBV	99
22) TERTIARY BUTYL ALCOHOL	6.53	59	185663	4.16	PPBV	98
23) METHYL TERTIARY BUTYL ETHE	7.73	73	275497	4.29	PPBV	99
24) TETRAHYDROFURAN	9.18	42	73632	3.91	PPBV	97
25) HEXANE	8.63	57	148311	4.63	PPBV	96
26) VINYL ACETATE	7.80	43	159621	4.00	PPBV	97
27) 1,1-DICHLOROETHANE	7.65	63	230505	4.57	PPBV	100
28) METHYL ETHYL KETONE	8.07	43	151968	4.26	PPBV	97
29) cis-1,2-DICHLOROETHYLENE	8.46	96	134707	4.36	PPBV	99
30) ETHYL ACETATE	8.69	43	165297	4.17	PPBV	99
31) CHLOROFORM	8.74	83	337126	4.62	PPBV	99
32) 1,1,1-TRICHLOROETHANE	9.68	97	389206	4.43	PPBV	99
33) CARBON TETRACHLORIDE	10.26	117	538078	4.82	PPBV	100
34) 1,2-DICHLOROETHANE	9.45	62	182824	4.33	PPBV	99
36) BENZENE	10.12	78	304284	4.16	PPBV	99
37) CYCLOHEXANE	10.37	84	165185	4.64	PPBV	98
38) TRICHLOROETHYLENE	11.12	95	183986	4.67	PPBV	99
39) 1,2-DICHLOROPROPANE	10.90	63	102207	4.34	PPBV	97
40) BROMODICHLOROMETHANE	11.08	83	319064	4.57	PPBV	100
41) 2,2,4-TRIMETHYLPENTANE	11.12	57	494974	4.28	PPBV	97
42) 1,4-DIOXANE	11.18	88	47550	3.44	PPBV	90
43) HEPTANE	11.36	43	143860	4.21	PPBV	98

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10282.D
Acq On : 9 May 2007 10:40 pm
Operator : danat
Sample : IC533-5 (M079)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 3 Sample Multiplier: 1

Quant Time: May 10 08:28:02 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 08:26:43 2007
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
44) METHYL ISOBUTYL KETONE	11.99	43	163941	3.87 PPBV	98
45) cis-1,3-DICHLOROPROPENE	11.94	75	147966	3.80 PPBV	97
46) TOLUENE	12.88	92	213818	3.93 PPBV	97
47) trans-1,3-DICHLOROPROPENE	12.44	75	138222	3.61 PPBV	98
48) 1,1,2-TRICHLOROETHANE	12.62	83	115078	4.36 PPBV	98
50) 2-HEXANONE	13.16	43	133852	3.89 PPBV	97
51) TETRACHLOROETHYLENE	14.01	164	204474	4.42 PPBV	99
52) DIBROMOCHLOROMETHANE	13.32	129	331347	4.68 PPBV	100
53) 1,2-DIBROMOETHANE	13.57	107	212135	4.30 PPBV	99
54) CHLOROBENZENE	14.72	112	333880	4.51 PPBV	99
55) ETHYLBENZENE	15.09	91	480853	4.18 PPBV	99
56) m,p-XYLENE	15.27	106	406843	8.53 PPBV	97
57) o-XYLENE	15.78	106	203652	4.22 PPBV	99
58) STYRENE	15.66	104	228166	3.39 PPBV	99
59) NONANE	15.97	43	186332	4.18 PPBV	100
60) BROMOFORM	15.39	173	330866	4.29 PPBV	99
62) 1,1,2,2-TETRACHLOROETHANE	15.79	83	289837	4.42 PPBV	99
63) ISOPROPYLBENZENE	16.42	105	561060	4.02 PPBV	99
64) 2-CHLOROTOLUENE	16.98	91	354631	3.75 PPBV	99
65) 4-ETHYLTOLUENE	17.17	105	458887	3.63 PPBV #	45
66) 1,3,5-TRIMETHYLBENZENE	17.25	105	427329	3.74 PPBV	99
67) 1,2,4-TRIMETHYLBENZENE	17.72	105	377565	3.58 PPBV	99
68) m-DICHLOROBENZENE	17.90	146	307821	3.98 PPBV	99
69) BENZYL CHLORIDE	17.89	91	224322	3.47 PPBV	98
70) p-DICHLOROBENZENE	17.98	146	290086	3.99 PPBV	98
71) o-DICHLOROBENZENE	18.37	146	259799	3.77 PPBV	98
72) HEXACHLOROBUTADIENE	20.76	225	180506	4.47 PPBV	98
73) 1,2,4-TRICHLOROBENZENE	20.27	180	80284	3.48 PPBV	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10282.D
Acq On : 9 May 2007 10:40 pm
Operator : danat
Sample : IC533-5 (M079)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 3 Sample Multiplier: 1

Quant Time: May 10 08:28:02 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 08:26:43 2007
Response via : Initial Calibration



ACCUTEST Laboratories M64467

J10283.D

Manual Integrations APPROVED (compounds with "m" flag)

Philip Bassignani 05/11/07 16:35

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10283.D
Acq On : 9 May 2007 11:27 pm
Operator : danat
Sample : IC533-2 (M079)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 3 Sample Multiplier: 1

Quant Time: May 10 08:30:24 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 08:28:13 2007
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) BROMOCHLOROMETHANE	8.62	128	364997	10.00	PPBV	0.00
35) 1,4-DIFLUOROBENZENE	10.44	114	1696449	10.00	PPBV	0.00
49) CHLOROBENZENE-D5	14.67	117	1441208	10.00	PPBV	0.00

System Monitoring Compounds

61) 4-BROMOFLUOROBENZENE	16.29	95	390695	4.19	PPBV	0.00

Spiked Amount 5.000 Range 51 - 142 Recovery = 83.80%

Target Compounds						Qvalue
2) DICHLORODIFLUOROMETHANE	4.07	85	265326	2.20	PPBV	99
3) PROPYLENE	3.99	41	27646	1.67	PPBV #	77
4) FREON 114	4.32	85	242119	2.43	PPBV	99
5) CHLOROMETHANE	4.23	50	53357	2.35	PPBV	94
6) VINYL CHLORIDE	4.44	62	59590	2.09	PPBV #	51
7) 1,3-BUTADIENE	4.59	39	38962	1.99	PPBV	96
8) BROMOMETHANE	4.86	94	79695	2.32	PPBV	99
9) CHLOROETHANE	5.03	64	23246	1.78	PPBV	99
10) TRICHLOROFLUOROMETHANE	5.80	101	285851	2.42	PPBV	98
11) ISOPROPYL ALCOHOL	5.95	45	36541	1.25	PPBV	82
12) ACETONE	5.67	43	62104	1.99	PPBV	66
13) PENTANE	6.14	42	49414	2.26	PPBV	98
14) 1,1-DICHLOROETHYLENE	6.42	96	68145	2.05	PPBV	96
15) CARBON DISULFIDE	6.85	76	212409	2.32	PPBV	95
16) ETHANOL	5.22	45	8373m	1.75	PPBV	
17) BROMOETHENE	5.38	106	69624	2.02	PPBV	99
18) METHYLENE CHLORIDE	6.55	84	62738	2.14	PPBV	96
19) 3-CHLOROPROPENE	6.67	39	38776	1.86	PPBV	93
20) FREON 113	6.80	151	163182	2.23	PPBV	99
21) TRANS-1,2-DICHLOROETHYLENE	7.46	96	62673	1.99	PPBV	99
22) TERTIARY BUTYL ALCOHOL	6.56	59	63841	1.50	PPBV	96
23) METHYL TERTIARY BUTYL ETHE	7.74	73	112307	1.82	PPBV	98
24) TETRAHYDROFURAN	9.20	42	27340	1.54	PPBV	89
25) HEXANE	8.63	57	64000	2.05	PPBV	92
26) VINYL ACETATE	7.81	43	59142	1.57	PPBV	95
27) 1,1-DICHLOROETHANE	7.65	63	103936	2.12	PPBV	99
28) METHYL ETHYL KETONE	8.08	43	64492	1.89	PPBV	95
29) cis-1,2-DICHLOROETHYLENE	8.46	96	61795	2.07	PPBV	99
30) ETHYL ACETATE	8.70	43	70793	1.87	PPBV #	98
31) CHLOROFORM	8.74	83	157428	2.21	PPBV	98
32) 1,1,1-TRICHLOROETHANE	9.68	97	173991	2.05	PPBV	99
33) CARBON TETRACHLORIDE	10.26	117	231505	2.10	PPBV	99
34) 1,2-DICHLOROETHANE	9.45	62	86021	2.12	PPBV	100
36) BENZENE	10.12	78	131797	1.73	PPBV	98
37) CYCLOHEXANE	10.37	84	65341	1.72	PPBV	99
38) TRICHLOROETHYLENE	11.11	95	87630	2.08	PPBV	100
39) 1,2-DICHLOROPROPANE	10.90	63	44533	1.80	PPBV	90
40) BROMODICHLOROMETHANE	11.08	83	152058	2.05	PPBV	99
41) 2,2,4-TRIMETHYLPENTANE	11.12	57	192800	1.59	PPBV	95
42) 1,4-DIOXANE	11.20	88	13702m	0.99	PPBV	
43) HEPTANE	11.36	43	60293	1.69	PPBV	99

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10283.D
Acq On : 9 May 2007 11:27 pm
Operator : danat
Sample : IC533-2 (M079)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 3 Sample Multiplier: 1

Quant Time: May 10 08:30:24 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 08:28:13 2007
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) METHYL ISOBUTYL KETONE	12.00	43	61589	1.42	PPBV	93
45) cis-1,3-DICHLOROPROPENE	11.93	75	57253	1.44	PPBV	97
46) TOLUENE	12.89	92	83319	1.49	PPBV	93
47) trans-1,3-DICHLOROPROPENE	12.44	75	53056	1.37	PPBV	99
48) 1,1,2-TRICHLOROETHANE	12.62	83	52926	1.91	PPBV	96
50) 2-HEXANONE	13.18	43	42072	1.22	PPBV	92
51) TETRACHLOROETHYLENE	14.01	164	92176	1.93	PPBV	99
52) DIBROMOCHLOROMETHANE	13.32	129	156054	2.10	PPBV	99
53) 1,2-DIBROMOETHANE	13.57	107	94616	1.86	PPBV	99
54) CHLOROBENZENE	14.72	112	158458	2.06	PPBV	98
55) ETHYLBENZENE	15.09	91	177378	1.51	PPBV	99
56) m,p-XYLENE	15.28	106	158064m	3.23	PPBV	
57) o-XYLENE	15.78	106	76083	1.54	PPBV	96
58) STYRENE	15.67	104	73413	1.11	PPBV	98
59) NONANE	15.97	43	72444	1.59	PPBV	99
60) BROMOFORM	15.39	173	150352	1.90	PPBV	99
62) 1,1,2,2-TETRACHLOROETHANE	15.79	83	125824	1.85	PPBV	99
63) ISOPROPYLBENZENE	16.42	105	213050	1.51	PPBV	99
64) 2-CHLOROTOLUENE	16.98	91	131457	1.39	PPBV	98
65) 4-ETHYLTOLUENE	17.17	105	162932	1.30	PPBV	98
66) 1,3,5-TRIMETHYLBENZENE	17.25	105	158041	1.38	PPBV	97
67) 1,2,4-TRIMETHYLBENZENE	17.72	105	129200	1.24	PPBV	99
68) m-DICHLOROBENZENE	17.90	146	130244	1.66	PPBV	97
69) BENZYL CHLORIDE	17.89	91	72946	1.15	PPBV	100
70) p-DICHLOROBENZENE	17.98	146	110415	1.50	PPBV	98
71) o-DICHLOROBENZENE	18.37	146	105939	1.54	PPBV	97
72) HEXACHLOROBUTADIENE	20.76	225	74693	1.78	PPBV	96
73) 1,2,4-TRICHLOROBENZENE	20.27	180	25976	1.14	PPBV	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10283.D
Acq On : 9 May 2007 11:27 pm
Operator : danat
Sample : IC533-2 (M079)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 3 Sample Multiplier: 1

Quant Time: May 10 08:30:24 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 08:28:13 2007
Response via : Initial Calibration



Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10283.D Vial: 3
Acq On : 9 May 2007 11:27 pm Operator: danat
Sample : IC533-2 (M079) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:28:50 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:28:13 2007
Response via : Multiple Level Calibration







(16) ETHANOL (m)

5.29min 0.27PPBV

response 1281

Ion	Exp%	Act%
45.00	100	100
46.00	14.10	14.99
42.00	2.30	6.17
0.00	0.00	0.00

J10283.D J050907T.M Thu May 10 08:29:19 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10283.D Vial: 3
Acq On : 9 May 2007 11:27 pm Operator: danat
Sample : IC533-2 (M079) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:28:50 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:28:13 2007
Response via : Multiple Level Calibration

(16) ETHANOL (m)

5.22min 1.75PPBV m

response 8373

Ion	Exp%	Act%
45.00	100	100
46.00	14.10	2.29
42.00	2.30	0.94
0.00	0.00	0.00

J10283.D J050907T.M Thu May 10 08:29:28 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10283.D Vial: 3
Acq On : 9 May 2007 11:27 pm Operator: danat
Sample : IC533-2 (M079) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:28:50 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:28:13 2007
Response via : Multiple Level Calibration







(42) 1,4-DIOXANE (m)

11.20min 0.80PPBV

response 11007

Ion	Exp%	Act%
88.00	100	100
57.00	26.70	18.48
58.00	51.70	32.19
43.00	24.80	28.34

J10283.D J050907T.M Thu May 10 08:29:58 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10283.D Vial: 3
Acq On : 9 May 2007 11:27 pm Operator: danat
Sample : IC533-2 (M079) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:28:50 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:28:13 2007
Response via : Multiple Level Calibration







(42) 1,4-DIOXANE (m)

11.20min 0.99PPBV m

response 13702

Ion	Exp%	Act%
88.00	100	100
57.00	26.70	18.48
58.00	51.70	32.19
43.00	24.80	28.34

J10283.D J050907T.M Thu May 10 08:30:04 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10283.D Vial: 3
Acq On : 9 May 2007 11:27 pm Operator: danat
Sample : IC533-2 (M079) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:28:50 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:28:13 2007
Response via : Multiple Level Calibration







(56) m,p-XYLENE (m)

15.27min 1.17PPBV

response 57465

Ion	Exp%	Act%
106.00	100	100
91.00	187.40	182.02
77.00	26.20	27.08
0.00	0.00	0.00

J10283.D J050907T.M Thu May 10 08:30:21 2007

Quantitation Report (Qedit)

Data File : C:\MSDChem\1\data\J10283.D Vial: 3
Acq On : 9 May 2007 11:27 pm Operator: danat
Sample : IC533-2 (M079) Inst : MSJ
Misc : MS13434, MSJ533,,,,,1 Multiplr: 1.00
MS Integration Params: LSCINT.P

Quant Time: May 10 08:28:50 2007 Results File: J050907T.RES

Method : C:\MSDCHEM\1\METHODS\J050907T.M (RTE Integrator)
Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
Last Update : Thu May 10 08:28:13 2007
Response via : Multiple Level Calibration







(56) m,p-XYLENE (m)

15.28min 3.23PPBV m

response 158064

Ion	Exp%	Act%
106.00	100	100
91.00	187.40	169.79
77.00	26.20	24.82
0.00	0.00	0.00

J10283.D J050907T.M Thu May 10 08:30:26 2007

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10286A.D
Acq On : 10 May 2007 10:33 am
Operator : danat
Sample : CC533-10 (M054)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 4 Sample Multiplier: 1

Quant Time: May 10 11:10:20 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 11:09:51 2007
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) BROMOCHLOROMETHANE	8.62	128	413189	10.00	PPBV	0.00
35) 1,4-DIFLUOROBENZENE	10.44	114	1835396	10.00	PPBV	0.00
49) CHLOROBENZENE-D5	14.67	117	1640141	10.00	PPBV	0.00

System Monitoring Compounds

61) 4-BROMOFLUOROBENZENE	16.29	95	531829	5.53	PPBV	0.00

Spiked Amount 5.000 Range 51 - 142 Recovery = 110.60%

Target Compounds						Qvalue
2) DICHLORODIFLUOROMETHANE	4.07	85	1305850	8.99	PPBV	100
3) PROPYLENE	3.99	41	178641	10.76	PPBV #	87
4) FREON 114	4.32	85	1056037	8.62	PPBV	99
5) CHLOROMETHANE	4.23	50	247208	9.17	PPBV	97
6) VINYL CHLORIDE	4.45	62	301533	8.91	PPBV	100
7) 1,3-BUTADIENE	4.58	39	204602	10.02	PPBV	97
8) BROMOMETHANE	4.85	94	360144	8.79	PPBV	99
9) CHLOROETHANE	5.03	64	140741	10.21	PPBV	99
10) TRICHLOROFLUOROMETHANE	5.79	101	1263256	8.54	PPBV	100
11) ISOPROPYL ALCOHOL	5.92	45	301406	9.67	PPBV	97
12) ACETONE	5.65	43	283156	7.72	PPBV	99
13) PENTANE	6.13	42	238500	9.59	PPBV	98
14) 1,1-DICHLOROETHYLENE	6.43	96	321653	8.69	PPBV	96
15) CARBON DISULFIDE	6.85	76	924747	8.26	PPBV	99
16) ETHANOL	5.20	45	49752	10.19	PPBV	96
17) BROMOETHENE	5.38	106	346747	9.30	PPBV	100
18) METHYLENE CHLORIDE	6.54	84	285766	7.73	PPBV	97
19) 3-CHLOROPROPENE	6.66	39	219659	9.18	PPBV	96
20) FREON 113	6.79	151	717493	7.93	PPBV	99
21) TRANS-1,2-DICHLOROETHYLENE	7.46	96	317265	8.82	PPBV	98
22) TERTIARY BUTYL ALCOHOL	6.52	59	409671	9.61	PPBV	99
23) METHYL TERTIARY BUTYL ETHE	7.73	73	534027	8.16	PPBV	98
24) TETRAHYDROFURAN	9.17	42	153701	8.55	PPBV	98
25) HEXANE	8.63	57	300535	7.52	PPBV	97
26) VINYL ACETATE	7.80	43	322349	8.07	PPBV	98
27) 1,1-DICHLOROETHANE	7.66	63	492394	8.46	PPBV	100
28) METHYL ETHYL KETONE	8.07	43	304589	8.22	PPBV	94
29) cis-1,2-DICHLOROETHYLENE	8.46	96	285978	8.56	PPBV	97
30) ETHYL ACETATE	8.68	43	324353	7.81	PPBV	99
31) CHLOROFORM	8.74	83	712177	8.31	PPBV	99
32) 1,1,1-TRICHLOROETHANE	9.68	97	854284	8.59	PPBV	99
33) CARBON TETRACHLORIDE	10.26	117	1127397	8.62	PPBV	99
34) 1,2-DICHLOROETHANE	9.45	62	382396	8.09	PPBV	100
36) BENZENE	10.12	78	669804	8.55	PPBV	100
37) CYCLOHEXANE	10.37	84	353671	9.72	PPBV	99
38) TRICHLOROETHYLENE	11.12	95	389434	8.76	PPBV	97
39) 1,2-DICHLOROPROPANE	10.90	63	214272	8.13	PPBV	98
40) BROMODICHLOROMETHANE	11.08	83	655320	7.76	PPBV	99
41) 2,2,4-TRIMETHYLPENTANE	11.12	57	1187978	10.49	PPBV	98
42) 1,4-DIOXANE	11.16	88	122751	8.14	PPBV	95
43) HEPTANE	11.36	43	323287	9.70	PPBV	99

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10286A.D
Acq On : 10 May 2007 10:33 am
Operator : danat
Sample : CC533-10 (M054)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 4 Sample Multiplier: 1

Quant Time: May 10 11:10:20 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 11:09:51 2007
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) METHYL ISOBUTYL KETONE	11.98	43	366654	8.87	PPBV	98
45) cis-1,3-DICHLOROPROPENE	11.94	75	349534	9.62	PPBV	98
46) TOLUENE	12.89	92	462783	9.02	PPBV	99
47) trans-1,3-DICHLOROPROPENE	12.44	75	290402	7.07	PPBV	99
48) 1,1,2-TRICHLOROETHANE	12.62	83	240426	8.02	PPBV	99
50) 2-HEXANONE	13.15	43	285755	7.38	PPBV	98
51) TETRACHLOROETHYLENE	14.01	164	445675	8.05	PPBV	99
52) DIBROMOCHLOROMETHANE	13.33	129	679264	7.62	PPBV	100
53) 1,2-DIBROMOETHANE	13.57	107	465149	8.46	PPBV	99
54) CHLOROBENZENE	14.71	112	714978	8.20	PPBV	98
55) ETHYLBENZENE	15.09	91	1072352	9.61	PPBV	99
56) m,p-XYLENE	15.28	106	907029	15.87	PPBV	100
57) o-XYLENE	15.78	106	461835	7.99	PPBV	100
58) STYRENE	15.67	104	610702	8.04	PPBV	100
59) NONANE	15.97	43	458405	9.71	PPBV	99
60) BROMOFORM	15.40	173	731566	8.07	PPBV	100
62) 1,1,2,2-TETRACHLOROETHANE	15.79	83	671390	8.73	PPBV	100
63) ISOPROPYLBENZENE	16.43	105	1349019	9.29	PPBV	99
64) 2-CHLOROTOLUENE	16.98	91	917142	9.59	PPBV	99
65) 4-ETHYLTOLUENE	17.17	105	1113566	7.77	PPBV	99
66) 1,3,5-TRIMETHYLBENZENE	17.25	105	1099713	9.63	PPBV	99
67) 1,2,4-TRIMETHYLBENZENE	17.72	105	998081	8.26	PPBV	99
68) m-DICHLOROBENZENE	17.91	146	797605	9.78	PPBV	99
69) BENZYL CHLORIDE	17.89	91	637252	9.90	PPBV	100
70) p-DICHLOROBENZENE	17.98	146	749201	10.05	PPBV	98
71) o-DICHLOROBENZENE	18.37	146	706204	9.72	PPBV	100
72) HEXACHLOROBUTADIENE	20.76	225	520415	11.87	PPBV	100
73) 1,2,4-TRICHLOROBENZENE	20.27	180	330262	13.97	PPBV	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (QT Reviewed)

Data Path : C:\MSDChem\1\data\
Data File : J10286A.D
Acq On : 10 May 2007 10:33 am
Operator : danat
Sample : CC533-10 (M054)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 4 Sample Multiplier: 1

Quant Time: May 10 11:10:20 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 11:09:51 2007
Response via : Initial Calibration



Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\
Data File : J10286.D
Acq On : 10 May 2007 10:33 am
Operator : danat
Sample : ICV533-10 (M054)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 4 Sample Multiplier: 1

Quant Time: May 10 11:10:20 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 11:09:51 2007
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) BROMOCHLOROMETHANE	8.62	128	413189	10.00	PPBV	0.00
35) 1,4-DIFLUOROBENZENE	10.44	114	1835396	10.00	PPBV	0.00
49) CHLOROBENZENE-D5	14.67	117	1640141	10.00	PPBV	0.00

System Monitoring Compounds

61) 4-BROMOFLUOROBENZENE	16.29	95	531829	5.53	PPBV	0.00

Spiked Amount 5.000 Range 51 - 142 Recovery = 110.60%

Target Compounds						Qvalue
2) DICHLORODIFLUOROMETHANE	4.07	85	1305850	8.99	PPBV	100
3) PROPYLENE	3.99	41	178641	10.76	PPBV #	87
4) FREON 114	4.32	85	1056037	8.62	PPBV	99
5) CHLOROMETHANE	4.23	50	247208	9.17	PPBV	97
6) VINYL CHLORIDE	4.45	62	301533	8.91	PPBV	100
7) 1,3-BUTADIENE	4.58	39	204602	10.02	PPBV	97
8) BROMOMETHANE	4.85	94	360144	8.79	PPBV	99
9) CHLOROETHANE	5.03	64	140741	10.21	PPBV	99
10) TRICHLOROFLUOROMETHANE	5.79	101	1263256	8.54	PPBV	100
11) ISOPROPYL ALCOHOL	5.92	45	301406	9.67	PPBV	97
12) ACETONE	5.65	43	283156	7.72	PPBV	99
13) PENTANE	6.13	42	238500	9.59	PPBV	98
14) 1,1-DICHLOROETHYLENE	6.43	96	321653	8.69	PPBV	96
15) CARBON DISULFIDE	6.85	76	924747	8.26	PPBV	99
16) ETHANOL	5.20	45	49752	10.19	PPBV	96
17) BROMOETHENE	5.38	106	346747	9.30	PPBV	100
18) METHYLENE CHLORIDE	6.54	84	285766	7.73	PPBV	97
19) 3-CHLOROPROPENE	6.66	39	219659	9.18	PPBV	96
20) FREON 113	6.79	151	717493	7.93	PPBV	99
21) TRANS-1,2-DICHLOROETHYLENE	7.46	96	317265	8.82	PPBV	98
22) TERTIARY BUTYL ALCOHOL	6.52	59	409671	9.61	PPBV	99
23) METHYL TERTIARY BUTYL ETHE	7.73	73	534027	8.16	PPBV	98
24) TETRAHYDROFURAN	9.17	42	153701	8.55	PPBV	98
25) HEXANE	8.63	57	300535	7.52	PPBV	97
26) VINYL ACETATE	7.80	43	322349	8.07	PPBV	98
27) 1,1-DICHLOROETHANE	7.66	63	492394	8.46	PPBV	100
28) METHYL ETHYL KETONE	8.07	43	304589	8.22	PPBV	94
29) cis-1,2-DICHLOROETHYLENE	8.46	96	285978	8.56	PPBV	97
30) ETHYL ACETATE	8.68	43	324353	7.81	PPBV	99
31) CHLOROFORM	8.74	83	712177	8.31	PPBV	99
32) 1,1,1-TRICHLOROETHANE	9.68	97	854284	8.59	PPBV	99
33) CARBON TETRACHLORIDE	10.26	117	1127397	8.62	PPBV	99
34) 1,2-DICHLOROETHANE	9.45	62	382396	8.09	PPBV	100
36) BENZENE	10.12	78	669804	8.55	PPBV	100
37) CYCLOHEXANE	10.37	84	353671	9.72	PPBV	99
38) TRICHLOROETHYLENE	11.12	95	389434	8.76	PPBV	97
39) 1,2-DICHLOROPROPANE	10.90	63	214272	8.13	PPBV	98
40) BROMODICHLOROMETHANE	11.08	83	655320	7.76	PPBV	99
41) 2,2,4-TRIMETHYLPENTANE	11.12	57	1187978	10.49	PPBV	98
42) 1,4-DIOXANE	11.16	88	122751	8.14	PPBV	95
43) HEPTANE	11.36	43	323287	9.70	PPBV	99

Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\
Data File : J10286.D
Acq On : 10 May 2007 10:33 am
Operator : danat
Sample : ICV533-10 (M054)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 4 Sample Multiplier: 1

Quant Time: May 10 11:10:20 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 11:09:51 2007
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) METHYL ISOBUTYL KETONE	11.98	43	366654	8.87	PPBV	98
45) cis-1,3-DICHLOROPROPENE	11.94	75	349534	9.62	PPBV	98
46) TOLUENE	12.89	92	462783	9.02	PPBV	99
47) trans-1,3-DICHLOROPROPENE	12.44	75	290402	7.07	PPBV	99
48) 1,1,2-TRICHLOROETHANE	12.62	83	240426	8.02	PPBV	99
50) 2-HEXANONE	13.15	43	285755	7.38	PPBV	98
51) TETRACHLOROETHYLENE	14.01	164	445675	8.05	PPBV	99
52) DIBROMOCHLOROMETHANE	13.33	129	679264	7.62	PPBV	100
53) 1,2-DIBROMOETHANE	13.57	107	465149	8.46	PPBV	99
54) CHLOROBENZENE	14.71	112	714978	8.20	PPBV	98
55) ETHYLBENZENE	15.09	91	1072352	9.61	PPBV	99
56) m,p-XYLENE	15.28	106	907029	15.87	PPBV	100
57) o-XYLENE	15.78	106	461835	7.99	PPBV	100
58) STYRENE	15.67	104	610702	8.04	PPBV	100
59) NONANE	15.97	43	458405	9.71	PPBV	99
60) BROMOFORM	15.40	173	731566	8.07	PPBV	100
62) 1,1,2,2-TETRACHLOROETHANE	15.79	83	671390	8.73	PPBV	100
63) ISOPROPYLBENZENE	16.43	105	1349019	9.29	PPBV	99
64) 2-CHLOROTOLUENE	16.98	91	917142	9.59	PPBV	99
65) 4-ETHYLTOLUENE	17.17	105	1113566	7.77	PPBV	99
66) 1,3,5-TRIMETHYLBENZENE	17.25	105	1099713	9.63	PPBV	99
67) 1,2,4-TRIMETHYLBENZENE	17.72	105	998081	8.26	PPBV	99
68) m-DICHLOROBENZENE	17.91	146	797605	9.78	PPBV	99
69) BENZYL CHLORIDE	17.89	91	637252	9.90	PPBV	100
70) p-DICHLOROBENZENE	17.98	146	749201	10.05	PPBV	98
71) o-DICHLOROBENZENE	18.37	146	706204	9.72	PPBV	100
72) HEXACHLOROBUTADIENE	20.76	225	520415	11.87	PPBV	100
73) 1,2,4-TRICHLOROBENZENE	20.27	180	330262	13.97	PPBV	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\
Data File : J10286.D
Acq On : 10 May 2007 10:33 am
Operator : danat
Sample : ICV533-10 (M054)
Misc : MS13434, MSJ533,,,,,1
ALS Vial : 4 Sample Multiplier: 1

Quant Time: May 10 11:10:20 2007
Quant Method : C:\MSDCHEM\1\METHODS\J050907T.M
Quant Title : TO15 by GCMS w/DB-1 60 m X .25 mm ID 1.0 um
QLast Update : Thu May 10 11:09:51 2007
Response via : Initial Calibration



ACCUTEST Laboratories M64467

Standard Lot #	
Calib STD 1	MSA332
Calib STD 2	MSA333
ICV	MSA335
Inter/Surr STD	MSA218
BS/BSD	MSA335

File ID	
Tune File 1	J10277
Tune File 2	J10286A
Initial Cal.	5/9/07
ID FILE	J050907T.M
Instrument ID	J
ICAL:	J10286
Column Type:	RTX-1
Sequence Verified By:	DT 5/9/07

Date:	5/9/2007
Analyst:	DANAT
Batch ID:	msj533
Matrix:	AIR
Approved By:	PHILIPB
Date Approved:	5/15/2007 4:11:59 PM

Data File	Sample ID	Bot #	Work Group #	Test	ALS #	Sample Vol (mls)	Diln Factor	Flow controller ID	Comment
J10267	IC533-4		MSA273		1		1		
J10268	IC533-0.2		MSA333		2		1		
J10269	IC533-0.5		MSA333		2		1		missing compounds at reporting minimum.
J10270	IC533-2		MSA333		2		1		
J10271	IC530-40		MSA332		3		1		Raise em volts here and recalibrate.
J10272	IC533-20		MSA332		3		1		
J10273	ICC533-10		MSA332		3		1		
J10274	IC533-5		MSA332		3		1		
J10275	IC533-2		MSA333		2		1		
J10276	IC533-0.5		MSA333		2		1		Compounds still low , raise em volts and recalibrate.
J10277	IC533-0.2		MSA333		2		1		
J10278	IC533-0.5		MSA333		2		1		
J10279	IC533-40		MSA332		3		1		
J10280	IC533-20		MSA332		3		1		
J10281	ICC533-10		MSA332		3		1		
J10282	IC533-5		MSA332		3		1		
J10283	IC533-2		MSA333		2		1		
J10284	ICV533-10	m087	MSA334		4		1		compounds low , remake std.
J10285	ICV533-10	m105	MSA329		4		1		used old std with low volume , remake std.
J10286	ICV533-10	m054	MSA335		4		1		
J10286a	CC533-10	m054	MSA335		4		1		
J10286b	BS	m054	MSA335		4		1		
J10287	MB	m009			5		1		
J10288	MB	m009			5		1		
J10289	M64467-1	m126	MS13540	VTO15STD	6		1		
J10290	M64502-1	m018	MS13540	VTO15STD	7		1		
J10291	M64502-2	m194	MS13540	VTO15STD	8		100		Re-run , over calibration.
J10292	M64502-2	m194	MS13540	VTO15STD	8		1000		Re-run , over calibration.
J10293	M64467-1DUP	m126	MS13540	VTO15STD	6		1		
J10294	M64502-2	m200	MS13540	VTO15STD	8		5000		Re-make dilution , results not match original run.
J10295	M64555-1	m056	MS13540	VTO15STD	9		1		
J10296	M64555-2	m055	MS13540	VTO15STD	10		1		
J10297	M64644-1	m123	MS13540	VTO15STD	11		1		Re-run , over calibration.

J10298	M64644-2	m103	MS13540	VTO15STD	12		1		Re-run , over calibration.
J10299	M64644-3	m003	MS13540	VTO15STD	13		1		Re-run , over calibration.
J10300	M64242-1R	m099	MS13540	VTO15STD	14		1		
J10301	M64246-2R	m064	MS13540	VTO15STD	15		1		
J10302	M64247-1R	m048	MS13540	VTO15STD	16		1		
J10303	M64319-2R	m071	MS13540	VTO15STD	5		1		
J10304	SCC	m050			6		1		Re-run , internal std low.
J10305	SCC	m131			7		1		
J10306	SCC	m132			8		1		
J10307	SCC	m069			9		1		
J10308	SCC	m066			10		1		
J10309	SCC	m073			11		1		
J10310	SCC	m089			12		1		
J10311	SCC	m070			13		1		